

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012592F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Apolo Scales/APTT
Range	40g to 30 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	2209576144	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
50g	50.000105	0.050	0.000	0.016
100g	100.00025	0.100	0.000	
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.001	-0.001	
2kg	1999.9963	2.001	-0.001	
5kg	5000.012	5.003	-0.003	
10kg	10000.040	10.007	-0.007	
20kg	20000.08	20.009	-0.009	
30kg	30000.12	30.012	-0.012	
50kg	50000.14	50.016	-0.016	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
20.006	20.007	20.005	20.008	20.007	20.008	20.006	20.008	20.007	20.006	20.008

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
20 kg	20.004	20.006	20.005	20.004	20.006	20.007	20.008	20.006	20.007	20.006

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012592F

Page 2 of 2

4. All our certificates are Traceable to national/International standard.

5. This Certificate relates to the item Calibrated total liability of Testnral Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012609F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Apolo Scales/-
Range	100g to 30 kg	Instrument Condition	OK
Least Count	0.001 kg	Location of Calibration	At Site
Serial No.	2003425	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
100g	100.00025	0.100	0.000	0.018
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.001	-0.001	
2kg	1999.9963	2.003	-0.003	
5kg	5000.012	5.006	-0.006	
10kg	10000.040	10.008	-0.008	
20kg	20000.08	20.013	-0.013	
30kg	30000.12	30.018	-0.018	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
10.006	10.007	10.004	10.006	10.005	10.006	10.007	10.005	10.004	10.006	10.008

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
10 kg	10.008	10.009	10.008	10.007	10.005	10.008	10.006	10.008	10.007	10.005

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012618F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Swisser/-
Range	50g to 50 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	2221041	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
50g	50.000105	0.050	0.000	0.012
100g	100.00025	0.100	0.000	
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.001	-0.001	
2kg	1999.9963	2.001	-0.001	
5kg	5000.012	5.002	-0.002	
10kg	10000.040	10.004	-0.004	
20kg	20000.08	20.007	-0.007	
30kg	30000.12	30.009	-0.009	
50kg	50000.14	50.012	-0.012	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
20.007	20.008	20.007	20.009	20.010	20.009	20.010	20.009	20.010	20.007	20.009

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
20 kg	20.004	20.006	20.005	20.004	20.006	20.007	20.008	20.006	20.007	20.006

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012618F

Page 2 of 2

4. All our certificates are Traceable to national/International standard.

5. This Certificate relates to the item Calibrated total liability of Testnral Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012619F Page 1 of 2

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Weighing Balance	Make / Model No.	Toyo/-
Range	0 to 10 kg	Instrument Condition	OK
Least Count	0.5 g	Location of Calibration	At Site
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/WB/10kg/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (g)	Deviation (g)	C MAX= Maximum Correction (g)
20g	20.000056	20.0	0.0	1.0
50g	50.000105	50.0	0.0	
100g	100.00025	100.0	0.0	
200g	200.00024	200.0	0.0	
500g	500.0011	500.0	0.0	
1kg	999.9992	1000.0	0.0	
2kg	1999.9963	2000.0	0.0	
5kg	5000.012	5000.5	0.5	
10kg	10000.027	10001.0	1.0	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
5000.0	5000.5	5000.0	5000.5	5000.0	5000.5	5000.5	5000.0	5000.5	5000.0	5000.0

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	Standard Dev. (g)	Maximum Error Between Centre and Other Points (g)	Uncertainty (g)
5000g	5000.5	5000.5	5001.0	5000.5	5000.5	5000.0	5000.5	5001.0	5000.5	5000.0	0.3	-0.1	0.7



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012619F

Page 2 of 2

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012620F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Weighing Balance	Make / Model No.	Apolo/APTT
Range	20g to 10 kg	Instrument Condition	OK
Least Count	0.5 g	Location of Calibration	At Site
Serial No.	2209535789	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (g)	Deviation (g)	C MAX= Maximum Correction (g)
20g	20.000056	20.0	0.0	1.5
50g	50.000105	50.0	0.0	
100g	100.00025	100.0	0.0	
200g	200.00024	200.0	0.0	
500g	500.0011	500.0	0.0	
1kg	999.9992	1000.0	0.0	
2kg	1999.9963	1999.5	-0.5	
5kg	5000.012	4999.0	-1.0	
10kg	10000.027	9998.5	-1.5	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
4999.5	4999.5	5000.0	4999.5	5000.0	4999.5	4999.5	5000.0	4999.5	5000.0	5000.0

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	Standard Dev. (g)	Maximum Error Between Centre and Other Points (g)	Uncertainty (g)
5000g	4999.5	5000.0	4999.5	5000.0	5000.0	4999.5	5000.0	4999.5	4999.5	5000.0	0.3	-0.1	0.9



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012621F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Swisser/Swit501
Range	50g to 50 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	2221033	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
50g	50.000105	0.050	0.000	0.017
100g	100.00025	0.100	0.000	
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.000	0.000	
2kg	1999.9963	2.001	-0.001	
5kg	5000.012	5.002	-0.002	
10kg	10000.040	10.005	-0.005	
20kg	20000.08	20.008	-0.008	
30kg	30000.12	30.0011	-0.001	
50kg	50000.14	50.017	-0.017	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
20.008	20.006	20.007	20.006	20.005	20.008	20.007	20.008	20.008	20.009	20.007

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
20 kg	20.004	20.006	20.005	20.004	20.006	20.007	20.008	20.006	20.007	20.006

Note:

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012622F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Weighing Balance	Make / Model No.	Microtech/MTT
Range	25g to 15 kg	Instrument Condition	OK
Least Count	0.5 g	Location of Calibration	At Site
Serial No.	202301312	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (g)	Deviation (g)	C MAX= Maximum Correction (g)
25g	25.000022	25.0	0.0	1.4
50g	50.000105	50.0	0.0	
100g	100.00025	100.0	0.0	
200g	200.00024	200.0	0.0	
500g	500.0012	500.0	0.0	
1kg	1000.0017	1000.5	0.5	
2kg	2000.0039	2000.5	0.5	
5kg	5000.011	5001.0	1.0	
10kg	10000.082	10001.0	0.9	
15kg	15000.093	15001.5	1.4	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
5000.5	5000.5	5000.0	5000.5	5000.0	5000.0	5000.5	5000.5	5000.0	5000.0	5000.5

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	Standard Dev. (g)	Maximum Error Between Centre and Other Points (g)	Uncertainty (g)



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:		CC205523000012622F										Page 2 of 2	
5kg	5000.5	5000.0	5000.5	5000.0	5000.5	5000.5	5000.5	5000.0	5000.5	5000.0	0.3	0.3	0.9

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012623F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Hot Air Oven	Make / Model No.	-/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	1°C	Location of Calibration	At Site
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/OVEN/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	20	20.12	0.12	0.66
2	40	50.17	10.17	0.66
3	60	60.21	0.21	0.66
4	80	80.24	0.24	0.66
5	100	100.28	0.28	0.66

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000012624F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Elongation Gauge	Make / Model No.	-/-
Range	As per Specification	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/EG/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20 ±2°C, Relative Humidity:- 50 ±10%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: ± μm
1	50	40	81.00	81.01	13.5
2	40	25	58.50	58.52	13.5
3	25	20	40.50	40.49	13.5
4	20	16	32.40	32.39	13.5
5	16	12.5	25.65	25.64	13.5
6	12.5	10	20.25	20.29	13.5
7	10	6.3	14.67	14.66	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012625F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	Instrument Receipt Date	5-Oct-2023
	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Flakiness Gauge	Make / Model No.	-/-
Range	As per Specification	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/FG/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 22.5°C, Relative Humidity:- 54%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: $\pm \mu\text{m}$
1	63	50	33.90	33.91	13.5
2	50	40	27.00	26.99	13.5
3	40	25	19.50	19.49	13.5
4	31.5	25	16.95	16.96	13.5
5	25	20	13.50	13.51	13.5
6	20	16	10.80	10.79	13.5
7	16	12.5	8.55	8.52	13.5
8	12.5	10	6.75	6.76	13.5
9	10	6.3	4.89	4.90	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012626F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Oven	Make / Model No.	itherm/AI-7781
Range	0 to 200 °C	Instrument Condition	OK
Least Count	0.1 °C	Location of Calibration	At Site
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/OVEN/02	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	50.0	50.19	0.19	0.48
2	70.0	70.23	0.23	0.48
3	100.0	100.28	0.28	0.48
4	150.0	150.34	0.34	0.48
5	200.0	200.41	0.41	0.48

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012627F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Water Bath	Make / Model No.	Thermotech/TH-012
Range	Amb. to 100 °C	Instrument Condition	OK
Least Count	0.1°C	Location of Calibration	At Site
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/WB/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	30.0	30.13	0.13	0.48
2	50.0	50.18	0.18	0.48
3	70.0	70.22	0.22	0.48
4	90.0	90.26	0.26	0.48
5	100.0	100.31	0.31	0.48

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205523000012628F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Hot Plate	Make / Model No.	-/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	5 °C	Location of Calibration	At Site
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/HP/07	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	20	20.10	0.10	2.90
2	40	40.17	0.17	2.90
3	60	60.26	0.26	2.90
4	80	80.33	0.33	2.90
5	100	100.41	0.41	2.90

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012629F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Thermometer	Make / Model No.	Multi/ST-9283B
Range	-50 to 300 °C/-58F to 572F	Instrument Condition	OK
Least Count	0.1 °C/0.1F	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/MST/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	-30.0	-30.16	-0.16	0.74
2	30.0	30.14	0.14	0.74
3	100.0	100.25	0.25	0.74
4	200.0	200.36	0.36	0.74
5	300.0	300.43	0.43	1.78

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.: CC205523000012630F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder	Make / Model No.	Lengbon/-
Range	0 to 1000 ml	Instrument Condition	OK
Least Count	10 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/MC/1000ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	100	100.018	-0.018	0.012
2	500	500.032	-0.032	1
3	1000	1000.049	-0.049	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012631F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate Glass/-
Range	0 to 500 ml	Instrument Condition	OK
Least Count	5 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/MC/500ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	50.014	-0.014	0.012
2	250	250.027	-0.027	1
3	500	500.035	-0.035	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
6. Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.: CC205523000012632F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder (Glass)	Make / Model No.	Borosilicate Glass/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/MC/100ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.0009	-0.001	0.012
2	50	50.0015	-0.002	0.012
3	100	100.0021	-0.002	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012655F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder (Glass)	Make / Model No.	Borosilicate Glass/-
Range	0 to 250 ml	Instrument Condition	OK
Least Count	2 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/MC/250ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	50.011	-0.011	0.012
2	150	150.024	-0.024	1
3	250	250.032	-0.032	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012661F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Thermometer	Make / Model No.	Multi/ST-9283B
Range	-50 to 300 °C/-58F to 572F	Instrument Condition	OK
Least Count	0.1 °C/0.1F	Location of Calibration	At Laboratory
Serial No.	-	Location	Bisauli 154+200 LHS
Instrument ID.	HG/MST/02	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	-30.0	-30.13	-0.13	0.74
2	30.0	30.17	0.17	0.74
3	100.0	100.28	0.28	0.74
4	200.0	200.43	0.43	0.74
5	300.0	300.56	0.56	1.78

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Total liability of Testncal Laboratory. will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012662F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Vernier Caliper	Make / Model No.	Insize/-
Range	0 to 300 mm	Instrument Condition	OK
Least Count	0.02mm	Location of Calibration	At Laboratory
Serial No.	1708190520	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C,Relative Humidity: - 50±10%	Calibration Reference Standard	IS 16491 (Part 1 & 2)
		Calibration Procedure	TNC -05-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Slip Gauge Block Set	TNC/M/C/001/15	TI/B/SGS/001/23	CC-2660 (TANSON INSTRUMENT)	02/02/2025

Calibration Results (mm)

Standard Value	Observed Value on UUC (Inner)	Deviation(In ner)	Uncertainty (± μm)	Observed Value on UUC(Outer)	Deviation(O uter)	Uncertainty (± μm)	Observed Value on UUC (Depth)	Deviation (Depth)	Uncertainty (± μm)
10.000	10.00	0.000	12	10.00	0.000	12	10.00	0.000	12
30.000	30.00	0.000	12	30.00	0.000	12	30.00	0.000	12
60.020	60.02	0.000	12	60.02	0.000	12	60.02	0.000	12
90.040	90.04	0.000	12	90.04	0.000	12	90.04	0.000	12
120.900	120.90	0.000	12	120.90	0.000	12	120.90	0.000	12
150.100	150.10	0.000	12	150.10	0.000	12	150.10	0.000	12
200.000	200.00	0.000	12	200.00	0.000	12	200.00	0.000	12
250.000	249.98	-0.020	12	249.98	-0.020	12	249.98	-0.020	12
300.000	299.98	-0.020	12	299.98	-0.020	12	299.98	-0.020	12

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012663F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Vaccum Gauge	Make / Model No.	Rocker/-
Range	0 to -760 mmHg	Instrument Condition	OK
Least Count	25 mmHg	Location of Calibration	At Laboratory
Serial No.	12305194	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Pressure
ENVIRONMENTAL CONDITION	Temperature-24±2°C RH-55±10%	Calibration Reference Standard	DKD-R 6-1
		Calibration Procedure	TNC-04-P 2.0

Calibration Result

Sr.No.	UUC Value (mmHg)	Average of Std. Readings (bar)	Average of Std. Readings (mmHg)	Error (mmHg)	Uncertainty at k=2 at 95 % Confidence Level (±) bar
1	0	0.0000	0.0000	0.0000	0.063
2	-100	-0.1334	-100.0583	0.0583	0.038
3	-200	-0.2668	-200.1165	0.1165	0.038
4	-400	-0.5337	-400.3081	0.3081	0.038
5	-600	-0.8003	-600.2746	0.2746	0.038
6	-700	-0.9337	-700.3329	0.3329	0.038

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012664F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Glass Thermometer	Make / Model No.	Omsons/-
Range	-10 to 50 °C	Instrument Condition	OK
Least Count	0.25 °C	Location of Calibration	At Laboratory
Serial No.	306704	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	-10.00	-10.09	-0.09	0.74
2	20.00	20.14	0.14	0.74
3	30.00	30.16	0.16	0.74
4	40.00	40.19	0.19	0.74
5	50.00	50.22	0.22	0.74

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205523000013462F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Dig. Compression Testing Machine	Make / Model No.	Aimil/-
Range	0 to 2000 KN	Instrument Condition	OK
Least Count	0.1 KN	Location of Calibration	At Site
Serial No.	222545	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 50±10	Calibration Reference Standard	IS 1828(Part1) :2022
		Calibration Procedure	TNC -01-F 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
PROVING RING 2000 KN	1821.2000KN	22011317/D1.05/C-445	CSIR	14/04/2024

Calibration Results

Force(kN)	Division at Std. Temp. 24°C	Division at room Temp. 24.4°C	Observed Reading (Divisions) 0°	Observed Reading (Divisions) 120°	Observed Reading (Divisions) 240°	Average of Reading (Division)	Relative accuracy error (q)%	Relative Repeatability error (b)%	Relative resolution Error (a)%	Relative Zero Error (fo)%	Expanded Uncertainty of Measurement ± %
200	84.8	84.8	85.0	85.0	85.5	85.17	0.44	0.59	0.03	0.00	1.00
400	167.0	167.0	167.0	167.5	167.5	167.33	0.20	0.30	0.01	0.00	1.00
600	249.5	249.5	249.5	249.5	250.0	249.67	0.07	0.20	0.01	0.00	1.00
800	333.3	333.3	333.5	334.0	334.5	334.00	0.21	0.30	0.01	0.00	1.00
1000	417.6	417.6	418.0	418.5	418.5	418.33	0.17	0.12	0.01	0.00	1.00
1200	500.3	500.2	500.5	501.0	501.5	501.00	0.16	0.20	0.00	0.00	1.00
1400	583.7	583.6	584.0	584.5	585.0	584.50	0.15	0.17	0.00	0.00	1.00
1600	669.2	669.1	669.5	669.5	670.5	669.83	0.11	0.15	0.00	0.00	1.00
1800	754.5	754.4	755.0	755.0	756.0	755.33	0.12	0.13	0.00	0.00	1.00
2000	838.2	838.1	839.0	839.0	839.5	839.17	0.13	0.06	0.00	0.00	1.00
0	0	0	0	0	0	0	0	0	0	0	0

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
6. Measurement of uncertainty at k=2, Approximate 95% confidence level.



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013493F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	40 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/40mm/06	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
40	40.13	40.12	40.14	40.12	40.13	40.13	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013494F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	37.5mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/37.5mm/05	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
37.5	37.59	37.58	37.60	37.59	37.61	37.59	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013496F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	31.5mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/31.5 mm/07	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
31.5	31.46	31.45	31.44	31.45	31.46	31.45	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013499F Page 1 of 1

Customer Details M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	Instrument Receipt Date	29-Sep-2023
	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	26.5mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/26.5mm/05	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
26.5	26.57	26.59	26.60	26.58	26.57	26.58	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013503F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	12.5mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/12.5 mm/04	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
12.5	12.61	12.59	12.60	12.62	12.61	12.61	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013508F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	12.5mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/12.5 mm/05	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
12.5	12.47	12.45	12.46	12.45	12.47	12.46	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013509F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	11.2mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/11.2mm/05	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
11.2	11.16	11.17	11.16	11.15	11.17	11.16	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013510F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	5.6mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/5.6 mm/04	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
5.6	5.62	5.63	5.61	5.62	5.63	5.62	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013511F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	4.75mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/4.75mm/06	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
4.75	4.76	4.77	4.76	4.75	4.76	4.76	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013512F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	28-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (300mm Dia)	Make / Model No.	-/-
Range	45mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/45mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
45	45.13	45.14	45.12	45.13	45.14	45.13	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000013513F Page 1 of 1

Customer Details	Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1383
	Date of Calibration	29-Sep-2023
	Next Calibration Date	29-Sep-2024
	Certificate Issue Date	30-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (200mm Dia)	Make / Model No.	-/-
Range	4.75mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/TS/4.75mm/04	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
4.75	4.77	4.78	4.76	4.77	4.78	4.77	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000013516F		Page 1 of 1	
Customer Details		Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1383
		Date of Calibration	29-Sep-2023
		Next Calibration Date	28-Sep-2024
		Certificate Issue Date	30-Sep-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Accelerated Curing Tank	Make / Model No.	-/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	0.1°C	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	HG/ACT/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	20.0	20.11	0.11	0.48
2	40.0	40.14	0.14	0.48
3	60.0	60.17	0.17	0.48
4	80.0	80.21	0.21	0.48
5	100.0	100.24	0.24	0.48

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000013520F		Page 1 of 1	
Customer Details		Instrument Receipt Date	29-Sep-2023
M/S H.G. INFRA ENGINEERING LIMITED, BINAWAR CAMP, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1383
		Date of Calibration	29-Sep-2023
		Next Calibration Date	28-Sep-2024
		Certificate Issue Date	30-Sep-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Cement Mortar Cube Machine	Make / Model No.	-/-
Range	12000±400 rpm	Instrument Condition	OK
Least Count	NA	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	HG/CMCM/01	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC233924000000443F	PRIMA	30/09/2024

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	12000	12014	-14	0.40

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000015857F Page 1 of 1

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/1245
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	11.2 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	HG/TS/11.2mm/02	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
11.2	11.26	11.27	11.25	11.26	11.28	11.26	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000015858F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	-/-
Range	12.5 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	HG/TS/12.5mm/02	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
12.5	12.58	12.57	12.55	12.58	12.56	12.57	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000015859F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Accelerated Curing Tank	Make / Model No.	-/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	1 °C	Location of Calibration	At Site
Serial No.	HG/ACT/01	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	30.0	30.18	0.18	0.66
2	50.0	50.21	0.21	0.66
3	70.0	70.25	0.25	0.66
4	80.0	80.29	0.29	0.66
5	100.0	100.34	0.34	0.66

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000015860F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Glass Hydrometer	Make / Model No.	-/-
Range	1.000 to 1.200 gm/cc	Instrument Condition	OK
Least Count	0.001 gm/cc	Location of Calibration	At Laboratory
Serial No.	HG/GH/01	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Density
ENVIRONMENTAL CONDITION	Temperature:- 25±2°C,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:3104 (Part I & II)
		Calibration Procedure	TNC-05/D/2.0

Calibration Results

S. No.	UUC. Value (Sp.Gr.)	STD. Value(g/ml)	Deviation (Sp. Gr.)	Uncertainty at k=2 ± (g/ml)
1	1.000	0.999	0.001	0.0029
2	1.100	1.098	0.002	0.0029
3	1.200	1.198	0.002	0.0029

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000015861F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/1245
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Stop Watch	Make / Model No.	-/-
Range	0 to 24 hrs	Instrument Condition	OK
Least Count	0.01 sec	Location of Calibration	At Laboratory
Serial No.	HG/SW/01	Location	Bisauli 154+200 LHS
Instrument ID.	-	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 50±10%	Calibration Reference Standard	IS : 5834
		Calibration Procedure	TNC -07- ET 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Time Interval Meter	TNC/ET/C/015	CC294023000000366F	Digital Calibration Technology	13/02/2024

Calibration Results

Sl. No	UUC Value (S)	Std. Value (S)	Error (S)	Uncertainty at k=2 ,± (S)
1	10.00	10.002	0.002	0.61
2	100.00	100.007	0.007	0.66
3	200.00	200.013	0.013	0.69
4	300.00	300.026	0.026	0.73
5	500.00	500.089	0.089	0.87
6	1000.00	1000.14	0.14	1.72
7	2000.00	2000.19	0.19	2.64
8	3600.00	3600.34	0.34	4.35

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205524000003218F Page 1 of 1

Customer Details	Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	SRF No.	TNC1/JC/23-24/2309
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Elongation Gauge	Make / Model No.	-/-
Range	6.3 - 50 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/EG/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20 $\pm$ 2°C, Relative Humidity:- 50 $\pm$ 10%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: $\pm$ μ m
1	50	40	81.00	81.05	13.5
2	40	25	58.50	58.52	13.5
3	25	20	40.50	40.49	13.5
4	20	16	32.40	32.46	13.5
5	16	12.5	25.65	25.71	13.5
6	12.5	10	20.25	20.29	13.5
7	10	6.3	14.67	14.74	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
6. Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205524000003219F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh	Instrument Receipt Date	5-Oct-2023
	SRF No.	TNC1/JC/23-24/2309
	Date of Calibration	5-Oct-2023
	Next Calibration Date	4-Oct-2024
	Certificate Issue Date	6-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Flakiness Gauge	Make / Model No.	-/-
Range	6.3 to 63 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	HG/FG/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 22.5°C, Relative Humidity:- 54%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: $\pm \mu\text{m}$
1	63	50	33.90	33.87	13.5
2	50	40	27.00	26.98	13.5
3	40	25	19.50	19.49	13.5
4	31.5	25	16.95	16.94	13.5
5	25	20	13.50	13.48	13.5
6	20	16	10.80	10.79	13.5
7	16	12.5	8.55	8.49	13.5
8	12.5	10	6.75	6.74	13.5
9	10	6.3	4.89	4.88	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205524000003220F		Page 1 of 1	
Customer Details		Instrument Receipt Date	5-Oct-2023
M/S. H.G. INFRA ENGINEERING LIMITED BASAULI Near Shankar Mandir Chungi Badaun, Site-Bisauli, Uttar Pradesh		SRF No.	TNC1/JC/23-24/2309
		Date of Calibration	5-Oct-2023
		Next Calibration Date	4-Oct-2024
		Certificate Issue Date	6-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Viscosity Bath	Make / Model No.	-/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	0.1°C	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	HG/VB/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	20.0	20.12	0.12	0.48
2	40.0	40.15	0.15	0.48
3	60.0	60.20	0.20	0.48
4	80.0	80.23	0.23	0.48
5	100.0	100.28	0.28	0.48

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By