

Calibration Certificate

Certificate/ULR No.:	CC205523000012361F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Brass Test Sieve (200 mm Dia)	Make / Model No.	Sethi/-
Range	10 mm	Instrument Condition	OK
Least Count	0	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/TS/10mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
10	9.96	9.98	9.97	9.98	9.96	9.97	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012362F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Brass Test Sieve (200 mm Dia)	Make / Model No.	Allied Sales/-
Range	4.75 mm	Instrument Condition	OK
Least Count	0	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/TS/4.75mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
4.75	4.76	4.75	4.76	4.77	4.76	4.76	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012363F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Brass Test Sieve (200 mm Dia)	Make / Model No.	Allied Sales/-
Range	4.75 mm	Instrument Condition	OK
Least Count	0	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/TS/4.75mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
4.75	4.76	4.76	4.75	4.77	4.76	4.76	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012364F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Toyo/-
Range	0 to 100 kg	Instrument Condition	OK
Least Count	0.020 kg	Location of Calibration	At Site
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/WB/100kg/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C RH-55±10%	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
10 kg	10000.040	10.000	0.000	0.081
20 kg	20000.08	20.020	0.020	
30kg	30000.12	30.000	0.000	
40 kg	40000.14	40.000	0.000	
50 kg	50000.20	49.980	-0.020	
60 kg	60000.40	59.980	-0.020	
70 kg	70000.28	69.960	-0.040	
80 kg	80000.364	79.940	-0.060	
90 kg	90000.46	89.940	-0.060	
100 kg	100000.60	99.920	-0.081	

Eccentricity Test @ 20kg

1	2	3	4	5	1	5	4	3	2	1
19.980	20.000	19.980	20.000	19.980	19.980	20.000	19.980	20.000	20.000	19.980



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012364F

Page 2 of 2

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	D= difference between Average readings of Centre and other positions (kg)	Uncertainty (kg)
20 kg	19.980	20.000	19.980	20.000	19.980	20.000	20.000	19.980	20.000	19.980	-0.01	0.095

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Total liability of Testncal Laboratory, will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012365F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Swisser/-
Range	0 to 50 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	2221050	Location	282+200 RHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
100g	100.00025	0.100	0.000	0.011
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.000	0.000	
2kg	1999.9963	2.000	0.000	
5kg	5000.012	5.001	-0.001	
10kg	10000.040	10.003	-0.003	
20kg	20000.08	20.004	-0.004	
30kg	30000.12	30.006	-0.006	
50kg	50000.14	50.011	-0.011	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
20.005	20.004	20.006	20.007	20.005	20.006	20.004	20.005	20.004	20.005	20.007

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
20 kg	20.005	20.006	20.007	20.005	20.007	20.008	20.006	20.005	20.005	20.007

Note:

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012366F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Swisser/-
Range	0 to 50 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	2221047	Location	282+200 RHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
100g	100.00025	0.100	0.000	0.012
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.000	0.000	
2kg	1999.9963	2.000	0.000	
5kg	5000.012	5.002	-0.002	
10kg	10000.040	10.004	-0.004	
20kg	20000.08	20.005	-0.005	
30kg	30000.12	30.008	-0.008	
50kg	50000.14	50.012	-0.012	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
20.008	20.006	20.007	20.008	20.006	20.007	20.005	20.006	20.008	20.007	20.009

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
20 kg	20.006	20.007	20.008	20.008	20.009	20.004	20.007	20.008	20.008	20.009

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012367F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		Instrument Receipt Date	3-Oct-2023
		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
A) Identification On UUC (Unit Under Calibration)		Certificate Issue Date	4-Oct-2023
Instrument Name	Weighing Balance	Make / Model No.	Apolo/-
Range	0 to 30 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/WB/30kg/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
100g	100.00025	0.100	0.000	0.013
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1 kg	999.9992	1.000	0.000	
2 kg	1999.9963	2.001	-0.001	
5 kg	5000.012	5.003	-0.003	
10 kg	10000.040	10.006	-0.006	
20 kg	20000.08	20.009	-0.009	
30 kg	30000.12	30.013	-0.013	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
10.006	10.008	10.006	10.007	10.006	10.008	10.007	10.009	10.006	10.005	10.006

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
10 kg	10.011	10.009	10.008	10.009	10.010	10.011	10.012	10.013	10.0011	10.008

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012413F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Scale-Tec/-
Range	0 to 600 g	Instrument Condition	OK
Least Count	0.01 g	Location of Calibration	At Site
Serial No.	202012189146	Location	282+200 RHS
Instrument ID.	-	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10%	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (g)	Deviation (g)	C MAX= Maximum Correction (g)
1g	1.000001	1.00	0.00	0.09
2g	2.000007	2.00	0.00	
5g	4.999966	5.00	0.00	
10g	10.000025	10.00	0.00	
20g	20.000006	20.00	0.00	
50g	50.000105	50.01	0.01	
100g	100.00025	100.02	0.02	
200g	200.00024	200.05	0.05	
500g	500.0012	500.07	0.07	
600g	600.0015	600.09	0.09	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
200.03	200.02	200.04	200.03	200.02	200.04	200.05	200.03	200.04	200.03	200.05
Standard Dev. (g)				D= difference between Average readings of Centre and other positions (g)				Uncertainty (g)		
0.01				0.01				0.10		

Note: -

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended only for guidance and not for legal purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated. Total liability of Testncl Laboratory Pvt.Ltd. will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012417F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Weighing Balance	Make / Model No.	Apolo/-
Range	0 to 10 kg	Instrument Condition	OK
Least Count	0.5 g	Location of Calibration	At Site
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/WB/10kg/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (g)	Deviation (g)	C MAX= Maximum Correction (g)
20g	20.000056	20.0	0.0	1.0
50g	50.000105	50.0	0.0	
100g	100.00025	100.0	0.0	
200g	200.00024	200.0	0.0	
500g	500.0011	500.0	0.0	
1kg	999.9992	1000.0	0.0	
2kg	1999.9963	2000.5	0.5	
5kg	5000.012	5000.5	0.5	
10kg	10000.027	10001.0	1.0	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
5000.5	5000.5	5000.0	5000.5	5000.0	5000.0	5000.5	5001.0	5000.5	5000.0	5000.0

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	Standard Dev. (g)	Maximum Error Between Centre and Other Points (g)	Uncertainty (g)
5kg	5000.5	5000.5	5001.0	5000.5	5000.5	5000.0	5000.5	5001.0	5000.5	5000.0	0.3	-0.1	0.7



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012417F

Page 2 of 2

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012419F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Vibrating Table	Make / Model No.	-/-
Range	2800 rpm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Site
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/VT/01	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC233924000000443F	PRIMA	30/09/2024

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	2800	2805	-5	0.40

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012421F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Motor Cube Vibrating Machine	Make / Model No.	-/-
Range	12000±400 rpm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Site
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MCVM/01	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC233924000000443F	PRIMA	30/09/2024

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	12000	12015	-15	0.40

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012422F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Hot Plate	Make / Model No.	Allied Sales/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	10 °C	Location of Calibration	At Site
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/HP/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	20	20.13	0.13	5.78
2	40	40.16	0.16	5.78
3	60	60.19	0.19	5.78
4	80	80.22	0.22	5.78
5	100	100.25	0.25	5.78

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012423F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Specific Gravity Weighing Balance	Make / Model No.	Imtex/-
Range	0 to 15 kg	Instrument Condition	OK
Least Count	0.5 g	Location of Calibration	At Site
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/WB/15kg/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (g)	Deviation (g)	C MAX= Maximum Correction (g)
50g	50.000105	50.0	0.0	0.9
100g	100.00025	100.0	0.0	
200g	200.00024	200.0	0.0	
500g	500.0012	500.0	0.0	
1kg	1000.0017	1000.0	0.0	
2kg	2000.0039	2000.0	0.0	
5kg	5000.011	5000.5	0.5	
10kg	10000.082	10000.5	0.4	
15kg	15000.093	15001.0	0.9	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
5000.5	5000.0	5000.5	5000.5	5000.0	5000.5	5000.5	5000.0	5000.0	5000.5	5000.5

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	Standard Dev. (g)	Maximum Error Between Centre and Other Points (g)	Uncertainty (g)
5000	5000.0	5000.5	5000.0	5000.5	5000.0	5000.5	5000.0	5000.5	5000.0	5000.5	0.3	0.2	0.7



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012423F

Page 2 of 2

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012424F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Flakiness Gauge	Make / Model No.	-/-
Range	As per Instrument	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/FG/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 22.5°C, Relative Humidity:- 54%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: $\pm \mu\text{m}$
1	63	50	33.90	33.91	13.5
2	50	40	27.00	26.99	13.5
3	40	25	19.50	19.49	13.5
4	31.5	25	16.95	16.94	13.5
5	25	20	13.50	13.49	13.5
6	20	16	10.80	10.81	13.5
7	16	12.5	8.55	8.56	13.5
8	12.5	10	6.75	6.76	13.5
9	10	6.3	4.89	4.90	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000012425F Page 1 of 1

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Flakiness Gauge	Make / Model No.	-/-
Range	6.3 to 63 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/FG/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 22.5°C, Relative Humidity:- 54%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: $\pm \mu\text{m}$
1	63	50	33.90	33.88	13.5
2	50	40	27.00	26.99	13.5
3	40	25	19.50	19.51	13.5
4	31.5	25	16.95	16.96	13.5
5	25	20	13.50	13.49	13.5
6	20	16	10.80	10.81	13.5
7	16	12.5	8.55	8.49	13.5
8	12.5	10	6.75	6.74	13.5
9	10	6.3	4.89	4.88	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012426F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Elongation Gauge	Make / Model No.	-/-
Range	As per Instrument	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/EG/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20 $\pm$ 2°C, Relative Humidity:- 50 $\pm$ 10%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: $\pm$ μ m
1	50	40	81.00	81.02	13.5
2	40	25	58.50	58.52	13.5
3	25	20	40.50	40.49	13.5
4	20	16	32.40	32.41	13.5
5	16	12.5	25.65	25.67	13.5
6	12.5	10	20.25	20.26	13.5
7	10	6.3	14.67	14.69	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
6. Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012427F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Elongation Gauge	Make / Model No.	-/-
Range	As per Instrument	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/EG/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20 $\pm$ 2°C, Relative Humidity:- 50 $\pm$ 10%	Calibration Reference Standard	-
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Sr. No.	Passing Through IS Sieve (mm)	Retained on IS Sieve (mm)	Thickness of gauge (mm)	Observed Value of Thickness Gauge (mm)	Measurements of Uncertainty approximate 95% confidence level at k=2: $\pm$ μ m
1	50	40	81.00	81.01	13.5
2	40	25	58.50	58.49	13.5
3	25	20	40.50	40.51	13.5
4	20	16	32.40	32.41	13.5
5	16	12.5	25.65	25.67	13.5
6	12.5	10	20.25	20.24	13.5
7	10	6.3	14.67	14.68	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205523000012428F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Compression Testing Machine	Make / Model No.	Aimil/-
Range	0 to 2000 KN	Instrument Condition	OK
Least Count	0.1 KN	Location of Calibration	At Site
Serial No.	222528	Location	282+200 RHS
Instrument ID.	-	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 50±10	Calibration Reference Standard	IS 1828(Part1) :2022
		Calibration Procedure	TNC -01-F 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
PROVING RING 2000 KN	1821.2000KN	22011317/D1.05/C-445	CSIR	14/04/2024

Calibration Results

Force(kN)	Division at Std. Temp. 24°C	Division at room Temp. 24.4°C	Observed Reading (Division s) 0°	Observed Reading (Division s) 120°	Observed Reading (Division s) 240°	Average of Reading (Division)	Relative accuracy error (q)%	Relative Repeatab ility error (b)%	Relative resolutio n Error (a)%	Relative Zero Error (fo)%	Expende d Uncertain ty of Measure ment ± %
200	84.8	84.8	85.0	85.0	85.5	85.17	0.44	0.59	0.03	0.00	1.00
400	167.0	167.0	167.0	167.5	167.5	167.33	0.20	0.30	0.01	0.00	1.00
600	249.5	249.5	249.5	249.5	250.0	249.67	0.07	0.20	0.01	0.00	1.00
800	333.3	333.3	333.5	334.0	334.5	334.00	0.21	0.30	0.01	0.00	1.00
1000	417.6	417.6	418.0	418.5	418.5	418.33	0.17	0.12	0.01	0.00	1.00
1200	500.3	500.2	500.5	501.0	501.5	501.00	0.16	0.20	0.00	0.00	1.00
1400	583.7	583.6	584.0	584.5	585.0	584.50	0.15	0.17	0.00	0.00	1.00
1600	669.2	669.1	669.5	669.5	670.5	669.83	0.11	0.15	0.00	0.00	1.00
1800	754.5	754.4	755.0	755.0	756.0	755.33	0.12	0.13	0.00	0.00	1.00
2000	838.2	838.1	839.0	839.0	839.5	839.17	0.13	0.06	0.00	0.00	1.00
0	0	0	0	0	0	0	0	0	0	0	0

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
6. Measurement of uncertainty at k=2, Approximate 95% confidence level.



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012429F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital MultistemThermometer	Make / Model No.	Multi/-
Range	-50 to 300 °C	Instrument Condition	OK
Least Count	0.1 °C	Location of Calibration	At Laboratory
Serial No.	73325662	Location	282+200 RHS
Instrument ID.	-	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	-50.0	-50.16	-0.16	0.74
2	50.0	50.18	0.18	0.74
3	100.0	100.31	0.31	0.74
4	200.0	200.44	0.44	0.74
5	300.0	300.52	0.52	1.78

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012430F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital MultistemThermometer	Make / Model No.	Multi/-
Range	-50 to 300 °C	Instrument Condition	OK
Least Count	0.1 °C	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MST/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	-50.0	-50.13	-0.13	0.74
2	50.0	50.14	0.14	0.74
3	100.0	100.25	0.25	0.74
4	200.0	200.37	0.37	0.74
5	300.0	300.46	0.46	1.78

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012432F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.004	-0.004	0.012
2	50	50.009	-0.009	0.012
3	100	100.013	-0.013	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012433F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/02	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.006	-0.006	0.012
2	50	50.011	-0.011	0.012
3	100	100.016	-0.016	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012434F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/03	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	9.993	0.007	0.012
2	50	49.988	0.012	0.012
3	100	99.982	0.018	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012435F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/04	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	9.996	0.004	0.012
2	50	49.991	0.009	0.012
3	100	99.984	0.016	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012436F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/05	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.008	-0.008	0.012
2	50	50.014	-0.014	0.012
3	100	100.021	-0.021	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012437F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/06	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.009	-0.009	0.012
2	50	50.017	-0.017	0.012
3	100	100.022	-0.022	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012438F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/07	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	9.992	0.008	0.012
2	50	49.984	0.016	0.012
3	100	99.979	0.021	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.: CC205523000012439F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/08	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.005	-0.005	0.012
2	50	50.012	-0.012	0.012
3	100	100.018	-0.018	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012440F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/09	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.008	-0.008	0.012
2	50	50.017	-0.017	0.012
3	100	100.026	-0.026	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012441F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/10	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	9.995	0.005	0.012
2	50	49.987	0.013	0.012
3	100	99.978	0.022	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012442F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/11	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.003	-0.003	0.012
2	50	50.009	-0.009	0.012
3	100	100.015	-0.015	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012443F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/12	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	9.997	0.003	0.012
2	50	49.990	0.010	0.012
3	100	99.983	0.017	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012444F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/13	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.006	-0.006	0.012
2	50	50.014	-0.014	0.012
3	100	100.020	-0.020	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012445F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/14	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	10.008	-0.008	0.012
2	50	50.018	-0.018	0.012
3	100	100.026	-0.026	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.: CC205523000012446F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 100 ml	Instrument Condition	OK
Least Count	1 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/100ml/15	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	10	9.995	0.005	0.012
2	50	49.987	0.013	0.012
3	100	99.976	0.024	0.012

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012447F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 500 ml	Instrument Condition	OK
Least Count	5 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/500ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	50.012	-0.012	0.012
2	250	250.029	-0.029	1
3	500	500.047	-0.047	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012449F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 500 ml	Instrument Condition	OK
Least Count	5 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/500ml/02	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	50.016	-0.016	0.012
2	250	250.038	-0.038	1
3	500	500.053	-0.053	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Total liability of Testncal Laboratory. will be limited to the invoiced amount only.
6. Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012450F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 500 ml	Instrument Condition	OK
Least Count	5 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/500ml/03	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	49.983	0.017	0.012
2	250	249.971	0.029	1
3	500	499.947	0.053	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012451F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 500 ml	Instrument Condition	OK
Least Count	5 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/500ml/04	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	49.986	0.014	0.012
2	250	249.973	0.027	1
3	500	499.944	0.056	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012452F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 1000 ml	Instrument Condition	OK
Least Count	10 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/1000ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	100	100.068	-0.068	0.012
2	500	500.113	-0.113	1
3	1000	1000.149	-0.149	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.: CC205523000012453F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 1000 ml	Instrument Condition	OK
Least Count	10 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/1000ml/02	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	100	100.053	-0.053	0.012
2	500	500.119	-0.119	1
3	1000	1000.232	-0.232	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.: CC205523000012455F		Page 1 of 1	
Customer Details		Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123		SRF No.	TNC1/JC/23-24/1218
		Date of Calibration	3-Oct-2023
		Next Calibration Date	2-Oct-2024
		Certificate Issue Date	4-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 1000 ml	Instrument Condition	OK
Least Count	10 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/1000ml/03	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	100	99.967	0.033	0.012
2	500	499.892	0.108	1
3	1000	999.806	0.194	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012456F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 1000 ml	Instrument Condition	OK
Least Count	10 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/1000ml/04	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	100	99.972	0.028	0.012
2	500	499.886	0.114	1
3	1000	999.791	0.209	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012457F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 250 ml	Instrument Condition	OK
Least Count	2 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/250ml/01	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	50.011	-0.011	0.012
2	150	150.023	-0.023	1
3	250	250.029	-0.029	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012458F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 250 ml	Instrument Condition	OK
Least Count	2 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/250ml/02	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	49.987	0.013	0.012
2	150	149.979	0.021	1
3	250	249.968	0.032	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012459F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	3-Oct-2023
	Next Calibration Date	2-Oct-2024
	Certificate Issue Date	4-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 250 ml	Instrument Condition	OK
Least Count	2 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/250ml/03	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Digital Weighing Balance (0 to 1kg)	TNC/M/C/003/19	CC205523000000002F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	49.985	0.015	0.012
2	150	149.973	0.027	1
3	250	249.961	0.039	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P 17 F03 00

Certificate/ULR No.:	CC205523000012460F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	3-Oct-2023
M/S H.G. INFRA ENGINEERING LTD. Vill-Semarjhala (Rupapur) - Khasra No:- 732,734 Teh. & PO- Sawayajpur, Thana-Pali Near Bhadauriya Dhaba, Allahganj- Bilram Road Dist. Hardoi, Uttar Pradesh Pin Code : 241123	SRF No.	TNC1/JC/23-24/1218
	Date of Calibration	4-Oct-2023
	Next Calibration Date	3-Oct-2024
	Certificate Issue Date	5-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Measuring Cylinder	Make / Model No.	Borosilicate/-
Range	0 to 250 ml	Instrument Condition	OK
Least Count	2 ml	Location of Calibration	At Laboratory
Serial No.	-	Location	282+200 RHS
Instrument ID.	HG/MC/250ml/04	Parameter	Volume
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS/ISO 4787 : 2010
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Weighing Balance (upto 220g)	TNC/M/C/001/17	CC205523000000001F	Testncal	01/01/2024

Calibration Results

SI No	Nominal Volume (ml)	Measured mean volume (ml)	Deviation (ml)	Uncertainty at k=2 ± (ml)
1	50	49.987	0.013	0.012
2	150	149.978	0.022	1
3	250	249.966	0.034	1

Result at Reference Temperature at 27°C

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Totalliability of Testncal Laboratory. will be limited to the invoiced amount only.
- 6.Measurement of uncertainty approximate 95% confidence level at k=2



Calibrated By

Issued & Approved By