

Calibration Certificate

Certificate No.	CALCERT-20260626-545A78	Job No.	JOB-HTTP-001
Client	AA Electro Magnetic Test Laboratory Private Limited	Calibration Date	2026-06-26
Address	Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon, Haryana	Next Calibration Date	2027-06-26
Instrument	Digital Multimeter	Issue Date	2026-06-26
Discipline	Electro Technical	Serial No.	NEW-SN-HTTP
Range	As per Manual	Least Count	As per Range

Retrieved Historical Data

Source ULR	CC405825000000872F	Source Date	4-Feb-2025
Confidence	0.952 (HIGH)	Source Pages	30-31
Reference Standard	IS:1248-2003	Procedure	TNC -03-WI 2.0
Environment	Temperature:- 24 +/- 2 deg C,Relative Humidity: - 55 +/- 10%		

Master Equipment

Name	ID / Serial	Certificate	Certified By	Valid Upto
6 1/2 Digit DMM	TNC/ET/C/001/15	AIPL/220125/2-1	Anshankan	25/01/2026

Calibration Results Copied From Historical Dataset

Calibration Table 1

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 +/- %
1	100uA	10uA	uA	10	9.9964	0.90
2		50uA	uA	50	49.9973	0.90

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 +/- %
3		99uA	uA	99	98.9955	0.90
4	100mA	10mA	mA	10	9.99858	0.30
5		50mA	mA	50	49.9828	0.30
6		99mA	mA	99	98.9917	0.30
7	3A	1A	A	1	0.999782	0.30
8		2A	A	2	1.99972	0.30
10	10A	5A	A	5	4.99791	0.30
11		9A	A	9	8.9945	0.30

Calibration Table 2

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 +/- %
1	4V	2V	V	2	1.99960	0.10
2		4V	V	4	3.99878	0.10
3	40V	20V	V	20	19.9953	0.10
4		40V	V	40	39.9885	0.10
5	400V	200V	V	200	199.934	0.10
6		400V	V	400	399.874	0.10
7	1000V	500V	V	500	499.818	0.10
8		900V	V	900	899.763	0.10

Calibration Table 3

Column 1	Column 2	Column 3	Column 4	Column 5	Column 6	Column 7
10		9A	A	9	8.99934	0.26

Calibration Table 4

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 +/- %
1	4V	2V	V	2	2.00014	0.21
2		4V	V	4	4.00024	0.21
3	40V	20V	V	20	20.0011	0.21
4		40V	V	40	40.0042	0.21
5	400V	200V	V	200	200.015	0.21
6		400V	V	400	400.047	0.21
7	1000V	500V	V	500	499.983	0.21
8		900V	V	900	899.942	0.21

Calibration Table 5

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 +/- %
1	400	400 ohm	ohm	400	399.9928	0.06
2	4	4k ohm	k ohm	4	3.999830	0.00
3	40	40k ohm	k ohm	40	39.99931	0.01
4	400	400k ohm	k ohm	400	399.9925	0.06
5	4	4M ohm	M ohm	4	3.999790	0.00
6	40	40M ohm	M ohm	40	39.99873	0.01

Calibration Table 6

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 +/- %
1	20Hz	10Hz	Hz	10	10.00012	0.82
2		19Hz	Hz	19	18.00015	0.82
3	100Hz	50Hz	Hz	50	50.0012	0.82
4		99Hz	Hz	99	99.0016	0.82
5	1kHz	0.5kHz	kHz	0.5	0.5000014	0.82

Calibration Table 7

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 +/- %
1	5uF	1uF	uF	1	0.998	4.13
2		4uF	uF	4	3.997	4.13
3	50uF	10uF	uF	10	10.02	4.13
4		49uF	uF	49	49.03	4.13
5	100uF	70uF	uF	70	69.98	4.13
6		99uF	uF	99	98.8	4.13

Audit: Calibration values marked as historical were retrieved from source certificate CC405825000000872F and copied into this new draft certificate for engineer review.