

# Calibration Certificate

|                 |                                                               |                       |                  |
|-----------------|---------------------------------------------------------------|-----------------------|------------------|
| Certificate No. | CALCERT-20260627-BC745B                                       | Job No.               | JOB-20260627-001 |
| Client          | AA Electro Magnetic Test Laboratory Private Limited           | Calibration Date      | 2026-08-23       |
| Address         | Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon, Haryana | Next Calibration Date | 2029-06-26       |
| Instrument      | Digital Multimeter                                            | Issue Date            | 2026-06-27       |
| Discipline      | Electro Technical                                             | Serial No.            | NEW-SN-001       |
| Range           | As per Manual                                                 | Least Count           | As per Range     |

## Retrieved Historical Data

|                    |                                                              |              |                |
|--------------------|--------------------------------------------------------------|--------------|----------------|
| Source ULR         | CC405825000000872F                                           | Source Date  | 4-Feb-2025     |
| Confidence         | 0.952 (HIGH)                                                 | Source Pages | 30-31          |
| Reference Standard | IS:1248-2003                                                 | Procedure    | TNC -03-WI 2.0 |
| Environment        | Temperature:- 24 +/- 2 deg C,Relative Humidity: - 55 +/- 10% |              |                |

## Master Equipment

| Name            | ID / Serial     | Certificate     | Certified By | Valid Upto |
|-----------------|-----------------|-----------------|--------------|------------|
| 6 1/2 Digit DMM | TNC/ET/C/001/15 | AIPL/220125/2-1 | Anshankan    | 25/01/2026 |

## Calibration Results Copied From Historical Dataset

**Calibration Table 1**

| Sl. No | Range | Range | UOM | Observed Value on UUC | Observed Value on Standard | Uncertainty at k=2 +/- % |
|--------|-------|-------|-----|-----------------------|----------------------------|--------------------------|
| 1      | 100uA | 10uA  | uA  | 10                    | 9.9964                     | 0.90                     |
| 2      |       | 50uA  | uA  | 50                    | 49.9973                    | 0.90                     |

| Sl. No | Range | Range | UOM | Observed Value on UUC | Observed Value on Standard | Uncertainty at k=2 +/- % |
|--------|-------|-------|-----|-----------------------|----------------------------|--------------------------|
| 3      |       | 99uA  | uA  | 99                    | 98.9955                    | 0.90                     |
| 4      | 100mA | 10mA  | mA  | 10                    | 9.99858                    | 0.30                     |
| 5      |       | 50mA  | mA  | 50                    | 49.9828                    | 0.30                     |
| 6      |       | 99mA  | mA  | 99                    | 98.9917                    | 0.30                     |
| 7      | 3A    | 1A    | A   | 1                     | 0.999782                   | 0.30                     |
| 8      |       | 2A    | A   | 2                     | 1.99972                    | 0.30                     |
| 10     | 10A   | 5A    | A   | 5                     | 4.99791                    | 0.30                     |
| 11     |       | 9A    | A   | 9                     | 8.9945                     | 0.30                     |

**Calibration Table 2**

| Sl. No | Range | Range | UOM | Observed Value on UUC | Observed Value on Standard | Uncertainty at k=2 +/- % |
|--------|-------|-------|-----|-----------------------|----------------------------|--------------------------|
| 1      | 4V    | 2V    | V   | 2                     | 1.99960                    | 0.10                     |
| 2      |       | 4V    | V   | 4                     | 3.99878                    | 0.10                     |
| 3      | 40V   | 20V   | V   | 20                    | 19.9953                    | 0.10                     |
| 4      |       | 40V   | V   | 40                    | 39.9885                    | 0.10                     |
| 5      | 400V  | 200V  | V   | 200                   | 199.934                    | 0.10                     |
| 6      |       | 400V  | V   | 400                   | 399.874                    | 0.10                     |
| 7      | 1000V | 500V  | V   | 500                   | 499.818                    | 0.10                     |
| 8      |       | 900V  | V   | 900                   | 899.763                    | 0.10                     |

**Calibration Table 3**

| Column 1 | Column 2 | Column 3 | Column 4 | Column 5 | Column 6 | Column 7 |
|----------|----------|----------|----------|----------|----------|----------|
| 10       |          | 9A       | A        | 9        | 8.99934  | 0.26     |

**Calibration Table 4**

| Sl. No | Range | Range | UOM | Observed Value on UUC | Observed Value on Standard | Uncertainty at k=2 +/- % |
|--------|-------|-------|-----|-----------------------|----------------------------|--------------------------|
| 1      | 4V    | 2V    | V   | 2                     | 2.00014                    | 0.21                     |
| 2      |       | 4V    | V   | 4                     | 4.00024                    | 0.21                     |
| 3      | 40V   | 20V   | V   | 20                    | 20.0011                    | 0.21                     |
| 4      |       | 40V   | V   | 40                    | 40.0042                    | 0.21                     |
| 5      | 400V  | 200V  | V   | 200                   | 200.015                    | 0.21                     |
| 6      |       | 400V  | V   | 400                   | 400.047                    | 0.21                     |
| 7      | 1000V | 500V  | V   | 500                   | 499.983                    | 0.21                     |
| 8      |       | 900V  | V   | 900                   | 899.942                    | 0.21                     |

**Calibration Table 5**

| Sl. No | Range | Range    | UOM   | Observed Value on UUC | Observed Value on Standard | Uncertainty at k=2 +/- % |
|--------|-------|----------|-------|-----------------------|----------------------------|--------------------------|
| 1      | 400   | 400 ohm  | ohm   | 400                   | 399.9928                   | 0.06                     |
| 2      | 4     | 4k ohm   | k ohm | 4                     | 3.999830                   | 0.00                     |
| 3      | 40    | 40k ohm  | k ohm | 40                    | 39.99931                   | 0.01                     |
| 4      | 400   | 400k ohm | k ohm | 400                   | 399.9925                   | 0.06                     |
| 5      | 4     | 4M ohm   | M ohm | 4                     | 3.999790                   | 0.00                     |
| 6      | 40    | 40M ohm  | M ohm | 40                    | 39.99873                   | 0.01                     |

**Calibration Table 6**

| Sl. No | Range | Range  | UOM | Observed Value on UUC | Observed Value on Standard | Uncertainty at k=2 +/- % |
|--------|-------|--------|-----|-----------------------|----------------------------|--------------------------|
| 1      | 20Hz  | 10Hz   | Hz  | 10                    | 10.00012                   | 0.82                     |
| 2      |       | 19Hz   | Hz  | 19                    | 18.00015                   | 0.82                     |
| 3      | 100Hz | 50Hz   | Hz  | 50                    | 50.0012                    | 0.82                     |
| 4      |       | 99Hz   | Hz  | 99                    | 99.0016                    | 0.82                     |
| 5      | 1kHz  | 0.5kHz | kHz | 0.5                   | 0.5000014                  | 0.82                     |

**Calibration Table 7**

| Sl. No | Range | Range | UOM | Observed Value on UUC | Observed Value on Standard | Uncertainty at k=2 +/- % |
|--------|-------|-------|-----|-----------------------|----------------------------|--------------------------|
| 1      | 5uF   | 1uF   | uF  | 1                     | 0.998                      | 4.13                     |
| 2      |       | 4uF   | uF  | 4                     | 3.997                      | 4.13                     |
| 3      | 50uF  | 10uF  | uF  | 10                    | 10.02                      | 4.13                     |
| 4      |       | 49uF  | uF  | 49                    | 49.03                      | 4.13                     |
| 5      | 100uF | 70uF  | uF  | 70                    | 69.98                      | 4.13                     |
| 6      |       | 99uF  | uF  | 99                    | 98.8                       | 4.13                     |

Audit: Calibration values marked as historical were retrieved from source certificate CC405825000000872F and copied into this new draft certificate for engineer review.