

Calibration Certificate

Certificate No.	CALCERT-20260629-04A51E	ULR No.	
Job No.	VERIFY-20260629-001	Issue Date	2026-06-29
Client	CalCert Verification Client	Calibration Date	2026-06-29
Address	Gurugram, Haryana	Next Calibration	2027-06-29
Instrument	Tachometer	Serial No.	TACH-VERIFY-001
Make / Model	Luolider / UT372	Discipline	Speed
Range	0 to 99999 rpm	Least Count	1 rpm
Procedure	TNC -02-RPM 2.0	Environment	temperature: 23.4, humidity: 51, pressure: 1012

Calibration Table 1

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 , +/- %
1	100.0	100.09	-0.09	0.611231
2	500.0	500.19	-0.19	1.15578
3	5000	5001.1	-1.1	10.027499
4	15000	15002.966667	-2.966667	30.035996
5	35000	35006	-6	70.070461

Reference Standards and Traceability

Calibration reference standard	IS: 12508
Historical structure source	CC405825000000210F, pages 10-10
Uncertainty statement	All uncertainty values were recalculated from current measurements.
Uncertainty model	Speed current-measurement model v1.0

Approved by engineer. End of certificate.