

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CALCERT-20260701-D73FBB		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Jul-2026
M/S. Browser Test Laboratory Gurugram, Haryana		SRF No.	BROWSER-TEST-20260701
		Date of Calibration	1-Jul-2026
		Next Calibration Date	1-Jul-2027
		Certificate Issue Date	1-Jul-2026
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Tachometer	Make / Model No.	Luolider/UT372
Range	0 to 99999 rpm	Instrument Condition	SATISFACTORY
Least Count	1 rpm	Location of Calibration	At Site
Serial No.	TACH-BROWSER-001	Location	-
Instrument ID.		Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 23.5 degC, Relative Humidity: 50%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC380824000000839F	CC-3808 (REALTECH)	30/09/2026

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	100.0	100.09	-0.09	0.61126
2	500.0	500.19	-0.19	1.15617
3	5000	5001.1	-1.1	10.031988
4	15000	15002.966667	-2.966667	30.049482
5	35000	35006	-6	70.101933

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By