Color calibration for smartphone-based anaemia detection

"even if the eye photo is taken there would be a problem where the lightning might not be good and the photo colour can change . so to address the issue we can place a Printed colour reference card so that the image of the refence card's photo can be taken as comparison for the lightning or brighnesss present there and accordingly the eye images can also be altered" *Generated by LeapSpace*

Yes — a **printed colour reference card** is a sensible way to reduce lighting and device-related colour variation in smartphone eye-image anaemia screening, because several sources show that **colour calibration improves consistency** and can outperform simple raw-image approaches [67] [68] [69] [70] [71].

What the evidence supports

Approach	What it does	What the evidence says
Colour reference card / colour chart	Provides a fixed reference in the image so colours can be calibrated against known values	A colourcard approach gave the **best results** for device-independent smartphone colourimetry, and a separate mobile-app study used a reference colour pattern for automatic colour calibration [70] [71].
Ambient subtraction / white balancing	Adjusts for changing illumination without needing a card	These methods can work well and are used in anaemia-related smartphone imaging, but results still vary across populations and lighting conditions [65] [69] [72].
Illumination normalization in software	Corrects image brightness/colour after capture	Anaemia studies have used colour correction, ambient subtraction, and brightness-adjustment algorithms to improve robustness across phones and light conditions [61] [67] [73].

How this applies to your idea

Your idea is **technically consistent with the literature**: include a printed reference card in the photo, then use it to estimate lighting/colour shifts and correct the eye image before classification [70] [71]. That is an **inference** from the calibration studies, but it is well supported by the way smartphone colourimetry papers handle device and ambient-light variation [67] [68] [69].

Important nuance

One anaemia study in pregnant women found that **no specialist hardware such as a colour reference card was required** for image acquisition, and instead relied on software methods like white balancing and ambient subtraction [72]. So the evidence does **not** show that a colour card is mandatory; rather, it shows that it is **one strong calibration option** among several [70] [72].

Bottom line

A printed colour reference card could **improve robustness** when eye photos are taken under uneven lighting or different phone cameras, and it is a defensible design choice for a smartphone anaemia-screening app [69] [70] [71]. But it should be treated as a **calibration aid**, not a guaranteed fix, because some anaemia-screening studies achieve acceptable performance without extra hardware by using software-based illumination correction [67] [72] [73].

Confidence level: **High** — multiple sources directly support colour calibration for smartphone imaging, while one anaemia study explicitly shows an alternative no-card workflow, so the overall conclusion is strong but not exclusive [69] [70] [71] [72].

References

61. Ghosal, S., Das, D., Udutalapally, V., Wasnik, P.N. (2022) INAP: A Hybrid Approach for NonInvasive Anemia-Polycythemia Detection in the IoMT. *ACM Transactions on Computing for Healthcare*. https://doi.org/10.1145/3503466

65. Wemyss, T.A., Rana, A., Hillman, S.L., (...), Leung, T.S. (2025) Diagnosing anaemia via smartphone colorimetry of the eye in a population of pregnant women. *Physiological Measurement*. https://doi.org/10.1088/1361-6579/adab4d

67. Sakthivel, Haripriya, Park, Sang Mok, Kwon, Semin, (...), Kim, Young L (2025) Machine learning of blood haemoglobin and haematocrit levels via smartphone conjunctiva photography in Kenyan pregnant women: a clinical study protocol. *BMJ Open*. https://doi.org/10.1136/bmjopen-2024-097342

68. Bao, X., Jiang, S., Wang, Y., (...), Han, J. (2018) A remote computing based point-of-care colorimetric detection system with a smartphone under complex ambient light conditions. *Analyst*. https://doi.org/10.1039/c7an01685a

69. Nixon, M., Outlaw, F., Leung, T.S. (2020) Accurate device-independent colorimetric measurements using smartphones. *PLoS ONE*. https://doi.org/10.1371/journal.pone.0230561

70. Nixon, M., Outlaw, F., MacDonald, L.W., Leung, T.S. (2019) The importance of a device specific calibration for smartphone colorimetry. *Final Program and Proceedings - IS and T/SID Color Imaging Conference*. https://doi.org/10.2352/issn.2169-2629.2019.27.10

71. Fletcher, R., Pignatelli, N., Jimenez-Galindo, A., Ghosh-Jerath, S. (2016) Development of smart phone tools for printed diagnostics: Challenges and solutions. *GHTC 2016 - IEEE Global Humanitarian Technology Conference: Technology for the Benefit of Humanity, Conference Proceedings*. https://doi.org/10.1109/GHTC.2016.7857355

72. Wemyss, T.A., Nixon-Hill, M., Outlaw, F., (...), Leung, T.S. (2023) Feasibility of smartphone colorimetry of the face as an anaemia screening tool for infants and young children in Ghana. *PLoS ONE*. https://doi.org/10.1371/journal.pone.0281736

73. Huamaní, E.L., Montufar-Barrientos, E., Romero-Alva, V., Velasquez-Jimenez, L. (2026) A Reproducible Mobile-Cloud Framework for Preliminary Non-Invasive Anemia Screening from Palpebral Conjunctiva Images. *International Journal of Advanced Computer Science and Applications*. https://doi.org/10.14569/ijacsa.2026.0170804