



Customer Information

Joinn Bio
2700 Hilltop Drive
joinn@bio.com
925 123 4567

Served By

Falco Automation, LLC
350 Industrial Rd, San Carlos, CA 94070
Tel: (650) 449-9011
www.falcoautomation.com

Equipment Information

Equipment Id JNB-TM-WM-01
Equipment Manufacturer Mettler Toledo
Model Number MT-700
Serial Number 7412253
Tolerance 1.0
Unit kg

Equipment Information

Calibration Method Weigt
Calibration Procedure Direct
Calibration Tool Id FAC-WT-001, FAC-WT-002
Manufacturer Weight. Co
Model Number WT-001, WT-002
Serial Number 12345, 12346

Calibration Results

NaN

Setpoint	As Found	As Left
0.0	0.0	0.0
50.0	49.6	49.6
100.0	89.9	96.6

Statements of Pass or Fail Conformance: The uncertainty of the measurement has been considered when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accepts does not exceed {guard percent} in compliance with {standard}

Field Environmental Conditions: Temperature: between 17 and 23 degrees Celsius; rate of change not to exceed 1.0 C per hour. Relative humidity: from 30 to 70%. These conditions satisfy NIST requirements.

Comments

Calibration Technician Petro Khalil Technician Signature _____ Date of Calibration 2023-06-08T05:34:18.421Z
Reviewed By Reviewer Signature _____ Calibration Interval -----
Calibration Due Date -----
Date Cert Issued -----

IMPORTANT: The "Calibration valid through" date shown above has been specified to us by this client. Falco Automation, LLC takes no position regarding the suitability or appropriateness of this interval.

Calibration Report Status:

PASS- Term used when compliance statement is given, and the measurement result is within accepted criteria.

FAIL- Term used when compliance statement is given, and the measurement result falls out of accepted criteria.

ADJUSTED- When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value. as left