



### Customer Information

Joinn Bio  
2700 Hilltop Drive  
support@joinnbio.com  
123 456 7890

### Serviced By

Falco Automation, LLC  
350 Industrial Rd, San Carlos, CA 94070  
Tel: (650) 449-9011  
www.falcoautomation.com

### Equipment Information

Equipment Id JNB-05560  
Equipment Manufacturer Mettler Toledo  
Model Number M700-23  
Serial Number 2045ss465f  
Tolerance 1.0

### Equipment Information

Calibration Procedure  
Calibration Tool Id  
Manufacturer  
Model Number  
Serial Number

### Calibration Results

#### {Procedure} – {Method}

Setpoint	As Found	As Left
0	0	0
50	50.1	50.1
100	89.9	99.4

Statements of Pass or Fail Conformance: The uncertainty of the measurement has been considered when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accepts does not exceed {guard percent} in compliance with {standard}

Field Environmental Conditions: Temperature: between 17 and 23 degrees Celsius; rate of change not to exceed 1.0 C per hour. Relative humidity: from 30 to 70%. These conditions satisfy NIST requirements.

### Comments

Calibration Technician	Petro Khalil	Technician Signature	_____	Date of Calibration	2023-06-08T05:12:19.796Z
				Calibration Interval	-----
Reviewed By		Reviewer Signature	_____	Calibration Due Date	-----
				Date Cert Issued	-----

IMPORTANT: The "Calibration valid through" date shown above has been specified to us by this client. Falco Automation, LLC takes no position regarding the suitability or appropriateness of this interval.

### Calibration Report Status:

PASS- Term used when compliance statement is given, and the measurement result is within accepted criteria.

FAIL- Term used when compliance statement is given, and the measurement result falls out of accepted criteria.

ADJUSTED- When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value. as left