

Multi-Imager Compatible, MR Safe, Remote Center of Motion Needle-Guide Robot

Dan Stoianovici^{1b}, Changhan Jun, Sunghwan Lim^{1b}, Pan Li, Doru Petrisor, Stanley Fricke, Karun Sharma, and Kevin Cleary

Abstract—We report the development of a new robotic system for direct image-guided interventions (DIGI; images acquired at the time of the intervention). The manipulator uses our previously reported pneumatic step motors and is entirely made of electrically nonconductive, nonmetallic, and nonmagnetic materials. It orients a needle-guide with two degrees of freedom (DoF) about a fulcrum point located below the guide using an innovative remote center of motion parallelogram type mechanism. The depth of manual needle insertion is preset with a third DoF, located remotely of the manipulator. Special consideration was given to the kinematic accuracy and the structural stiffness. The manipulator includes registration markers for image-to-robot registration. Based on the images, it may guide needles, drills, or other slender instruments to a target ($OD < 10 \text{ mm}$). Comprehensive preclinical tests were performed. The manipulator is MR safe (ASTM F2503-13). Electromagnetic compatibility (EMC) testing (IEC 60601-1-2) of the system shows that it does not conduct or radiate EM emissions. The change in the signal to noise ratio of the MRI due to the presence and motion of the robot in the scanner is below 1%. The structural stiffness at the needle-guide is 33 N/mm. The angular accuracy and precision of the manipulator itself are 0.177° and 0.077° . MRI-guided targeting accuracy and precision *in vitro* were 1.71 mm and 0.51 mm, at an average target depth of $\sim 38 \text{ mm}$, with no adjustments. The system may be suitable for DIGI where $1.07 + 0.0031 \cdot d \text{ [mm]}$ accuracy lateral to the needle (2D) or $1.59 + 0.0031 \cdot d \text{ [mm]}$ in 3D is acceptable. The system is also multi-imager compatible and could be used with other imaging modalities.

Index Terms—Compatible, direct IGI, DIGI, MR safe, multi-imager, motor, pneumatic, robot, remote center of motion RCM.

Manuscript received February 15, 2017; revised March 15, 2017; accepted April 16, 2017. Date of publication April 25, 2016; date of current version December 20, 2017. This work was supported in part by award 1R01CA172244 from the National Cancer Institute, in part by award RC1EB010936 from the National Institute of Biomedical Imaging and Bioengineering, and in part by award W81XWH0810221 from the Department of Defense. (Corresponding author: Dan Stoianovici.)

D. Stoianovici is with the Urology Department, Robotics Laboratory, Johns Hopkins University, Baltimore, MD 21218 USA (e-mail: dss@jhu.edu).

C. Jun, S. Lim, P. Li, and D. Petrisor are with the Urology Department, Robotics Laboratory, Johns Hopkins University.

S. Fricke, K. Sharma, and K. Cleary are with the Children's National Health System, Sheikh Zayed Institute for Pediatric Surgical Innovation.

This paper contains supplemental material available online at <http://ieeexplore.ieee.org> (File size: 800 Kb).

Digital Object Identifier 10.1109/TBME.2017.2697766

I. INTRODUCTION

IMAGE-GUIDED interventions (IGI) may be performed based on pre-acquired images, but using the imager at the time of the intervention to guide the procedure directly provides updated feedback and improves quality control [1], [2]. But Direct-IGI (DIGI) often require special devices such as medical robots to help the physician and improve the accuracy [3], [4]. Compatibility of robotic devices with the imaging equipment has been a challenging engineering task [5], especially with the MR environment [6], [7].

The American Society for Testing and Materials (ASTM) standard F2503-13 [8] classifies devices for the MR environment. In the United States this is mandated by the Food and Drug Administration (FDA) for clinical use of Significant Risk (SR) devices (21 CFR 812, [9]). The highest class of compatibility of devices with the MR environment is MR Safe. These are devices “that poses no known hazards resulting from exposure to any MR environment. MR Safe items are composed of materials that are electrically nonconductive, nonmetallic and nonmagnetic” [8].

Traditional active robotic components are not MR Safe. But in the last decade researchers developed and adapted several actuation and sensing components that are MR Safe. Examples of electricity-free manipulators built with a minimal amount of nonferrous metals are [10], [11]. A few MR Safe manipulators have already been built [12]–[16] and some meet the ASTM F2503 classification [16]–[18].

All current MR Safe robots are pneumatically actuated, either with purposely-built pneumatic cylinders [12], turbines [13], [19], or stepper motors [16], [18]. These were commonly powered through long hoses by pneumatic valves, either voice-coil type located outside the MRI room (ACR Zone III [20]) [18], or piezoelectric type valves located in the MR scanner room (ACR Zone IV) [16]. Position sensors have been typically built with code-wheel quadrature encoding and optical sensors, that emit an optical signal through a flexible fiber and recapture it through another fiber. The fiber sensors were not built for the MR environment, but were adapted from harsh-environment (wet, corrosive, explosive) conditions.

As described in this paper, by using several purposely made and adapted components, the development of MR Safe manipulators is no longer a daunting challenge, but actually similar in complexity to traditional robot manipulators.

Several essential advantages are also derived from being able to make “plastic” manipulators. First, a manipulator made entirely of electrically nonconductive, nonmetallic and non-magnetic materials may be determined to be MR Safe by providing a scientifically based rationale rather than test data” [8]. This facilitates FDA approvals and clinical translation. Second, this construction also makes the medical device more versatile, since it works within any MR environment. Third, since the MR environment imposes the highest compatibility requirements, a manipulator that is MR Safe and passes electromagnetic compatibility (EMC) tests is compatible with other imaging equipment based on the scientific rational, in other words is Multi-Imager Compatible [5]. This enables its use for various applications and even in cross-modality DIGI, for example PET-CT (positron emission tomography - computed tomography) and MRI.

The manipulator that we present includes a Remote Center of Motion (RCM) kinematic structure. Since its invention in 1995 [21], this has become the most common kinematic component of medical robots. It allows a medical instrument to pivot about a fulcrum point that is not occupied by the robot (the “remote” in RCM), and is typically placed at the skin entry point/port of numerous minimally invasive interventions/surgeries. In fact, from the kinematics perspective the RCM has become homologous to minimally invasive.

Many RCM types have since been developed [22], including parallelogram bar [23] (the original RCM) and belt drive mechanisms [24], [25]. The latest daVinci robots (Intuitive Surgical, Sunnyvale, CA) use the belt type. The RCM reported here is a parallelogram bar type mechanism with an original joint arrangement which facilitates the mounting of the medical instrument and offers improved clearance relative to the patient. This was designed to guide not only slender needles but also larger diameter instruments such as bone biopsy cannulas and drills. While the values are unknown, the forces exerted by the larger instruments on the needle-guide in bony applications are expectedly higher than those encountered for slender needle insertion into soft tissue. Therefore, high structural stiffness was an important consideration in the new RCM design.

Our robot is not the first RCM type robot for the MRI. A parallelogram type with an interesting curved link structure that goes around the head for stereotactic neurosurgery was built [26]. This robot would probably be considered MR Conditional since it includes few metallic components and uses piezoelectric motors. Piezoelectric (ultrasonic) motors have significant SNR loss [27]. A custom controller that reduces substantially the loss of SNR to less than 15% in motion was also developed [26].

This paper presents the technical details of our new robot: the new RCM mechanism, the design of the manipulator, and a comprehensive set of tests. The conclusions include lessons learned that we hope will inform other researchers in the field.

II. NON-COLLINEAR PARALLELOGRAM RCM

Basic mechanical RCM mechanisms present 2 rotary DoF that intersect at the fulcrum (RCM) point. However, only one of these mechanisms requires special kinematics, since the other can be




Fig. 1. 1-DoF parallelogram RCM types: (a) classic – collinearly aligned joints, (b) non-collinear vertical links, (c) non-collinear.

simply implemented with a revolute joint whose axis passes the fulcrum. Therefore, the RCM mechanisms are commonly considered as a 1-DoF planar mechanism that implements the actual remote revolute axis.

The 1-DoF mechanism of the original RCM [21] is schematically represented in Fig. 1(a). This is a 6-bar linkage with 8 joints. Here, several of the links and joints are redundant. For example, the kinematics of the mechanism is identical if either link A_1A_2 , A_1B_1 , or both A_2B_2 and B_1B_2 are missing. Therefore, the mechanism uses at least 2 parallelogram 4-bar structures to implement a virtual parallelogram OAA_1O_1 , which makes link AB rotate about the remote point O , the RCM point. The remote axis of rotation is normal to the plane of the mechanism at the RCM point. The use of the redundant parallelogram requires precise manufacturing, but improves structural stiffness.

The points A , B , and O are collinear. Since the instrument (I) must cross the point O , it should be angled relative to AB as shown in the figure, either in the mechanism plane or another. Therefore, the end-effector tool/instrument (I) is attached (G , guide) to the link AB and passes the point O , as shown in Fig. 1(a). An example of a robot with the instrument angled is the LARS robot [28] of the original RCM patent [21]. Another option is to place the instrument I in a parallel plane to clear the interference. An example of a robot with this arrangement is the previously mentioned RCM robot built for the MRI [26].

A variation of the parallelogram RCM mechanism is presented in Fig. 1(b). Here, the vertical links $O_1A_1B_1$ and $O_2A_2B_2$ have been “bent” so that their joints are non-collinear. As such, the virtual parallelogram is OBB_1O_1 . This makes the RCM point O non-collinear with AB , allowing more flexibility in mounting the instrument. With this, the instrument may be mounted parallel to AB in the RCM plane ($OI \parallel AB$).

Finally, the non-collinear joint arrangement can also be applied to links AA_1A_2 and BB_1B_2 . As shown in Fig. 1(c), this helps to further displace the RCM point in the other direction. A variation of this with link $O_2A_2B_2$ straight, $O_1A_1B_1$ and AA_1A_2 bent and no redundant $A_1A_2B_2B_1$ parallelogram was constructed for the Steady-Hand Eye Robot [29].

For our application, high structural stiffness was desired to withstand inadvertent forces exerted on the needle-guide by the




Fig. 2. Robot RCM and actuation schematic.




Fig. 3. Manipulator design.

procedure needle/drill. Since force levels required for clinical use have not been quantified, the design goal was to create a structure as stiff as possible within the MR Safe and size constraints in the MR gantry. Our solution was to use short RCM links, a low OA/OB ratio (relatively short distance to the RCM point), and involve the redundant RCM parallelograms.

The kinematic diagram of the RCM mechanism devised for the robot is shown in Fig. 2. This mechanism uses a vertically non-collinear joint arrangement, the redundant parallelogram, relatively short distance to the RCM point, and the needle-guide centered on the joint B. The non-collinear arrangement enabled the bottom joint A to be located aside of the needle-guide. The advantages of this unique arrangement will be presented in the design section and verified experimentally in the results.

The parallelogram mechanism generates the primary RCM rotation R_1 . This is driven by a spinning-screw (J, revolute joint L) sliding-nut (K) drive, linked with revolute joints to B_1 . The other RCM rotation R_2 is implemented by rotating the base of the parallelogram mechanism about an axis that passes the RCM point (coaxial with O_1O_2 for design convenience), as shown in Fig. 2. The 2-DoF are driven by motors located in the actuation module which includes a harmonic drive for R_2 .

III. THE ROBOT MANIPULATOR

The manipulator design, analysis, and manufacturing codes were built in Creo (PTC, Inc., Needham, MA). The manipulator consists of the 2-DoF RCM module, an actuation module, a passive arm with 3 degrees of adjustment (DoA), and a mount that latches onto the channels of the MRI table, as shown in Fig. 3. The arm and mount are made for a long bone biopsy procedure. Other procedures may require different mounts. The




Fig. 4. RCM module detached from the actuation base: (abc) isometric and orthogonal views, (d) link AB, needle-guide, and latching mechanism.

position of the robot and respectively its RCM point can be adjusted with 4-DoA, sliding the base on the table channels (A_1) and adjusting the positions of the arm (A_{2-4}). These are locked by tightening the respective knobs.

For versatility, the manipulator is constructed with a modular structure, so that the actuation module could be reused to drive other robots. This construction is similar to our previously developed robot for prostate biopsy [16]. It consists of two coaxially placed PneuStep motors ($\varnothing 70 \text{ mm} \times 25 \text{ mm}$, $4^\circ/\text{step}$, 2 encoder counts/step) [30] and a harmonic drive (49:1 transmission ratio) driven by one of the motors (see Fig. 2).

The RCM module detached from the actuation module is shown in Fig. 4. The guide is centered on the top joint B by designing the RCM mechanism with a wide structure on both sides of the guide. Joint B is therefore implemented by two coaxial joints BB on the sides, as shown in Fig. 4(b). Here, if the classic collinear RCM mechanism were used instead, the bottom joints AA would have to be similarly spaced apart, and the angular clearance with respect to the patient would have been γ_1 [see Fig. 4(b)]. However, the non-collinear design, enabled the two joints A to be located closer together, increasing the clearance to $\gamma_2(40^\circ) > \gamma_1(20^\circ)$. In the other direction, clearance with respect to the patient is increased by angling the entire device with angle $\alpha(20^\circ)$, as shown in Fig. 4(c). The attached animation shows the R_1 motion.

The range of the R_1 rotation is -50° to $+30^\circ$ implemented with a custom ACME $\varnothing 10 \times 2 \text{ mm}$ pitch screw. The step size varies between 0.029–0.044°/step, depending on the position of the RCM mechanism. The R_2 rotation is unrestricted by the joint, but effectively bounded by the interference with the clinical site ($\pm\gamma_2 = \pm 40^\circ$ on a flat surface). Its resolution is 0.082°/step.

For sterilization purposes, the entire device can be covered with a sterile bag (Universal Medical EZ-3030). The only




Fig. 5. Depth driver to preset the depth of needle insertion.




Fig. 6. Photo of the robot shown with bone biopsy needle.

component to be sterilized is the needle-guide. The needle-guide can be easily detached by pressing a button [see Fig. 4(d)]. It also includes a pin that pierces the bag and attaches to link AB, which has a concave structure that follows the outer geometry of the needle-guide to provide a secure and stiff attachment when the latch of the button engages the pin.

Needles or instruments with diameters up to 10 mm can be guided by the robot by constructing needle-guides with the corresponding bore. The needle used in the current design (see Fig. 3) is the Invivo 15100 (Schwerin, Germany) MRI bone biopsy set with a cannula diameter of 4.9 mm.

While the needle is manually inserted through the guide, the maximum depth of insertion may be preset by the robot before the insertion. A depth limiter consisting of a rubber O-ring is placed over the barrel of the needle (see Figs. 5 and 6). The needle is inserted through the guide until the ring bottoms on the top surface of the guide. A separate device (1 remote DoF) is used to adjust the location of the ring on the needle barrel prior to its insertion.

The depth presetting device is schematically represented in Fig. 5. This consists of a third identical PneuStep motor equipped with a central nut (ACME Ø19 × 4.5 mm pitch). The nut drives a screw with 3 flat sides spaced at 120°. Six rollers placed in the casing support the screw to implement a prismatic joint. This axis (T) adjusts the location of a limit ring (O-Ring) on the barrel of the needle. The limit ring is initially placed on the barrel close to the point. The needle is then inserted in the device until its point bottoms on the needle limiter, while the limit ring slides on the barrel up to the desired preset depth. The maximum depth of adjustment is 100 mm with a resolution of 0.05 mm/step. These should satisfy numerous clinical applications.




Fig. 7. System diagram showing placement of MR safe, conditional, and unsafe components.

A set of 4 registration markers has been built into the BB₂ link of the RCM module, as shown in Figs. 4(a) and 6. These are made of glass tubes, filled with Radiance MRI imaging liquid (Beekley, Bristol, CT).

Two development prototypes of the RCM mechanism were assembled of mostly 3D printed components. Even for an experienced design team, we found these prototypes to be helpful in the design refinement process along with the CAD models. Computer Numerically Controlled (CNC) manufacturing was then used for the final prototype (except for 7 3D-printed small cover pieces). The manipulator (see Fig. 6) and depth presetting device including the motors and encoders are entirely built of electrically nonconductive, nonmetallic, and nonmagnetic materials. These materials are plastics (ABS, Acetal Copolymer, Delrin, Nylon 6, Peek 1000, Polycarbonate, Polyetherimide (Ultem 1000), Polyimide, Polyethersulfone (Radel), PTFE (Teflon)), rubber, composites (Garolite, Torlon Polyamide-imide), glass, and high-alumina ceramic. In addition to mechanical property considerations, material selection included high electrical resistivity and dielectric strength considerations. The links of the RCM are built of Garolite and Ultem, the pins of the joints are made of Garolite rods, with tightly fitted (local interference H7/p6 tolerance classification) Delrin bushings. The needle-guide, which is the only part that comes in direct contact with the patient was built of material with ISO-10993 biocompatible certification (Radel BL033) and withstands steam sterilization at 135 °C.

IV. THE ROBOTIC SYSTEM

A schematic of the robotic system in the MR environment is presented in Fig. 7. The control system is an updated version of our prostate biopsy robot [16]. The Robot Controller is a PC running Windows 10 (Microsoft Corp.) and equipped with a motion control card (MC4000, PMDI, Victoria, BC, Canada). The Controller is MR Unsafe and must remain outside the ACR (American College of Radiology) Zone IV (scanner room), in the ACR Zone III (control room or equipment room) [31].

An electro-pneumatic and electro-optical Interface to the 3 PneuStep motors is located in the scanner room (ACR

Zone IV), but outside the 0.5 mTesla (5 Gauss) line. The interface is nonmagnetic but uses electricity. For each of the 3 axes, it includes a PneuStep driver, 3 piezoelectric valves (Hoerbiger PS10021-641A, Germany), and 2 fiber optic sensors (D10BFP, Banner Engineering, Minneapolis, MN). The R_1 axis of the robot has no limit switch, and instead a hardware stop is used to home it. The Axis R_2 has no limit and uses an arbitrary home. The needle depth Axis T uses an additional fiber optic sensor for homing. The interface also includes an emergency button and status LEDs.

The Depth Driver is conveniently mounted on top of the Interface box. The Interface is connected with 3 m long hoses to the robot and with DB25 cables (through the access panel waveguide or its RF Shield) to the Controller. The air supply is typically taken from the medical air port in the MR room, but could be passed through the waveguide as well.

Motion control, safety, and robot kinematics are implemented on the Robot Controller PC under C++ (Visual Studio 2015, Microsoft Corp.) using libraries of the MCC. Safety features include a watchdog, emergency stop buttons, and visual alerts. The watchdog (built on hardware and software) checks the state of several components of the system at 6 Hz, immediately disabling power to the pneumatic valves of the electro-pneumatic interface, should a faulty condition occur.

Images are transferred over the network in DICOM format (Digital Imaging and Communications in Medicine) to the Image Registration & Navigation controller implemented on another PC. Image-to-robot registration of the set of markers of the RCM is used to register the robot and MRI spaces. Our previous imaging programs [16], [18] were based on the Amira Visualization software (Visage Imaging, San Diego, CA). for this robot, C++ programs were developed using the open source VTK and ITK toolkits and the GDICM open library for DICOM processing. A custom image-to-model registration algorithm similar to the one that we presented for our transperineal biopsy MrBot [18] robot was used. In short, a CAD model of the markers in the robot coordinate system is superimposed over the MR image of the markers to derive the registration matrix. This then enables the conversion of points between the two coordinate systems, so that points selected in the image are converted to robot coordinates and then joint angles based on the inverse kinematics of the robot.

V. ROBOT TESTS

Extensive bench, imaging, and targeting tests of the robot have been conducted. Errors were calculated as the vector difference between the actual (measured) and set point (target) datasets. Accuracy and precision results were calculated based on the mean and standard deviation of the error vectors. Tests in the MRI were performed in a Siemens Magnetom Aera.

A. Joins Space Set Point Testing

The first tests were performed to verify the positioning ability of the 3 axes of the robot one at a time. The axes were commanded to move over their entire range of motion with 5° increments for R_1 , R_2 and 5 mm for the depth axis T.




Fig. 8. Robot tests with an optical tracker.

Measurements were acquired with a protractor (0.25° resolution) and a caliper (0.01 mm resolution). Each experiment was repeated 3 times.

B. RCM Check and Robot Space Set Point Testing

An optical tracker (Polaris, NDI, Canada) was used to measure the position of an optical marker on a slender rod placed through the needle-guide of the robot (at ~ 208 mm from the RCM point), as shown in Fig. 8. The manufacturer stated RMS error of the Polaris is 0.35 mm. Measurements were performed close to the tracker, with the marker at rest, and 500 repeated measurements were averaged at each static location. For these measurement conditions, the accuracy of this optical tracker may be as low as 0.055 mm [32].

The R_1 axis was commanded to move to $N_1 = 15$ positions (-45° to 25° with 5° increments), and for each the R_2 axis was moved to $N_2 = 16$ positions (0° to 180° with 12° increments). The actual position of the marker in tracker coordinates, ${}^T a_{ij} \mid i = 1, 2, \dots, N_1, j = 1, 2, \dots, N_2$, was measured at all these 240 positions. Since R_1 and R_2 are the axes of the RCM mechanism, the set of points ${}^T a$ should lie on a sphere. As such, the set was fitted to a sphere using a Gauss-Newton least squares algorithm (in Matlab, MathWorks, Inc.). The center of the sphere is ${}^T O$. The length of the rod from the RCM point to the center of the marker (L) was estimated as the radius of the sphere. Having L and the joint axes positions, allowed the calculation of the set point dataset in robot coordinates ${}^R s_{ij} \mid i = 1, 2, \dots, N_1, j = 1, 2, \dots, N_2$ through the direct kinematics. Then, a rigid point-cloud registration (F) between the robot and tracker spaces, ${}^T a = {}^R F \cdot {}^R s$ was derived using Horn's method [33]. The registration allowed transformations of the datasets between the robot and tracker spaces: ${}^T s = {}^R F \cdot {}^R s$ and ${}^R a = {}^T F \cdot {}^T a$. Robot positioning errors were then calculated as the difference between the actual and set point datasets in the same frame ${}^R e = {}^R a - {}^R s$. The error of the RCM location is the fitted sphere center in the robot coordinates ${}^R O = {}^T F \cdot {}^T O$. The angular error of the robot was calculated in degrees as $\varepsilon = 180 {}^R e / (L\pi)$. For clinical purposes, the estimated targeting error that may be expected due to the robot is $e_R = {}^R e / L$ for each mm of target depth.




Fig. 9. Repeatability test points in joint space.




Fig. 10. Stiffness test setup in a milling machine.

C. Set Point Repeatability Testing

Using the same optical tracker setup, the robot was sent to 9 set positions at $R_1 = \{-15^\circ, 0^\circ, 15^\circ\}$ and $R_2 = \{-15^\circ, \{0^\circ, 15^\circ\}\}$. Each position was approached on each axis from $\pm 5^\circ$ of the position, as shown in Fig. 9. The experiment was repeated 5 times for a total of $9 \times 4 \times 5 = 180$ tests. Tracker measurements were acquired at the 9 set point positions. The angular errors were calculated as in the previous test. The repeatability was calculated as the standard deviation of the angular errors.

D. Stiffness Testing

The robot was mounted in the vise of a vertical milling machine as shown in Fig. 10. A compression spring (McMaster Carr 9657K145, zinc-plated steel, 3.125" L, 0.75" OD, 0.091" wire diameter, calibrated elastic constant 4.123 N/mm) was mounted in the spindle of the machine (stopped, no spin). A feed of the machine (0.01 mm resolution) compressed the spring against the needle-guide in a direction passing the RCM point, to generate the desired load. A micrometer (0.01 mm resolution) was placed on the opposite side to measure the deflection under the load. Experiments were performed in axial (-Y) and lateral (X) directions to the guide. Forces between 0 and 35 N were applied in 20 values and the corresponding deflections were recorded. The force-deflection data was then fitted linearly.

e. MR Safe Testing

The robot was tested according to ASTM F2503-13 and its derived test standards [8, Table X1.1]. The scientific rational




Fig. 11. (a) Robot and mockup within MRI coils, (b) the mockup (grid slots 10 mm, grid wall 1 mm, top grid height 5 mm, distance between top and bottom grids 30 mm).

was first employed based on the component materials of the manipulator. The robot was then tested for the Static magnetic field-induced displacement force (ASTM F2052) and torque (ASTM F2213) by suspending the manipulator at the entrance of the MR system bore by its hoses and by a string and measuring the hose/string deflection angle and observing the torque. The sting supported device was moved with the table through the entire bore while inspecting the possible motion of the device due to induced effects. The RF field-induced heating (ASTM F2182, ISO TS 10974) and Gradient field-induced heating and vibration (ISO TS 10974) were performed by direct observation. The Gradient and RF field-induced voltage tests (ISO TS 10974) do not apply to dielectric materials. Similarly, malfunction of the device due to the static, gradient, or RF fields (ISO TS 10974) does not apply. However, it was explored during the EMC testing and DIGI targeting tests presented next. Evaluation of the MR images according to ASTM F2119 is designed for implants, and was not possible (it was not advisable to immerse the robot in a solution). Instead, alternative tests were used as shown in the following sections. The Interface controller is designed to remain outside the 0.5 mTesla (5 Gauss) line. This was carefully brought in the MR room while observing possible induced forces. The system components were labeled according to ASTM F2503-13 (see Figs. 6 and 7).

F. Electromagnetic Compatibility (EMC) Testing

Although the manipulator is immune to electromagnetic disturbances from the MR scanner due to its MR Safe materials, there is the potential for electromagnetic (EM) interference from the control components of the robot which are electric. The Interface is connected with a shielded DB25 cable to the Robot Controller (see Fig. 6) either directly through the waveguide or passing through the RF filter of the access panel. The waveguide connection is preferable because it is independent of the RF filter implementation, and is portable across imaging modalities. For EMC testing this connection provides the reasonably foreseeable maximum of the EM disturbances.

The same mockup was used in all tests involving the MR. This is a geometric mockup designed in the shape of a rectangular basin with square grids at the top and the bottom, and with linear MR markers that can be used for registration [see Fig. 11(b)].

TABLE I
TEST CONFIGURATIONS WITH AND WITHOUT THE ROBOT

Test with	Description
No Robot (N)	Mockup at MR isocenter with MR coils in place, manipulator and Interface controller outside of the MR shield and unpowered.
Robot Active (A)	Mockup and coils unchanged from N, manipulator in operative position about the MR isocenter, and the Interface controller outside the 5-Gauss line. All system components connected, powered, robot in ready to move state but at rest.
Robot in Motion (M)	Same as A-test, but with the robot in continuous motion.

Note: We found that it is preferable to perform the tests in the reversed order, because it is easier to remove rather than placing components without changing the mockup and coils position.

The mockup was 3D printed of plastic and filled with a solution of 0.5% Magnevist (Bayer HealthCare LLC) in water by volume. The markers were filled with the Radiance contrast, as used in the robot markers. The mockup was mounted with a support on the MR table and positioned at the bore isocenter.

EMC testing followed the IEC 60601-1-2 [34] standard, which the FDA recognizes as a consensus standard [35]. Reasonably foreseeable RF emissions may be transmitted by the Interface controller to the MR environment and induced by the MR environment into the Interface and subsequently retransmitted into the RF antenna of the MRI. The white pixel or noise test is a test commonly performed at the installation of the MR scanner and its room shield, and then performed during service and maintenance of the MR to measure EM noise levels in the MR environment. Here, we used the test to measure the EM emissions augmented by the robotic system. The test is commonly available on MR scanners and consists of a series of MR sequences run with the RF transmitter of the MR turned off. Instead, the RF signals that the MR antennas capture are originating from external sources, from improper shielding of the MR room, or, in our case from the robotic system. To separate the causes, three test setups were used, as shown in Table I.

Six noise tests were performed in each of the 3 test configurations. The base line test (N) establishes the normal level of noise in the MR environment. The other two are used to determine the additional noise caused by the presence of the robotic system (A) and its motion (M) relative to the no robot case (N).

G. Image Quality Testing

The tests are based on the methods presented in [5], [36], updated in [16], and are further refined herein. The N, A, and M test conditions described in Table I were used. Images were acquired over the entire mockup using a T1 (Short T1 Inversion Recovery) and a T2 sequence with coronal slices [ZX plane in Fig. 11(a)]. Their sequence parameters were maintained throughout all the tests. Coronal images were chosen as being normal to the typical direction of needle insertion for this robot/procedure. All images were acquired with two Body 18 coils, Siemens Healthcare [see Fig. 11(a)]. For each sequence type, we acquired three N image sets, two A sets, and two M sets. For all tests it is essential that the MR coils are not moved. These should be placed so that

they do not interfere with the robot and its motion, and that the robot may be placed or removed without moving the coils.

The tests were designed to establish quantitative image deterioration metrics due to the robot active (A) relative to the N case, and due to its motion (M) relative to the A case. Images were analyzed for: 1) Signal to noise ratio (SNR) changes, 2) Image change factors, and 3) Geometric artifacts.

G1. SNR Testing: The SNR is calculated based on the NEMA standard MS 1-2008 [37], Method 1. A region of interest (ROI) is defined within signal-containing images centered over the mockup. The same ROI is used in all tests. The SNR calculation requires two equivalent image sets $\{I, J\}$. For each slice $s : 1 \rightarrow n^{ROI}$ of the ROI, the signal S is calculated as the average pixel intensity values p^{Is} in one of the sets. The noise N is calculated based on the standard deviation of the pixel-by-pixel difference image between the two sets:

$$\begin{aligned} S_{\{I\}}^s &= \text{Avrg}_{\{u,v\}:ROI} \left[p_{(u,v)}^{Is} \right] \\ N_{\{I,J\}}^s &= \text{StDev}_{\{u,v\}:ROI} \left[p_{(u,v)}^{Is} - p_{(u,v)}^{Js} \right] / \sqrt{2} \\ SNR_{\{I,J\}}^s &= \frac{S_{\{I\}}^s}{N_{\{I,J\}}^s} \end{aligned} \quad (1)$$

The SNR is calculated for each of the ROI slices, for the following pairs of sets: {N1, N3}, {N2, N3}, {A1, A2}, {M1, M2}. The relative changes in SNR are:

$$\begin{aligned} \Delta SNR_{\{N-N\}}^s &= 100 \frac{SNR_{\{N1, N3\}}^s - SNR_{\{N2, N3\}}^s}{SNR_{\{N2, N3\}}^s} [\%] \\ \Delta SNR_{\{A-N\}}^s &= 100 \frac{SNR_{\{A1, A2\}}^s - SNR_{\{N2, N3\}}^s}{SNR_{\{N2, N3\}}^s} [\%] \\ \Delta SNR_{\{M-A\}}^s &= 100 \frac{SNR_{\{M1, M2\}}^s - SNR_{\{A1, A2\}}^s}{SNR_{\{A1, A2\}}^s} [\%] \end{aligned} \quad (2)$$

Equation (2) gives the three ΔSNR^s values for each slice of the ROI. These are plotted versus the image space coordinate of the slices, separately for the T1 and T2 sequences. The results are also averaged over the slices, for each sequence type, to determine the global metrics:

$$\begin{aligned} \Delta SNR_{\{N-N\}} &= \text{Avrg}_{s:1 \rightarrow n^{ROI}} [\Delta SNR_{\{N-N\}}^s] \\ \Delta SNR_{\{A-N\}} &= \text{Avrg}_{s:1 \rightarrow n^{ROI}} [\Delta SNR_{\{A-N\}}^s] \\ \Delta SNR_{\{M-A\}} &= \text{Avrg}_{s:1 \rightarrow n^{ROI}} [\Delta SNR_{\{M-A\}}^s] \end{aligned} \quad (3)$$

As such, $\Delta SNR_{\{N-N\}}$ quantifies the normal change in SNR between consecutively acquired image sets, in the absence of the robot to be tested, and sets the non-interference goal. The $\Delta SNR_{\{A-N\}}$ quantifies the SNR changes caused by the robot presence. Last, $\Delta SNR_{\{M-A\}}$ quantifies the component of SNR changes caused by the robot motion.

G2. Image Change Factors: The SNR tests do not account for possible image artifacts induced by the presence and motion of the robot. Image change tests quantify the differences between the A-N and M-A sets. The N1, N2, N3, A1, and M1 sets from both the T1 and T2 sequences are used. There is no need to reacquire these, since they are available from the SNR tests. The ROI is the same one used in the SNR.

Image change (deterioration) metrics [36] are reciprocal to image similarity metrics commonly used in image-based registration [38]. Here, we use the image deterioration factors described in [5]. These are updated to reflect the combined test with the SNR and to increase the sensitivity.

For a pair of image sets $\{I, J\}$, where I and J are the tested respectively the reference images, an Image Change Factor (ICF) can be calculated for each pair of slices $s : 1 \rightarrow n^{ROI}$ within the ROI as:

$$ICF_{\{I, J\}}^s = -100 \frac{\text{Avrg}_{\{u, v\}:ROI} \left[|p_{(u,v)}^{Is} - p_{(u,v)}^{Js}| \right]}{\text{Avrg}_{\{u, v, s\}:ROI} \left[p_{(u,v)}^{Js} \right]} [\%] \quad (4)$$

where the numerator is the average of the absolute values of the pixel-by-pixel difference between the pixel intensities form the 2 sets. The denominator is the average of the pixel intensities over the entire ROI. With this, $ICF_{\{N_2, N_3\}}^s$ quantifies the normal variability between consecutively acquired image sets, in the absence of the robot, and gives an estimate of the non-change goal. The ICF differences relative to it are:

$$\begin{aligned} \Delta ICF_{\{N - N\}}^s &= ICF_{\{N_1, N_3\}}^s - ICF_{\{N_2, N_3\}}^s \\ \Delta ICF_{\{A - N\}}^s &= ICF_{\{A_1, N_3\}}^s - ICF_{\{N_2, N_3\}}^s \\ \Delta ICF_{\{M - A\}}^s &= ICF_{\{M_1, A_1\}}^s - ICF_{\{N_2, N_3\}}^s \end{aligned} \quad (5)$$

Equation (5) gives ΔICF^s values for each slice of the ROI. These are plotted versus the image space coordinate of the slices, for T1 and T2. The results are averaged over the slices, for each sequence type, to determine the global metrics:

$$\begin{aligned} \Delta ICF_{\{N - N\}} &= \text{Avrg}_{s:1 \rightarrow n^{ROI}} \left[\Delta ICF_{\{N - N\}}^s \right] \\ \Delta ICF_{\{A - N\}} &= \text{Avrg}_{s:1 \rightarrow n^{ROI}} \left[\Delta ICF_{\{A - N\}}^s \right] \\ \Delta ICF_{\{M - A\}} &= \text{Avrg}_{s:1 \rightarrow n^{ROI}} \left[\Delta ICF_{\{M - A\}}^s \right] \end{aligned} \quad (6)$$

Similar to the SNR tests, $\Delta ICF_{\{A - N\}}$ quantifies the changes caused by the robot presence, and $\Delta ICF_{\{M - A\}}$ quantifies the component of the image changes caused by the robot motion alone. As in the SNR tests, negative values represent a loss.

G3. Geometric Artifacts: The geometric accuracy of the images is critical for the accuracy of image-guided targeting using the robot. The ICF test above is sensitive to A-N geometric artifacts. However, it is unlikely but possible that a geometric artifact would not be detected by the ICF (i.e. a shift equal to the pattern increment of a checker pattern image). Moreover, the ICF does not depict artifacts of the N set, of the MRI itself.




Fig. 12. The robot and mockup attached to the MRI table for the image-guided targeting tests.

For this, an additional visual inspection of the imaged shape of the grid of the mockup is helpful.

The mockup [see Fig. 11(b)] was instrumented with a set of image-markers that are similar to those of the robot. One of the N image sets (that contains the mockup markers) is used. The mockup markers are used to register the CAD model of the mockup to the image. The mockup-to-image registration is performed similar to the robot-to-image registration described in Section IV. Based on the registered model, a grid of lines that corresponds to the grid of the mockup is overlaid over the images. Visual inspection is then used to observe if the lines overlap the imaged grid. If observable, the misalignment could be quantified as a measure of the geometric artifacts.

H. Image-Guided Targeting

The same mockup was used. A new set of images (T1) was acquired to include the robot markers as well as the mockup (see Fig. 12). The robot and the mockup were registered to the image space. An additional marker (similar construction) was initially placed through the needle guide to image the guide direction.

The clinical MRI bone biopsy needle Invivo 15100 is metallic. For the experiments, a needle with the same diameter was made of clear Acrylic plastic. Even though clinically the targeting errors will be more difficult to measure due to the metal artifacts from the needle, for this test the location can be more exactly measured without the artifacts. The plastic needle does not influence the accuracy results of the robot, because the artifact from the needle comes after aligning the guide on target, and inserting it. Then, no final adjustments were made. In the real case a marker may be placed through the guide for trajectory verification before inserting the actual needle, as shown in Fig. 12.

Nine target points were defined at the centers of the bottom grid of the mockup. These were successively selected as targets in the image navigation software. The robot adjusted the orientation of the needle-guide and preset the depth of insertion. The plastic needle was manually inserted through the guide up to the indicated depth. No manual adjustments were made. The entire experiment was then repeated, for a total of 18 targets. Images were acquired for confirmation after each insertion with the




Fig. 13. RCM mechanism and set point accuracy tests.

needle in place. The images show the void of signal in the region displaced by the needle. This region was segmented based on a Marching Cube algorithm [39]. The direction of the needle was determined with a principal components analysis. The depth of the needle was adjusted along the direction so that the model of the needle overlapped its image. The distance between the point of the needle model and target point was measured as the 3D targeting error. The distance from the target to the needle direction was also measured as the 2D error. The average of the target depths (from the RCM point) is calculated and named d_t . The accuracy ($A_{(d_t)}$) and precision are calculated as usual, as the average respectively the standard deviation of the errors.

Based on these and the bench test results (Section B2), the accuracy components that are due to factors other than the robot may be estimated as:

$$A_{(0)} \approx A_{(d_t)} - e_R d_t \quad (7)$$

Therefore, robot's best expected accuracy at a depth of d under MRI guidance can be estimated as:

$$A_{(d)} \approx A_{(0)} + e_R \cdot d \quad (8)$$

Here, the coefficient e_R depends on the robot itself, but $A_{(0)}$ depends on the imaging modality.

VI. RESULTS

A. Join Space Positioning Testing

The accuracy and (precision) for the R_1 , R_2 , and T axes were 0.11° (0.14°), 0.0° (0.0°), respectively 0.01 mm (0.02 mm).

B1. RCM Mechanism Check: The radius of the sphere fitted to the measured dataset was $L = 208.574$ [mm]. The center of the fitted sphere in robot coordinates was ${}^R O = (0.174, 0.061, 0.204)$ [mm] (see Fig. 13). The average distance from the measured dataset ${}^T a$ to the surface of the fitted sphere was 0.136 mm (SD 0.175 mm).

B2. Robot Space Set Point Testing: The average errors between the actual and set point datasets was 0.645 mm (SD 0.279 mm). The angular accuracy and (precision) of the




Fig. 14. Robot stiffness in axial and lateral directions to the needle-guide.




Fig. 15. Induced force (ASTM F2052) and torque (F2213) tests with robot suspended by (a) its hoses and (b) a string.

robot were 0.177° (0.077°). The estimated component of targeting errors due to the robot is $e_{R(10)} = 0.031$ mm (SD 0.013 mm) per cm of target depth.

C. Set Point Repeatability Testing

The repeatability measured 0.136° . The manipulator has no perceivable play or backlash in any direction and appears stiff.

D. Stiffness Testing

The force/deflection diagrams of the robot at the needle-guide in the axial and lateral directions are presented in Fig. 14. Based on the linear fit, the stiffness of the robot at the needle-guide is 33.38 N/mm axially and 25.53 N/mm laterally. Therefore, a displacement between 0.29 – 0.38 mm can be expected if the guide is inadvertently pushed with 1 kgf. During the forcing experiments the needle-guide did not detach from the robot.

E. MR Safe Testing

The manipulator is MR Safe according to ASTM F2503-13 because “An item composed entirely of electrically nonconductive, nonmetallic and nonmagnetic materials may be determined to be MR Safe by providing a scientifically based rationale rather than test data” [8]. Several tests were still performed for additional verification. Photos taken during the force and torque measurement experiments are shown in Fig. 15. The induced




Fig. 16. EMC test results. Spectrum of the noise in frequency domain (au = arbitrary units).




Fig. 17. SNR change due to the presence and motion of the robot: (a) Slice 1 images N, A, M from the T2 set and the ROI, (b) SNR change for T1, and (c) T2 sequences.

force and torques were below a perceivable level. There was no observable induced heat or vibrations. There was no observable induced force on the Interface controller outside the 0.5 mTesla (5 Gauss) line.

F. EMC Testing

Six each of the N, A, M test results are overlaid in the Fig. 16 graph. For all 18 tests the results were virtually identical, with somewhat different peaks around 13.5 MHz. On the Siemens MR used, the noise was reported in arbitrary units [au]. However, the results show that the robotic system and its motion did not observably change the EM emissions on a wide spectrum of frequencies. This is a very sensitive test. For example, additional noise is observable if the test is run with the MR door open.

G1. SNR Testing: The results of the SNR tests are presented in Fig. 17. A rectangular ROI, 140×140 [pixels], centered over the mockup, comprising coronal slices starting on bottom grid and ending on the top grid (11 slices, thickness

TABLE II
GLOBAL SNR AND ICF RESULTS (− IS LOSS, + IS GAIN)

Sequence type:	T1 [%]		T2 [%]	
	ΔSNR	ΔICF	ΔSNR	ΔICF
Normal: $\{N - N\}$	0.728	-0.045	-0.105	-0.128
Due to robot: $\{A - N\}$	0.144	-1.601	-0.352	-1.425
Due to motion: $\{M - A\}$	-0.855	-0.039	-0.638	-0.419




Fig. 18. ICF due to the presence and motion of the robot: (a) Slice 1 ROI images N, A, and A-N from the T2 set, (b) ICF for the T and (c) T2 sequences.

= 4 mm) was used. Corresponding images taken at the same location throughout the mockup from the N, A, and M sets are shown in Fig. 17(a), together with the ROI. The images show no observable relative differences.

The results of the SNR changes for the T1 and T2 sequences are shown in Fig. 17(b) and (c). The normal fluctuations in SNR of the scanner itself are approx. -1% to $+2\%$, for both sequences. With the robot the SNR change was similar (-1.6% to $+1.5\%$). Setting the robot in motion makes the range 0.5% wider (-2.5% to 1.35%). The changes were similar for the 2 sequence types. The global ΔSNR metrics are presented in Table II.

G2. ICF Testing: The results of the ICF tests are presented in Fig. 18(b) and (c). An $\{R - N\}$ change is observable in several slices. These correspond to the grid of the mockup. The difference image at one of these slices is shown in Fig. 18(a). Here, a very slight contour of the grid is visible, suggesting a relative shift of the N and A images. The cause is unknown. It could be due to slightly displacing the mockup at the time when the robot was removed before the N experiments (order: A, M, N). Comparatively, a simulated shift of the mockup in the T2 set with one pixel (pixel size 0.468 mm) makes $\Delta IDF^S \cong 8\%$. The global ΔIDF metrics are presented in Table II. Based on




Fig. 19. (a) Model grid superimposed over image of grid, and (b) Augmented reality displayed by the navigation software: Robot and mockup models registered to the image space. Nine targets defined at the centers of the bottom grid.




Fig. 20. Imaged needle at targets: (a) Segmented needle at one location together with 2 image slices showing the grid and a needle slice, and (b) Needle model registered to the needle segmentation. Actual needle point from the model and target.

the original IDF definitions of [5], the passive and active image deterioration factors over the entire mockup are $E_P = -1.04 \cdot 10^{-6}$ [%] and $E_A = -4.84 \cdot 10^{-8}$ [%], which are typically unobservable to radiologists.

G3. Geometric Artifacts: An image of the bottom grid with the model grid overlapped is shown in Fig. 19(a). The images show a slight misalignments of the grid towards the perimeter on the order of 0.4 mm.

H. Image-Guided Targeting

The augmented reality display of the navigation software is shown in Fig. 19(b). This displays the robot and mockup registered to the segmented markers, together with the 9 targets. One of the 9 targets was excluded at the time of the experiment because at that orientation the needle touched the top grid. Images of 8 of the inserted needles are shown in Fig. 20(b). Average target depth was $d_t = 37.91$ ($SD 1.12$). The accuracy and precision results of the 16 experiments are listed in Table III, together with the estimated accuracy at a target depth $A_{(d)}$, and an example at 7 cm, $A_{(70)}$.

VII. DISCUSSION

We reported the development and preclinical testing of a new robot for image-guided interventions. This uses an innovative parallelogram type RCM structure with non-collinear links. This design results in increased lateral clearance with

TABLE III
IMAGE-GUIDED TARGETING RESULTS

	2D [mm]	3D [mm]
Max Error	1.82	2.65
Accuracy $A_{(d_t)}$	1.19	1.71
Precision	0.45	0.51
Accuracy $A_{(d)}$	$1.07 + 0.0031 \cdot d$	$1.59 + 0.0031 \cdot d$
Accuracy $A_{(70)}$	1.29	1.81

respect to the patient and high structural stiffness. The stiffness at the needle guide measured 33 N/mm, the RCM mechanism has no perceivable backlash, and the RCM mechanism describes a sphere within a tolerance of 0.136 mm. These results were achieved with relatively high-power motors, a stiff structure that included numerous Garolite (fiberglass) components, and tightly fitted joints. A relative comparison with similar robots is difficult because stiffness is not normally reported, but subjectively it appears very stiff. Numerically, a 1 kgf of inadvertent force exerted in either direction on the guide at the RCM point yields to a deflection smaller than 0.38 mm.

The needle-guide is the only component that must be sterilized. Its latching mechanism and convex-concave mating profiles work through the sterile bag, mount precisely, and are extremely stiff (the force was exerted on the guide in the stiffness experiments, through the latch). The latch did not detach or break up to the 35 N level exerted axially or laterally. Many other applications could use a similar design.

Most IGI robots orient a needle guide, but the depth of insertion is manually set. Setting the depth manually works well if real-time images are readily available from the specific imager, for example under ultrasound guidance. Under MRI guidance, however, the access to the needle in the scanner combined with the special sequences required for near-real-time imaging make it helpful to have the depth of insertion preset by the robot. Here, we use a remote DoF to preset the depth. This feature was derived from our driver for the transrectal biopsy robot [16] and is updated with the O-Ring type spacer and a new mechanism. This feature fits into the anticipated clinical workflow, and did not appear to contribute to errors (our 3D minus 2D accuracy is 0.52 mm, which includes this as well as the registration errors of the needle model to its image). Of course the O-Ring mark can be moved if the operator pushes it excessively into guide, but we found that it was easy to operate properly during our mockup study. The O-Ring stopper could be used in many other applications.

The control system is similar to our previous design [16]. Piezoelectric valves are used and these are located in the MR room outside the 5 Gauss line. We found this to be a substantial advance over the voice-coil type valves that we used before [40]. Having the valves in the room makes the hoses shorter and eliminates the need for a connector on the hoses as well as the optical fibers. We feel that the system diagram in Fig. 7 is highly appropriate for MR Safe robots, and this configuration could be used for other applications.

A comprehensive set of tests were performed. These were synthesized based on international standards, our experience of nearly 2 decades of building MRI robots, and with the FDA approval process of another MR robot [18]. These start with kinematic tests of the robot itself, which are essential to set the accuracy base and quantify error components. ASTM tests for the magnetic resonance environment are included. If a device requires FDA approval (SR), the ASTM MR Safety and EMC tests are required by the FDA. For the EMC testing, we describe here how the MRI itself can be used as an EM emission meter. Low EM test results are also supportive for multi-imager compatibility, since the MR tests it on a wide frequency spectrum.

The markers that we use in the mockup are similar to those of the robot, and so were the registration algorithms used for the robot-to-image and mockup-to-image registrations. As such, we believe that if the robot is MR Safe (test E), if the robot is accurate (test B2), and if the grid properly overlaps the mockup (test G3), then the robot should be accurate in DIGI targeting. Since MRI testing is expensive, one could initially use only tests E, B2, and G3 and refine the robot and software, before proceeding to the other tests. In corollary, if tests B2 or G3 are not acceptable, IGI targeting (test H) should probably not be undertaken.

If used with other imaging modalities, the glass tubes of the robot registration markers could be filled with an appropriate contrast, or with mixtures of the contrast solutions if used in cross-modality IGI. In this case, test G3 should use the same type of contrast.

The change in SNR tests described in [16] were updated here with a base level measurement $\{N-N\}$. We feel this is very useful to set the base level of the tested device. Then, the image deterioration factors of [5], [36] were updated here to integrate with the SNR tests and to increase their sensitivity. The measures are similar, but the magnitude is higher (i.e. -0.352% vs. $-1.04 \cdot 10^{-6}$).

The tests were devised to isolate the errors components as much as possible. This enabled the estimation of targeting accuracy at a given depth, $A_{(d)}$. Of course, additional errors such as soft tissue and needle deflection will be present in-vivo. But the in-vitro estimate should help to estimate if the device is inappropriate to a specific clinical case. If the target is smaller than $A_{(d)}$, the robot will probably not work under the tested imaging modality (In (8), $A_{(0)}$ depends on the imaging modality). Additional tests with the specific imager are needed if another is used. This accuracy estimation is a new approach, which we believe could be useful for other researchers.

In summary, our research was built upon numerous previous technical achievements. The principal new developments include the RCM mechanism, stiff structure, and the detachable needle-guide. The paper also reports a very comprehensive set of preclinical tests. To the best of our knowledge this is the most comprehensive testing methodology reported to date for MR robots. It incorporates previous methods, regulatory tests, but also the new Image Change Factors and the new analysis of component errors in Image-Guided Targeting. This is also one of the few papers to report the stiffness of the manipulator, a critical element for IGI targeting. The paper also reports

several of our know-how lessons that we have learned, for example how to setup the tests in order to reduce the expense with the MRI.

VIII. CONCLUSION

Preclinical tests show that the robot is MR Safe, no EM emissions were found, has no observable mutual interference with the MRI, has a stiff structure, suggest that the robot is multi-imager compatible, and may be applicable to direct MR-guided interventions that require more than $1.07 + 0.0031 \cdot d$ [mm] accuracy in 2D or $1.59 + 0.0031 \cdot d$ [mm] in 3D. At a target depth of 7 cm for example, these are 1.29 mm respectively 1.81 mm. Since other imagers (i.e. the CT) are typically more accurate geometrically than the MR ($A_{(0)}^{CT} < A_{(0)}^{MR}$), the accuracy limits above appear safe to use for other imaging modalities. If more accuracy is needed, the error components related to the images ($A_{(0)}$) could be further improved for the MR or other modalities.

In summary, we report a new MR Safe robot, that is relatively simple and accurate.

REFERENCES

- [1] R. H. Taylor and D. Stoianovici, "Medical robotics in computer-integrated surgery," *IEEE Trans. Robot. Autom.*, vol. 19, no. 5, pp. 765–781, Oct. 2003. PMCID:ISI:000185906400002.
- [2] P. Mozer, A. W. Partin, and D. Stoianovici, "Robotic image-guided needle interventions of the prostate," *Rev. Urology*, vol. 11, no. 1, pp. 7–15, Winter 2009. PMCID:19390670.
- [3] K. Cleary *et al.*, "Interventional robotic systems: Applications and technology state-of-the-art," *Minimally Invasive Therapy Allied Technol.*, vol. 15, no. 2, pp. 101–113, 2006. PMCID:16754193.
- [4] D. Stoianovici, "Technology advances for prostate biopsy and needle therapies," *J. Urology*, vol. 188, no. 4, pp. 1074–1075, Oct. 2012. PMCID:22901579.
- [5] D. Stoianovici, "Multi-imager compatible actuation principles in surgical robotics," *Int. J. Med. Robot. Comput. Assist. Surg.*, vol. 1, no. 2, pp. 86–100, Jan. 2005. PMCID:17518382.
- [6] H. Elhawary *et al.*, "The case for MR-compatible robotics: A review of the state of the art," *Int. J. Med. Robot.*, vol. 4, no. 2, pp. 105–113, Jun. 2008. PMCID:18481822.
- [7] G. S. Fischer *et al.*, "MRI compatibility of robot actuation techniques—A comparative study," *Med. Image Comput. Comput. Assist. Intervention*, vol. 11, no. 2, pp. 509–517, 2008. PMCID:18982643.
- [8] *Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment*, ASTM F2503 – 13, 2013. [Online]. Available: http://compass.astm.org/EDIT/html_annot.cgi?F2503+13.
- [9] *CFR—Code of Federal Regulations Title 21, Part 812 Investigational Device Exemptions*. [Online]. Available: <http://www.accessdata.fda.gov/SCRIPTS/cdrh/cfdocs/cfCFR/CFRSearch.cfm?CFRPart=812>.
- [10] J. Tokuda *et al.*, "Preclinical evaluation of an MRI-compatible pneumatic robot for angulated needle placement in transperineal prostate interventions," *Int. J. Comput. Assist. Radiol. Surg.*, vol. 7, no. 6, pp. 949–957, Nov. 2012. PMCID:22678723.
- [11] Z. T. H. Tse *et al.*, "A 3-DOF MR-compatible device for magic angle related in vivo experiments," *IEEE-ASME Trans. Mechatron.*, vol. 13, no. 3, pp. 316–324, Jun. 2008. WOS:000256966700007.
- [12] S. Zangos *et al.*, "MR-compatible assistance system for biopsy in a high-field-strength system: Initial results in patients with suspicious prostate lesions," *Radiology*, vol. 259, no. 3, pp. 903–910, Jun. 2011. PMCID:21364080.
- [13] D. Yakan *et al.*, "Feasibility of a pneumatically actuated MR-compatible robot for transrectal prostate biopsy guidance," *Radiology*, vol. 260, no. 1, pp. 241–247, Jul. 2011. PMCID:21406625.
- [14] M. G. Schouten *et al.*, "The accuracy and safety aspects of a novel robotic needle guide manipulator to perform transrectal prostate biopsies," *Med. Phys.*, vol. 37, no. 9, pp. 4744–4750, Sep. 2010. PMCID:20964192.

- [15] M. Muntener *et al.*, "Transperineal prostate intervention: Robot for fully automated MR imaging-system description and proof of principle in a canine model," *Radiology*, vol. 247, no. 2, pp. 543–549, May 2008. PMCID:18430882.
- [16] D. Stoianovici *et al.*, "MRI-safe robot for endorectal prostate biopsy," *IEEE-ASME Trans. Mechatron.*, vol. 19, no. 4, pp. 1289–1299, Aug. 2014. PMCID:4219418 ISI:000335915800019.
- [17] G. Srinathveeravalli *et al.*, "MRI-safe robot for targeted transrectal prostate biopsy: Animal experiments," *Brit. J. Urology Int.*, vol. 113, no. 6, pp. 977–985, 2014. PMCID:24118992.
- [18] D. Stoianovici *et al.*, "MR safe robot, FDA clearance, safety, and feasibility of prostate biopsy clinical trial," *IEEE-ASME Trans. Mechatron.*, vol. 22, no. 1, pp. 115–126, Feb. 2017. WOS:000395750100014.
- [19] A. Squires *et al.*, "Robot for magnetic resonance imaging guided focal prostate laser ablation," *J. Med. Devices-Trans. ASME*, vol. 10, no. 3, pp. (030942) 1–3, Sep. 2016.
- [20] E. Kanal *et al.*, "ACR guidance document on MR safe practices," *J. Magn. Reson. Imag.*, vol. 37, no. 3, pp. 501–530, Mar. 2013. PMCID:23345200.
- [21] R. H. Taylor *et al.*, "Remote center-of-motion robot for surgery," US Patent 5 397 323, 1995.
- [22] C. H. Kuo, J. S. Dai, and P. Dasgupta, "Kinematic design considerations for minimally invasive surgical robots: an overview," *Int. J. Med. Robot.*, vol. 8, no. 2, pp. 127–145, Jun. 2012. PMCID:22228671.
- [23] S. E. Salcudean *et al.*, "A robot system for medical ultrasound," in *Proc. 9th Int. Symp. Robot. Res.*, 2000, pp. 195–202. WOS:000087440500024.
- [24] D. Stoianovici *et al.*, "A modular surgical robotic system for image guided percutaneous procedures," *Lecture Notes Comput. Sci.*, vol. 1496, pp. 404–410, 1998.
- [25] D. Stoianovici *et al.*, "AcuBot: A robot for radiological interventions," *IEEE Trans. Robot. Autom.*, vol. 19, no. 5, pp. 926–930, Oct. 2003.
- [26] G. Li *et al.*, "Robotic system for MRI-guided stereotactic neurosurgery," *IEEE Trans. Biomed. Eng.*, vol. 62, no. 4, pp. 1077–1088, Apr. 2015. WOS:000351769700008.
- [27] P. Shokrollahi, J. M. Drake, and A. A. Goldenberg, "Signal-to-noise ratio evaluation of magnetic resonance images in the presence of an ultrasonic motor," *Biomed. Eng. Online*, vol. 16, no. 1, p. 45, Apr. 2017. PMCID:28410615.
- [28] R. H. Taylor *et al.*, "A telerobotic assistant for laparoscopic surgery," *IEEE Eng. Med. Biol. Mag.*, vol. 14, no. 3, pp. 279–288, May/Jun. 1995. WOS:A1995QZ42400011.
- [29] A. Uneri *et al.*, "New steady-hand eye robot with micro-force sensing for vitreoretinal surgery," *Proc. IEEE RAS EMBS Int. Conf. Biomed. Robot Biomechatron.*, vol. 2010, nos. 26–29, pp. 814–819, Sep. 2010. PMCID:21461178.
- [30] D. Stoianovici *et al.*, "A new type of motor: Pneumatic step motor," *IEEE/ASME Trans. Mechatron.*, vol. 12, no. 1, pp. 98–106, Feb. 2007. PMCID:21528106.
- [31] E. Kanal *et al.*, "ACR guidance document for safe MR practices," *Amer. J. Roentgenol.*, vol. 188, no. 6, pp. 1447–1474, Jun. 2007. PMCID:17515363.
- [32] R. Decker *et al.*, "How to use an optical tracker for more accurate measurements," in *Proc. 27th Annu. Meeting Eng. Urology Soc.*, 2012, p. 35. [Online]. Available: http://engineering-urology.org/am/27EUS_2012.pdf.
- [33] B. K. P. Horn, "Closed-form solution of absolute orientation using unit quaternions," *J. Opt. Soc. Amer.-Opt. Image Sci. Vis.*, vol. 4, no. 4, pp. 629–642, Apr. 1987.
- [34] *Medical Electrical Equipment. Part 1-2: General Requirements for Basic Safety and Essential Performance - Collateral Standard: Electromagnetic Disturbances - Requirements and Tests (General II (ES/EMC))*, IEC 60601-1-2 Edition 4.0 2014-02, 2014. [Online]. Available: <http://webstore.ansi.org/RecordDetail.aspx?sku=IEC+60601-1-2+Ed.+4.0+b%3a2014>.
- [35] *Administration USFaD: Recognized Consensus Standards*. [Online]. Available: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/detail.cfm?standard_id=identification_no=32630.
- [36] N. Yu, R. Gassert, and R. Riener, "Mutual interferences and design principles for mechatronic devices in magnetic resonance imaging," *Int. J. Comput. Assist. Radiol. Surg.*, vol. 6, no. 4, pp. 473–488, Jul. 2011. PMCID:20811816.
- [37] *Determination of Signal-to-Noise Ratio (SNR) in Diagnostic Magnetic Resonance Imaging*, NEMA MS 1-2008, 2015. [Online]. Available: www.nema.org.
- [38] *ITK: Insight Segmentation and Registration Toolkit - Image Similarity Metrics*. [Online]. Available: <https://itk.org/Doxxygen/html/ImageSimilarityMetricsPage.html>.
- [39] W. E. Lorensen and H. E. Cline, "Marching cubes: A high resolution 3D surface construction algorithm," in *Proc. 14th Annu. Conf. Comput. Graph. Interact. Tech.*, vol. 21, 1987, pp. 163–169.
- [40] D. Stoianovici *et al.*, "MRI stealth," *Robot Prostate Interventions. Minimally Invasive Therapy Allied Technol.*, vol. 16, no. 4, pp. 241–248, 2007. PMCID:17763098.