

Quality Control Report (Both Cameras)

Timestamp: 2026-04-01T11:46:41

Beam used: 488 nm 

Intensity: 1 mW


Included cameras: Camera 1, Camera 2

Camera	Angle Tilt (deg)	Focus (col,row)	Beam Waist at focus
Camera 1	0.626	(958,1107.83)	4.006 μm
Camera 2	0.233	(1125,1100.09)	3.872 μm


Results - Camera 1

Angle Tilt (deg): 0.626
Focus (col,row): (958,1107.83)
Beam Waist (FWHM at focus): 4.006 μm

FWHM graph:




Camera snapshot + overlays (Camera 1):



Results - Camera 2

Angle Tilt (deg): 0.233
Focus (col,row): (1125,1100.09)
Beam Waist (FWHM at focus): 3.872 μm

FWHM graph:



Camera snapshot + overlays (Camera 2):

