LIFEGENE
*Supporting Science*

# Pipette Calibration Certificate

| DOCUMENT No. | CALIBRATED BY | CALIBRATION DATE | CALIBRATION RESULT |
|---|---|---|---|
| **LG-CAL-2026-0004** | **-** | **12.11.2026** | **PASS** |

## Pipette data

Manufacturer / model
**Gilson P200**

Type
**Single channel**

Volume range
**20 - 200 uL**

Serial number
**GS-2026-0042**

Tip
**D200**

## Environmental conditions

Water temperature
**22.0 °C**

Relative humidity
**50 %**

Air pressure
**1013.2 hPa**

Z factor (uL/mg)
**1.00328**

Balance
**RADWAG analytical balance**

## Calibration

Method
**Gravimetric - for academic / internal use only**

Replicates
**4 per test volume**

Place of calibration
**LifeGene Lab, Mevo Horon, Israel**

Next calibration
**12.11.2027**

Comments
**-**

## AVERAGE RESULTS

| LOW - nominal 20 uL | | | | | |
|---|---|---|---|---|---|
| **Average volume (uL)** | **Systematic error (uL)** | **Systematic error (%)** | **Random error SD (uL)** | **CV (%)** | **Uncertainty U (uL)** |
| **19.989** | -0.011 | -0.06 | 0.033 | 0.17 | ± 0.033 |
| Limit | ± 1.6 | PASS | ≤ 0.6 | PASS | k = 2 |
| Weights (g) | 0.019970 | 0.019922 | 0.019897 | 0.019904 | |
| Volumes (uL) | 20.035 | 19.987 | 19.962 | 19.969 | |

| MID - nominal 100 uL | | | | | |
|---|---|---|---|---|---|
| **Average volume (uL)** | **Systematic error (uL)** | **Systematic error (%)** | **Random error SD (uL)** | **CV (%)** | **Uncertainty U (uL)** |
| **100.033** | +0.033 | +0.03 | 0.193 | 0.19 | ± 0.193 |
| Limit | ± 1.6 | PASS | ≤ 0.6 | PASS | k = 2 |
| Weights (g) | 0.099636 | 0.099748 | 0.099948 | 0.099491 | |
| Volumes (uL) | 99.963 | 100.075 | 100.276 | 99.817 | |

| HIGH - nominal 200 uL | | | | | |
|---|---|---|---|---|---|
| **Average volume (uL)** | **Systematic error (uL)** | **Systematic error (%)** | **Random error SD (uL)** | **CV (%)** | **Uncertainty U (uL)** |
| **200.013** | +0.013 | +0.01 | 0.342 | 0.17 | ± 0.342 |
| Limit | ± 1.6 | PASS | ≤ 0.6 | PASS | k = 2 |
| Weights (g) | 0.198988 | 0.199494 | 0.199764 | 0.199191 | |
| Volumes (uL) | 199.641 | 200.148 | 200.419 | 199.844 | |

## Signatures

| Calibrated by | Approved by | Contact |
|---|---|---|
| _____ | _____ | **LifeGene Ltd.**<br>LifeGene, Mevo Horon, Israel<br>Call us now: +972-(0)8-6444314<br>Email: info@lifegene.co.il |
| Date: 12.11.2026 | Date: _____ | |

## Notes

The results reported in this certificate relate only to the instrument identified above, as received and tested on the calibration date. Measurements were taken with distilled water at the recorded ambient conditions and converted to volume using the density of water and an air-buoyancy correction (Z factor). The systematic error is the difference between the mean delivered volume and the nominal test volume; the random error is the standard deviation of the replicate deliveries. The reported expanded uncertainty is based on a coverage factor k = 2 (approximately 95 % confidence). The next calibration date is a recommendation; the user remains responsible for the suitability of the instrument for its intended use.

**IMPORTANT NOTICE - SCOPE OF CALIBRATION: This calibration certificate is issued for internal quality-control and academic reference purposes only. The calibration was performed using a gravimetric method and reflects the performance of the instrument at the time of testing under the environmental conditions recorded herein. This certificate does not represent accredited calibration and has not been issued under any national or international accreditation standard (including ISO/IEC 17025 or equivalent). The results shall not be used as a substitute for accredited calibration where regulatory or legal compliance requires one. LifeGene Ltd. assumes no liability for decisions made solely on the basis of this document beyond its stated internal use.**

- End of report -