LIFEGENE

Supporting Science
Mevo Horon, Israel
+972-(0)8-6444314 | info@lifegene.co.il
www.LifeGene.co.il

Pipette
Calibration Certificate

Document No.	Operator	Calibration date	CALIBRATION RESULT
LG-CAL-2026-0001	Dmitri Trakhtman	12/11/2026	**PASS**

Pipettes data
Manufacturer: **GILSON**
Model: **Classic P200**
Range: **20 - 200 µL**
Channels: **1**
Tip: **DF200**
Serial No: **GS-2026-0042**

Environmental conditions
Air temperature [°C] **22.00**
Pressure [hPa] **1013.2**
Humidity [%] **50.0**
Temp. of water [°C] **22.00**
Z factor [mL/g] **1.003280**
Balance: **RADWAG analytical balance**

Calibration
Place of calibration: **LifeGene Lab, Mevo Horon, Israel**
Calibration method: **Gravimetric - For academic / internal use only**
Measurements: **4 per test volume**
Mass unit: **g**
Standard: **W-004-04 Calibration of Pipette**
Recommended next calibration: **12/11/2027**

AVERAGE RESULTS

Volume [µL]	20	Weight [g]	0.019921	Norm. systematic error [µL]	± 1.6	Norm. random error [µL]	≤ 0.6

Canal No / Weighings	Average volume [µL]	Syst. err. [%]	Syst. err. [µL]	Rand. err. [%]	Rand. err. [µL]	Measurement uncertainty [µL]	Result A	Result CV
Canal 1 n = 4	**19.9858**	-0.0708	-0.0142	0.0773	0.0155	± 0.0155	✓	✓

Individual weights (g): 0.019921 0.019904 0.019916 0.019941 *volumes (µL):* 19.986 19.969 19.981 20.006

Volume [µL]	100	Weight [g]	0.099782	Norm. systematic error [µL]	± 1.6	Norm. random error [µL]	≤ 0.6

Canal No / Weighings	Average volume [µL]	Syst. err. [%]	Syst. err. [µL]	Rand. err. [%]	Rand. err. [µL]	Measurement uncertainty [µL]	Result A	Result CV
Canal 1 n = 4	**100.0037**	+0.0037	+0.0037	0.1405	0.1405	± 0.1405	✓	✓

Individual weights (g): 0.099782 0.099807 0.099520 0.099598 *volumes (µL):* 100.109 100.134 99.846 99.925

Volume [µL]	200	Weight [g]	0.199316	Norm. systematic error [µL]	± 1.6	Norm. random error [µL]	≤ 0.6

Canal No / Weighings	Average volume [µL]	Syst. err. [%]	Syst. err. [µL]	Rand. err. [%]	Rand. err. [µL]	Measurement uncertainty [µL]	Result A	Result CV
Canal 1 n = 4	**199.9800**	-0.0100	-0.0200	0.0971	0.1943	± 0.1943	✓	✓

Individual weights (g): 0.199316 0.199056 0.199467 0.199466 *volumes (µL):* 199.970 199.709 200.121 200.120

Copyright © LifeGene Ltd. Calibration certificate shall not be reproduced except in full.

Calibrated by: **Dmitri Trakhtman**



Dmitri Trakhtman

Authorized Signature - signed electronically 12/11/2026

LifeGene Ltd.
Mevo Horon, Israel
+972-(0)8-6444314
info@lifegene.co.il
www.LifeGene.co.il

Accounting for the uncertainty will be taken to mean that at a 95 % confidence level the measurement results plus and minus the expanded uncertainty (k = 2) shall be totally within the specification limits and the risk of false acceptance / rejection will not be greater than 2.5 %.

A binary decision rule is used for the statement of conformity according to the ILAC Guidelines on Decision Rules and Statements of Conformity (ILAC-G8:09/2019).

The systematic error (A) is the difference between the mean delivered volume and the nominal test volume; the random error (CV) is the standard deviation of the replicate deliveries. Weights were converted to volume with the Z factor (density of water at the test temperature, corrected for air buoyancy).

We certify that the instrument stated above meets the acceptance limits published in this certificate at the time of testing.

IMPORTANT NOTICE - SCOPE OF CALIBRATION: This calibration certificate is issued for internal quality-control and academic reference purposes only. The calibration was performed using a gravimetric method and reflects the performance of the instrument at the time of testing under the environmental conditions recorded herein. This certificate does not represent accredited calibration and has not been issued under any national or international accreditation standard (including ISO/IEC 17025 or equivalent). The results shall not be used as a substitute for accredited calibration where regulatory or legal compliance requires one. LifeGene Ltd. assumes no liability for decisions made solely on the basis of this document beyond its stated internal use.

End of report