
1 Loading height


Adjusting and memorising the stretcher loading height.

1 Pull the release lever (A) of the extractable tray.

Note: this way, the device is positioned at the bottom to allow the extractable tray to come out.




2 Pull the lever (A) again and accompany the exit of the tray until it engages in the "fully extended" position.



3 Adjust the loading height (H) using the foot side buttons (B).

Note: move the stretcher close and optimise its tilting angle to facilitate loading.





4 Memorise the desired position by simultaneously pressing the two arrows on the control panel.

PRESS **Up** + **Down** **simultaneously**

