

## **Varian OSMS Clinical Report**

**Patient Name: TestJHV ZZUWQA (Female)**

**Patient Number: 20170307 JHV**


**Treatment Date: 07, March, 2017**

**Treatment Site: Intracranial SRS**

### **OSMS Positioning Report prior to the Delivery of Treatment**

The OSMS system was used to position the patient immediately prior to treatment and to track the patient during treatment in 3D mode. The patient was monitored for 00:06:02.

The following parameters are the couch shifts which were computed by OSMS in the treatment position, in order to assess the accuracy of patient set-up and to prevent geographical miss with the radiation beam, in standard couch coordinates.



## Couch Angle: 0 Shifts

D\_VRT (mm): -28.60

D\_LNG (mm): 8.10

D\_LAT (mm): -18.52

D\_Rtn (degrees): 0.12

D\_Roll (degrees): 4.66

D\_Pitch (degrees): -0.55

The patient was aligned properly prior to treatment initiation and the patient position was tracked during treatment delivery. Imaging using the OSMS system was performed by the radiation therapist.

**RTT   Name: Physicist (Physicist)**

**Signature**

**MD   Name:**

**Signature**