

CARDIOTRACK CT-200 HOME BLOOD PRESSURE MONITOR USER MANUAL

1.0 INTRODUCTION

The CardioTrack CT-200 is an easy-to-use digital monitor designed for home use. It provides fast and reliable measurements of systolic and diastolic blood pressure, as well as heart rate. Please keep this manual for future reference.

2.0 SAFETY INSTRUCTIONS

Please read this section carefully before operating the device. Keep it in a safe place.

2.1 Pressure Safety Limits

If the cuff pressure exceeds the safe limit of 300 mmHg, the device should automatically deflate the cuff and display error code E3 within 1 second to prevent patient injury.

2.2 Battery Safety

Use only authorized AA batteries. Do not mix old and new batteries. If the device is not used for a long time, remove the batteries.

3.0 OPERATING INSTRUCTIONS

Follow these steps to ensure accurate blood pressure measurements.

3.1 Applying the Cuff

Wrap the cuff around your upper left arm, about 2-3 cm above the elbow. The tube should run down the center of the arm.

3.2 Taking a Measurement

Press the START button. Remain still and quiet during the measurement. The cuff will inflate automatically.

3.3 Battery Safety

Open the battery cover on the back of the device. Install four AA batteries matching the polarity markings (+ and -). Close the cover securely.

4.0 ERROR CODES

Below are the system error codes, their causes, and solutions.

Error Code	Description & Solution
E1	Pulse sensor failed. Adjust cuff and try again.
E2	Movement detected. Stay still and measure again.
E3	Cuff pressure exceeded 300 mmHg. Deflated automatically.