



Analytical Laboratory Report

Report ID: S34111.01(01)
Generated on 04/12/2022

Report to

Attention: David Cook
Global Environmental Consulting
223 W. Michigan Ave.
Clinton, MI 49236

Phone: 517-456-6881 FAX:
Email: gec223@earthlink.net

Additional Contacts: Doug Ervin

Report produced by

Merit Laboratories, Inc.
2680 East Lansing Drive
East Lansing, MI 48823

Phone: (517) 332-0167 FAX: (517) 332-6333

Contacts for report questions:
John Lavery (johnlavery@meritlabs.com)
Barbara Ball (bball@meritlabs.com)

Report Summary

Lab Sample ID(s): S34111.01
Project: Brighton Township WWTP
Collected Date(s): 03/18/2022
Submitted Date/Time: 03/22/2022 13:30
Sampled by: David Cook
P.O. #: D1873

Table of Contents

Cover Page (Page 1)
General Report Notes (Page 2)
Report Narrative (Page 2)
Laboratory Certifications (Page 3)
Qualifier Descriptions (Page 3)
Glossary of Abbreviations (Page 3)
Method Summary (Page 4)
Sample Summary (Page 5)

Maya Murshak
Technical Director



Analytical Laboratory Report

General Report Notes

Analytical results relate only to the samples tested, in the condition received by the laboratory.

Methods may be modified for improved performance.

Results reported on a dry weight basis where applicable.

'Not detected' indicates that parameter was not found at a level equal to or greater than the reporting limit (RL).

When MDL results are provided, then 'Not detected' indicates that parameter was not found at a level equal to or greater than the MDL.

40 CFR Part 136 Table II Required Containers, Preservation Techniques and Holding Times for the Clean Water Act specify that samples for acrolein, acrylonitrile, and 2-chlorovinylethyl ether need to be preserved at a pH in the range of 4 to 5 or if not preserved, analyzed within 3 days of sampling.

QA/QC corresponding to this analytical report is a separate document with the same Merit ID reference and is available upon request.

Full accreditation certificates are available upon request. Starred (*) analytes are not NELAP accredited.

Samples are held by the lab for 30 days from the final report date unless a written request to hold longer is provided by the client.

Report shall not be reproduced except in full, without the written approval of Merit Laboratories, Inc.

Limits for drinking water samples, are listed as the MCL Limits (Maximum Contaminant Level Concentrations)

PFAS requirement: Section 9.3.8 of U.S. EPA Method 537.1 states "If the method analyte(s) found in the Field Sample is present in the

FRB at a concentration greater than 1/3 the MRL, then all samples collected with that FRB are invalid and must be recollected and reanalyzed."

Samples submitted without an accompanying FRB may not be acceptable for compliance purposes.

Wisconsin PFAs analysis: MDL = LOD; RL = LOQ. LOD and LOQ are adjusted for dilution.

Report Narrative

There is no additional narrative for this analytical report

Laboratory Certifications

Authority	Certification ID
Michigan DEQ	#9956
DOD ELAP/ISO 17025	#69699
WBENC	#2005110032
Ohio VAP	#CL0002
Indiana DOH	#C-MI-07
New York NELAC	#11814
North Carolina DENR	#680
North Carolina DOH	#26702
Alaska CSLAP	#17-001
Pennsylvania DEP	#68-05884
Wisconsin DNR	FID# 399147320

Qualifier Descriptions

Qualifier	Description
!	Result is outside of stated limit criteria
B	Compound also found in associated method blank
E	Concentration exceeds calibration range
F	Analysis run outside of holding time
G	Estimated result due to extraction run outside of holding time
H	Sample submitted and run outside of holding time
I	Matrix interference with internal standard
J	Estimated value less than reporting limit, but greater than MDL
L	Elevated reporting limit due to low sample amount
M	Result reported to MDL not RDL
O	Analysis performed by outside laboratory. See attached report.
R	Preliminary result
S	Surrogate recovery outside of control limits
T	No correction for total solids
X	Elevated reporting limit due to matrix interference
Y	Elevated reporting limit due to high target concentration
b	Value detected less than reporting limit, but greater than MDL
e	Reported value estimated due to interference
j	Analyte also found in associated method blank
p	Benzo(b)Fluoranthene and Benzo(k)Fluoranthene integrated as one peak.
x	Preserved from bulk sample

Glossary of Abbreviations

Abbreviation	Description
RL/RDL	Reporting Limit
MDL	Method Detection Limit
MS	Matrix Spike
MSD	Matrix Spike Duplicate
SW	EPA SW 846 (Soil and Wastewater) Methods
E	EPA Methods
SM	Standard Methods
LN	Linear
BR	Branched

Method Summary

Method	Version
ASTM D7968-17M	ASTM Method D7968 - 17 Modified (Isotopic Dilution)
SM2540B	Standard Method 2540 B 2015

Parameter Summary

Parameter	Synonym	Cas #
PFBA	Perfluorobutanoic Acid	375-22-4
PFPeA	Perfluoropentanoic Acid	2706-90-3
4:2 FTSA	4:2 Fluorotelomer Sulfonic Acid	757124-72-4
PFHxA	Perfluorohexanoic Acid	307-24-4
PFBS	Perfluorobutane sulfonic Acid	375-73-5
PFHpA	Perfluoroheptanoic Acid	375-85-9
PFPeS	Perfluoropentane Sulfonic Acid	2706-91-4
6:2 FTSA	6:2 Fluorotelomer Sulfonic Acid	27619-97-2
PFOA	Perfluorooctanoic Acid	335-67-1
PFHxS	Perfluorohexane Sulfonic Acid	355-46-4
PFHxS-LN	Perfluorohexane Sulfonic Acid - LN	355-46-4-LN
PFHxS-BR	Perfluorohexane Sulfonic Acid - BR	355-46-4-BR
PFNA	Perfluorononanoic Acid	375-95-1
8:2 FTSA	8:2 Fluorotelomer Sulfonic Acid	39108-34-4
PFHpS	Perfluoroheptane Sulfonic Acid	375-92-8
PFDA	Perfluorodecanoic Acid	335-76-2
N-MeFOSAA	N-methyl perfluorooctanesulfonamidoacetic acid	2355-31-9
EtFOSAA	N-Ethyl Perfluorooctane Sulfonamidoacetic Acid	2991-50-6
PFOS	Perfluorooctane Sulfonic Acid	1763-23-1
PFOS-LN	Perfluorooctane Sulfonic Acid - LN	1763-23-1-LN
PFOS-BR	Perfluorooctane Sulfonic Acid - BR	1763-23-1-BR
PFUnDA	Perfluoroundecanoic Acid	2058-94-8
PFNS	Perfluorononane Sulfonic Acid	68259-12-1
PFDoDA	Perfluorododecanoic Acid	307-55-1
PFDS	Perfluorodecane Sulfonic Acid	335-77-3
PFTTrDA	Perfluorotridecanoic Acid	72629-94-8
FOSA	Perfluorooctane Sulfonamide	754-91-6
PFTeDA	Perfluorotetradecanoic Acid	376-06-7
11Cl-PF3OUdS	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	763051-92-9
9Cl-PF3ONS	9-chlorohexadecafluoro-3-oxanone1-sulfonic acid	756426-58-1
ADONA	4,8-dioxa-3H-perfluorononanoic acid	919005-14-4
HFPO-DA	Hexafluoropropylene oxide dimer	13252-13-6



Analytical Laboratory Report

Sample Summary (1 samples)

Sample ID	Sample Tag	Matrix	Collected Date/Time
S34111.01	Sludge Waste	Sludge	03/18/22 10:30

Lab Sample ID: S34111.01

Sample Tag: Sludge Waste

Collected Date/Time: 03/18/2022 10:30

Matrix: Sludge

COC Reference: 126637

Sample Containers

#	Type	Preservative(s)	Refrigerated?	Arrival Temp. (C)	Thermometer #
1	15ml Centrifuge Tube	None	Yes	5.0	IR
1	250ml Plastic	None	Yes	5.0	IR

Extraction / Prep.

Parameter	Result	Method	Run Date	Analyst	Flags
Initial wt. (g) / Final wt. (g) / Volume (ml)*	12.92/7.05/10	ASTM D7968-17M	03/29/22 13:30	KCV	

Inorganics

Method: SM2540B, Run Date: 03/22/22 14:13, Analyst: MAM

Parameter	Result	RL	MDL	Units	Dilution	CAS#	Flags
Total Solids*	3.6	1		%	1		

Organics

28 PFAs, Method: ASTM D7968-17M, Run Date: 03/30/22 22:12, Analyst: KCV

Parameter	Result	RL	MDL	Units	Dilution	CAS#	Flags
PFBA*	Not detected	2.8		ug/kg	47.3	375-22-4	IX
PFPeA*	2.2	0.47		ug/kg	47.3	2706-90-3	I
4:2 FTSA*	Not detected	0.47		ug/kg	47.3	757124-72-4	I
PFHxA*	3	0.47		ug/kg	47.3	307-24-4	I
PFBS*	0.55	0.47		ug/kg	47.3	375-73-5	
PFHpA*	0.64	0.47		ug/kg	47.3	375-85-9	I
PFPeS*	Not detected	0.47		ug/kg	47.3	2706-91-4	
6:2 FTSA*	Not detected	0.47		ug/kg	47.3	27619-97-2	I
PFOA*	11	0.47		ug/kg	47.3	335-67-1	
PFHxS*	Not detected	0.47		ug/kg	47.3	355-46-4	
PFHxS-LN*	Not detected	0.47		ug/kg	47.3	355-46-4-LN	
PFHxS-BR*	Not detected	0.47		ug/kg	47.3	355-46-4-BR	
PFNA*	2.3	0.47		ug/kg	47.3	375-95-1	
8:2 FTSA*	0.53	0.47		ug/kg	47.3	39108-34-4	I
PFHpS*	Not detected	0.47		ug/kg	47.3	375-92-8	
PFDA*	47	0.47		ug/kg	47.3	335-76-2	I
N-MeFOSAA*	8.3	0.47		ug/kg	47.3	2355-31-9	
EtFOSAA*	15	0.47		ug/kg	47.3	2991-50-6	
PFOS*	15	0.47		ug/kg	47.3	1763-23-1	
PFOS-LN*	14	0.47		ug/kg	47.3	1763-23-1-LN	
PFOS-BR*	1.6	0.47		ug/kg	47.3	1763-23-1-BR	
PFUnDA*	8.8	0.47		ug/kg	47.3	2058-94-8	I
PFNS*	Not detected	0.47		ug/kg	47.3	68259-12-1	
PFDoDA*	14	0.47		ug/kg	47.3	307-55-1	I
PFDS*	Not detected	0.47		ug/kg	47.3	335-77-3	
PFTTrDA*	1.7	0.47		ug/kg	47.3	72629-94-8	I
FOSA*	6.2	0.47		ug/kg	47.3	754-91-6	
PFTeDA*	2.9	0.47		ug/kg	47.3	376-06-7	I1
11CI-PF3OUdS*	Not detected	0.47		ug/kg	47.3	763051-92-9	

I-Matrix interference with internal standard X-Elevated reporting limit due to matrix interference

1-IS recovery <10%



Analytical Laboratory Report

Lab Sample ID: S34111.01 (continued)

Sample Tag: Sludge Waste

28 PFAs, Method: ASTM D7968-17M, Run Date: 03/30/22 22:12, Analyst: KCV (continued)

Parameter	Result	RL	MDL	Units	Dilution	CAS#	Flags
9CI-PF3ONS*	Not detected	0.47		ug/kg	47.3	756426-58-1	
ADONA*	Not detected	0.47		ug/kg	47.3	919005-14-4	
HFPO-DA*	Not detected	0.47		ug/kg	47.3	13252-13-6	I

I-Matrix interference with internal standard

Merit Laboratories Login Checklist

Lab Set ID:S34111

Client:GLOBAL02 (Global Environmental Consulting)

Project: Brighton Township WWTP

Submitted:03/22/2022 13:30 Login User: PFD

Attention: David Cook

Address: Global Environmental Consulting
223 W. Michigan Ave.
Clinton, MI 49236

Phone: 517-456-6881

FAX:

Email: gec223@earthlink.net

Selection	Description	Note
Sample Receiving		
01.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Samples are received at 4C +/- 2C Thermometer # IR 5.0
02.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Received on ice/ cooling process begun
03.	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	Samples shipped
04.	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	Samples left in 24 hr. drop box
05.	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	Are there custody seals/tape or is the drop box locked
Chain of Custody		
06.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	COC adequately filled out
07.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	COC signed and relinquished to the lab
08.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Sample tag on bottles match COC
09.	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	Subcontracting needed? Subcontracted to:
Preservation		
10.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Do sample have correct chemical preservation
11.	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	Completed pH checks on preserved samples? (no VOAs)
12.	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	Did any samples need to be preserved in the lab?
Bottle Conditions		
13.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	All bottles intact
14.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Appropriate analytical bottles are used
15.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Merit bottles used
16.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Sufficient sample volume received
17.	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	Samples require laboratory filtration
18.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Samples submitted within holding time
19.	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	Do water VOC or TOX bottles contain headspace

Corrective action for all exceptions is to call the client and to notify the project manager.

Client Review By: _____ Date: _____

REPORT TO

CONTACT NAME David Cook			
COMPANY GEC			
ADDRESS 223 W Michigan Ave			
CITY Clinton			STATE MI
			ZIP CODE 49236
PHONE NO. 734649 2998		FAX NO.	P.O. NO.
E-MAIL ADDRESS GEC223@earthlink.net			QUOTE NO. D1873

CHAIN OF CUSTODY RECORD

CONTACT NAME		<input type="checkbox"/> SAME	
COMPANY			
ADDRESS			
CITY		STATE	ZIP CODE
PHONE NO.		E-MAIL ADDRESS	

INVOICE TO

PROJECT NO./NAME <u>Brighton Township WWT</u>	SAMPLER(S) - PLEASE PRINT/SIGN NAME <u>David Cook</u>
TURNAROUND TIME REQUIRED <input type="checkbox"/> 1 DAY <input type="checkbox"/> 2 DAYS <input type="checkbox"/> 3 DAYS <input checked="" type="checkbox"/> STANDARD <input type="checkbox"/> OTHER _____	
DELIVERABLES REQUIRED <input type="checkbox"/> STD <input type="checkbox"/> LEVEL II <input type="checkbox"/> LEVEL III <input type="checkbox"/> LEVEL IV <input type="checkbox"/> EDD <input type="checkbox"/> OTHER _____	

ANALYSIS (ATTACH LIST IF MORE SPACE IS REQUIRED)

Certifications

☐ OHIO VAP ☐ Drinking Water
☐ DoD ☐ NPDES

Project Locations

☐ Detroit ☐ New York☐ Other _____









Special Instructions

MATRIX CODE:	GW=GROUNDWATER SL=SLUDGE	WW=WASTEWATER DW=DRINKING WATER	S=SOIL O=OIL	L=LIQUID WP=WIPE	SD=SOLID A=AIR W=WASTE	# Containers & Preservatives
--------------	-----------------------------	------------------------------------	-----------------	---------------------	------------------------------	------------------------------

[illegible]

PFAS (28)

ILC

RELINQUISHED BY: SIGNATURE/ORGANIZATION		<input checked="" type="checkbox"/> Sampler	DATE	TIME
RECEIVED BY: SIGNATURE/ORGANIZATION			3/21/22	9:00
RELINQUISHED BY: SIGNATURE/ORGANIZATION			DATE	TIME
RECEIVED BY: SIGNATURE/ORGANIZATION			3/22/22	10:00
RELINQUISHED BY: SIGNATURE/ORGANIZATION			DATE	TIME
RECEIVED BY: SIGNATURE/ORGANIZATION			3/22/22	1:30
RELINQUISHED BY: SIGNATURE/ORGANIZATION			DATE	TIME
RECEIVED BY: SIGNATURE/ORGANIZATION			3/22/22	13:30

RELINQUISHED BY: SIGNATURE/ORGANIZATION			DATE	TIME
RECEIVED BY: SIGNATURE/ORGANIZATION			DATE	TIME
SEAL NO.	SEAL INTACT YES <input type="checkbox"/> NO <input type="checkbox"/>	INITIALS	NOTES:	TEMP. ON ARRIVAL _____ 5.0
SEAL NO.	SEAL INTACT YES <input type="checkbox"/> NO <input type="checkbox"/>	INITIALS		

PLEASE NOTE: SIGNING ACKNOWLEDGES ADHERENCE TO MERIT'S SAMPLE ACCEPTANCE POLICY ON REVERSE SIDE