

## **After a Power Failure**

R.C. Stewart

19/05/2025

If the generator at MVO fails, and the seismic data-acquisition computers go off, the following problems have been observed when power is restored.

### **Linux Computers**

Disks are not mounted.

### **Windows Computers**

Software might not start until after the user has responded to the prompt asking about the downtime.

### **earthworm03**

The helicorder and spectrogram plots are not updating. This is because the file has been corrupted in such a way that you can view it, but *earthworm* can't write to it.

In this case, do the following:

- Stop earthworm
- Rename all of today's files with an "a" before the ".gif".
- Uncomment the *#BuildOnRestart* line in
  - *heli\_ewII.d*
  - *heli\_ewII\_wide.d*
  - *sgram.d*
- Start earthworm
- Recoment the *#BuildOnRestart* line.

### **opsrm04 (display computer above centre windows in Ops Room)**

If the three monitors are in the wrong order, reboot the computer.