1 Example module (synthetic)

Lt Module Output · Source: c1-r101 · captured 2000-01-01 · exported 2000-01-01

⚠ This page is your own module output, written by lt-module-export. lt-module-answer skips every page stamped with lt-module::, so it is never read back as a source.

Page 3 — Calibration slide images

Upload your own labelled sketch of the light path and label each part.

`i-upann-1 · medium confidence`

- Light passes from the lamp through the condenser, the specimen, the objective lens and the eyepiece to the eye.
  EX-PRAC, slide-9
- A published survey of teaching microscopes reports a typical working distance of a few millimetres at low power.
  10.5555/synthetic.0001, table 2

  ⚠ **Uncertain.** Is a working distance wanted, or the path alone? — An untaught figure can look like filler to a marker.
  - Tier 1 → 2: insufficient-depth. The notes trace the light path but give no distance.
  - Tier 2 → 3: absent. The Tier 2 source has nothing on optics.
  - Tier 3 → 4: absent. Nothing at Tier 3 either.

Explain why immersion oil improves resolution at high magnification.

`i-text-1 · high confidence`

- Immersion oil has a refractive index close to that of glass, so light leaving the slide is not bent away from the objective lens.
  EX-PRAC, slide-14
- This matters for two reasons.
- More of the light cone enters the lens, which raises the numerical aperture and so the finest detail the lens can resolve.
  EX-PRAC, slide-14 · **EX-LECT**, 00:23:40

Name the control that adjusts fine focus.

`i-text-2 · low confidence`

- The fine focus knob adjusts fine focus.
  EX-PRAC, slide-22
- A maintenance guide names the same control the micrometer drive.
  EX-GUIDE, section-4

  ⚠ **Tier conflict.** (**EX-PRAC**, slide-22) says “fine focus knob”; (**EX-GUIDE**, section-4) says “micrometer drive”. Shown to the reviewer, not merged. The course name wins in a course answer.

  ⚠ **Uncertain.** Which name does the practical expect? — Two sources, two names, and no rule here for picking one.
  - Tier 1 → 2: contradicted. Slide 22 and the guide name the knob differently; both stay on record.
  - Tier 2 → 3: absent. The Tier 2 source has nothing on microscopes.

Not drafted

- i-table-1 — already-answered

TODO Record the measured field diameter.

`i-unmapped-1`

- pluginKey 'kcDataValue' has no tested write path.
- Placement: Page 3, panel pk-12, the measured field diameter field. Type the value by hand.
- i-cat-1 — single-attempt-drag