

Regulatory

Intraosseous radiofrequency ablation of the basivertebral nerve is a surgical procedure and thus not regulated by the FDA.

K190504 Trade/Device Name: Intracept Intraosseous Nerve Ablation System (RF Probe),
Intracept Intraosseous Nerve Ablation System (Access Instruments),
Relievent RF Generator Regulation Number: 21 CFR 882.4725
Regulation Name: Radiofrequency Lesion Probe Regulatory Class: Class II Product Code: GXI Dated:
February 28, 2019 Received: March 1, 2019
501 K premarket approval: May 2019

Intracept Intraosseous Nerve Ablation System Relievent Medsystems RF Generator (RFG) indications:

The Intracept Intraosseous Nerve Ablation System is intended to be used in conjunction with radiofrequency (RF) generators for the ablation of basivertebral nerves of the L3 through S1 vertebrae for the relief of chronic low back pain of at least six months duration that has not responded to at least six months of conservative care, and is also accompanied by features consistent with Type 1 or Type 2 Modic changes on an MRI such as inflammation, edema, vertebral endplate changes, disruption and fissuring of the endplate, vascularized fibrous tissues within the adjacent marrow, hypointensive signals (Type 1 Modic change), and changes to the vertebral body marrow including replacement of normal bone marrow by fat, and hyperintensive signals (Type 2 Modic change).

The Relievent RFG is intended to be used with RF probes FDA cleared as part of the Relievent Intracept Intraosseous Nerve Ablation System for the ablation of basivertebral nerves of the L3 through S1 vertebrae for the relief of chronic low back pain of at least six months duration that has not responded to at least six months of conservative care, and is also accompanied by features consistent with Type 1 or Type 2 Modic changes on an MRI such as inflammation, edema, vertebral endplate changes, disruption and fissuring of the endplate, vascularized fibrous tissues within the adjacent marrow, hypointensive signals (Type 1 Modic change), and changes to the vertebral body marrow including replacement of normal bone marrow by fat, and hyperintensive signals (Type 2 Modic change).

K250213 Trade/Device Name: OptaBlate BVN Intraosseous Nerve Ablation System (10mm Probe Single Kit);

OptaBlate BVN Intraosseous Nerve Ablation System (10mm Probe Dual Kit)

Date Received 01/24/2025

Decision Date 05/15/2025

501 K premarket approval: May 2025

The Stryker OptaBlate BVN Intraosseous Nerve Ablation System indications:

The Stryker OptaBlate BVN Intraosseous Nerve Ablation System is intended for the ablation of basivertebral nerves of the L3 through S1 vertebrae for the relief of chronic low back pain of at least six months duration that has not responded to at least six months of conservative care, and is also accompanied by features consistent with Type 1 or Type 2 Modic changes on an MRI such as inflammation, edema, vertebral endplate changes, disruption and fissuring of the endplate, vascularized fibrous tissues within the adjacent marrow, hypointensive signals (Type 1 Modic change), and changes