Pre-adoption checklist for fetal-ultrasound plane classifiers

#	Item	Question to ask before adopting a plane classifier
1	Per-device results	Are accuracy and rejection rates reported separately for each ultrasound machine?
2	External sites	Was the model tested at centres and on devices not seen in training?
3	Rejection rate	How often are true standard planes rejected as 'Other' on new devices?
4	Overlay robustness	Does performance hold when calipers, text and UI differ from training?
5	Confound audit	Can metadata or the frame alone (no anatomy) predict the label in the training data?
6	Saliency sanity check	Do explanation maps pass a model-randomisation test before being shown?
7	Calibration	Are confidence scores calibrated on the target device, with a stated abstention rule?
8	Local pilot	Is there a silent local pilot on the site's own machines before clinical use?
9	Data governance	Are data handled under PCPNDT and institutional rules (no sex determination, consent)?