




Instrument Name: ZenoTOF™ 7600 System
Instrument Model: ZenoTOF 7600 System
Ion Source: OptiFlow 50-200µL Micro/MicroCal
Serial Number: FB24962403

Manufacturer: AB Sciex
Target Instrument: ZenoTOF 7600 System
Optimization Set: Positive Quick Status Check



1. Positive MS Check

Step	Result	Details																														
Achieve Stable Spray / Modify	Pass	Slope: 0.000490128761020068, previous was 0.000490128761020068 Delay: -13.4385170172153, previous was -13.4385170172153																														
Average of 30 spectra - Confirmation Mode Complete																																
																																
<table><tr><th>Item</th><th>Min</th><th>Max</th><th>Actual</th><th>Status</th></tr><tr><td>Mass (Da)</td><td>132.8049</td><td>133.0049</td><td>132.9032</td><td>Pass</td></tr><tr><td>Sum Intensity (cps)</td><td>1000</td><td>1e+11</td><td>10678</td><td>Pass</td></tr></table>	Item	Min	Max	Actual	Status	Mass (Da)	132.8049	133.0049	132.9032	Pass	Sum Intensity (cps)	1000	1e+11	10678	Pass	<table><tr><th>Item</th><th>Min</th><th>Max</th><th>Actual</th><th>Status</th></tr><tr><td>Mass (Da)</td><td>829.4393</td><td>829.6393</td><td>829.5310</td><td>Pass</td></tr><tr><td>Sum Intensity (cps)</td><td>1000</td><td>1e+11</td><td>18558</td><td>Pass</td></tr></table>		Item	Min	Max	Actual	Status	Mass (Da)	829.4393	829.6393	829.5310	Pass	Sum Intensity (cps)	1000	1e+11	18558	Pass
Item	Min	Max	Actual	Status																												
Mass (Da)	132.8049	133.0049	132.9032	Pass																												
Sum Intensity (cps)	1000	1e+11	10678	Pass																												
Item	Min	Max	Actual	Status																												
Mass (Da)	829.4393	829.6393	829.5310	Pass																												
Sum Intensity (cps)	1000	1e+11	18558	Pass																												
																																

Channel Alignment

Pass

Sum of 10 spectra

									
Item	Min	Max	Actual	Status	Item	Min	Max	Actual	Status
Mass (Da)	132.6549	133.1549	132.9028	Pass	Mass (Da)	829.2893	829.7893	829.530	Pass
Resolution	0	100000	34233	Pass	Resolution	0	100000	37917	Pass
Sum Intensity (cps)	0	1e+11	6.6813e+05	Pass	Sum Intensity (cps)	0	1e+11	1.9885e+06	Pass
Align to channel: 2 Result: Pass									
Item	Min	Max	Initial	Actual	Status				
Channel 1 vs. 2 Slope	-50	50	-0.122697399	-0.119468368	Pass				
Channel 1 vs. 2 Intercept	-50	50	-11.630417737	-10.987371767	Pass				
Channel 3 vs. 2 Slope	-50	50	0.074599955	0.075420505	Pass				
Channel 3 vs. 2 Intercept	-50	50	-3.730581109	-3.943745481	Pass				
Channel 4 vs. 2 Slope	-50	50	0.128690242	0.123578614	Pass				
Channel 4 vs. 2 Intercept	-50	50	8.172561522	8.874604557	Pass				

Instrument Name: ZenoTOF™ 7600 System
Instrument Model: ZenoTOF 7600 System
Ion Source: OptiFlow 50-200μL Micro/MicroCal
Serial Number: FB24962403


Manufacturer: AB Sciex
Target Instrument: ZenoTOF 7600 System
Optimization Set: Positive Quick Status Check

TOF MS Mass Check

Pass

Slope: 0.000490130851918769, previous was 0.000490128761020068
Delay: -13.4822019365703, previous was -13.4385170172153

Average of 10 spectra - Confirmation Mode Complete




TOF MS/MS Mass Check

Pass

Slope: 0.000490130009192977, previous was 0.000490127867977034
Delay: -13.5006217186637, previous was -13.5185464050735

Average of 10 spectra - Confirmation Mode Complete



TOF MS/MS Mass Check Zeno On

Pass

Slope: 0.000490130996394524, previous was 0.000490129121738077
Delay: -13.4857096858094, previous was -13.50770592866

Average of 10 spectra - Confirmation Mode Complete

