


**Instrument Name:** ZenoTOF™ 7600 System  
**Instrument Model:** ZenoTOF 7600 System  
**Ion Source:** OptiFlow 50-200µL Micro/MicroCal  
**Serial Number:** FB24962403

**Manufacturer:** AB Sciex  
**Target Instrument:** ZenoTOF 7600 System  
**Optimization Set:** Positive Quick Status Check

## 1. Positive MS Check

| Step                          | Result | Details  |
|-------------------------------|--------|--|
| Achieve Stable Spray / Modify | Pass   | Slope: 0.0004901089968472, previous was 0.0004901089968472<br>Delay: -13.4117255093855, previous was -13.4117255093855 |


Average of 30 spectra - Confirmation Mode Complete



## Channel Alignment

Pass

Sum of 10 spectra



**Instrument Name:** ZenoTOF™ 7600 System  
**Instrument Model:** ZenoTOF 7600 System  
**Ion Source:** OptiFlow 50-200µL Micro/MicroCal  
**Serial Number:** FB24962403


**Manufacturer:** AB Sciex  
**Target Instrument:** ZenoTOF 7600 System  
**Optimization Set:** Positive Quick Status Check

## TOF MS Mass Check

**Pass**

Slope: 0.00049011030874611, previous was 0.0004901089968472  
Delay: -13.4671347450144, previous was -13.4117255093855

### Average of 10 spectra - Confirmation Mode Complete




## TOF MS/MS Mass Check

**Pass**

Slope: 0.000490109247627954, previous was 0.000490108117016973  
Delay: -13.4950844003287, previous was -13.4846053051093

### Average of 10 spectra - Confirmation Mode Complete



## TOF MS/MS Mass Check Zeno On

**Pass**

Slope: 0.000490110468318482, previous was 0.000490109423469443  
Delay: -13.4615987827408, previous was -13.4700449290315

### Average of 10 spectra - Confirmation Mode Complete

