




Instrument Name: ZenoTOF™ 7600 System
Instrument Model: ZenoTOF 7600 System
Ion Source: OptiFlow 50-200µL Micro/MicroCal
Serial Number: FB24962403

Manufacturer: AB Sciex
Target Instrument: ZenoTOF 7600 System
Optimization Set: Positive Quick Status Check



1. Positive MS Check

Step	Result	Details																														
Achieve Stable Spray / Modify	Pass	Slope: 0.000490061723338294, previous was 0.000490061723338294 Delay: -13.3773477147312, previous was -13.3773477147312																														
Average of 30 spectra - Confirmation Mode Complete																																
																																
<table><tr><th>Item</th><th>Min</th><th>Max</th><th>Actual</th><th>Status</th></tr><tr><td>Mass (Da)</td><td>132.8049</td><td>133.0049</td><td>132.9024</td><td>Pass</td></tr><tr><td>Sum Intensity (cps)</td><td>1000</td><td>1e+11</td><td>13793</td><td>Pass</td></tr></table>	Item	Min	Max	Actual	Status	Mass (Da)	132.8049	133.0049	132.9024	Pass	Sum Intensity (cps)	1000	1e+11	13793	Pass	<table><tr><th>Item</th><th>Min</th><th>Max</th><th>Actual</th><th>Status</th></tr><tr><td>Mass (Da)</td><td>829.4393</td><td>829.6393</td><td>829.5263</td><td>Pass</td></tr><tr><td>Sum Intensity (cps)</td><td>1000</td><td>1e+11</td><td>15764</td><td>Pass</td></tr></table>	Item	Min	Max	Actual	Status	Mass (Da)	829.4393	829.6393	829.5263	Pass	Sum Intensity (cps)	1000	1e+11	15764	Pass	
Item	Min	Max	Actual	Status																												
Mass (Da)	132.8049	133.0049	132.9024	Pass																												
Sum Intensity (cps)	1000	1e+11	13793	Pass																												
Item	Min	Max	Actual	Status																												
Mass (Da)	829.4393	829.6393	829.5263	Pass																												
Sum Intensity (cps)	1000	1e+11	15764	Pass																												
																																

Channel Alignment

Pass

Sum of 10 spectra

									
Item	Min	Max	Actual	Status	Item	Min	Max	Actual	Status
Mass (Da)	132.6549	133.1549	132.9018	Pass	Mass (Da)	829.2893	829.7893	829.5252	Pass
Resolution	0	100000	38356	Pass	Resolution	0	100000	38938	Pass
Sum Intensity (cps)	0	1e+11	7.2565e+05	Pass	Sum Intensity (cps)	0	1e+11	1.8046e+06	Pass
Align to channel: 2 Result: Pass									
Item	Min	Max	Initial	Actual	Status				
Channel 1 vs. 2 Slope	-50	50	-0.123370267	-0.11904876	Pass				
Channel 1 vs. 2 Intercept	-50	50	-10.870397251	-11.481762602	Pass				
Channel 3 vs. 2 Slope	-50	50	0.069486241	0.074529109	Pass				
Channel 3 vs. 2 Intercept	-50	50	-3.722826498	-3.866621726	Pass				
Channel 4 vs. 2 Slope	-50	50	0.109851305	0.120056138	Pass				
Channel 4 vs. 2 Intercept	-50	50	9.060881839	8.656859793	Pass				

Instrument Name: ZenoTOF™ 7600 System
Instrument Model: ZenoTOF 7600 System
Ion Source: OptiFlow 50-200µL Micro/MicroCal
Serial Number: FB24962403


Manufacturer: AB Sciex
Target Instrument: ZenoTOF 7600 System
Optimization Set: Positive Quick Status Check

TOF MS Mass Check

Pass

Slope: 0.00049006483391588, previous was 0.000490061723338294
Delay: -13.4494703241482, previous was -13.3773477147312

Average of 10 spectra - Confirmation Mode Complete




TOF MS/MS Mass Check

Pass

Slope: 0.000490063898310161, previous was 0.000490061595921675
Delay: -13.4742421648903, previous was -13.3973249013344

Average of 10 spectra - Confirmation Mode Complete



TOF MS/MS Mass Check Zeno On

Pass

Slope: 0.000490065074272785, previous was 0.000490062456910012
Delay: -13.442075498609, previous was -13.3947222678976

Average of 10 spectra - Confirmation Mode Complete

