


# PARTICLE COUNTER

Date: January 18, 2025 | Time: 9:20 | Booth ID: SB-02 | Model: TH-DL2000

## Particle Count Monitoring



## DAILY SUMMARY STATISTICS

Parameter	Value	Range	Status
Particle Count	2800	NMT 3520 ≥0.5µm/m³	✓ PASS

## INSTRUMENT INFORMATION

HEPA Integrity	Pass
Comments	Weekly check. All parameters OK.



System generated | Serial: TH-DL2000-2558 | Cal Due: 2025-12-31