

DIGITAL THERMO-HYGROMETER / DATA LOGGER

Date: January 04, 2025 | Time: 9:10 | Booth ID: SB-02 | Model: TH-DL2000

Environmental Data



DAILY SUMMARY STATISTICS

Parameter	Value	Range	Status
Temperature	22.3°C	20.0-25.0°C	✓ PASS
Relative Humidity	48.0%	NMT 55.0%	✓ PASS
Overall Status			✓ PASS

INSTRUMENT INFORMATION

HEPA Integrity	Pass
Comments	Weekly check. All parameters OK.