

# DIGITAL THERMO-HYGROMETER / DATA LOGGER

Date: February 08, 2025 | Time: 9:10 | Booth ID: SB-01

## Environmental Data



## MEASUREMENT DETAILS

Parameter	Value	Unit	Acceptance Criteria	Status
Temperature	21.8	°C	20.0 - 25.0 °C	✓ PASS
Relative Humidity	49	%	NMT 55.0%	✓ PASS
Overall Status				✓ PASS

## INSTRUMENT INFORMATION

HEPA Integrity	Pass
Type of Monitoring	Weekly
Instrument Calibration Status	Y
CFU Count	N/A
Remarks	Weekly check

## SYSTEM GENERATED DOCUMENT

Calibration Due: August 29, 2026

This document is automatically generated and does not require manual signature.