


PARTICLE COUNTER

Date: January 25, 2025 | Time: 9:00 | Booth ID: SB-02 | Model: TH-DL2000

Particle Count Monitoring



DAILY SUMMARY STATISTICS

Parameter	Value	Range	Status
Particle Count	3050	NMT 3520 ≥0.5µm/m³	✓ PASS

INSTRUMENT INFORMATION

HEPA Integrity	Pass
Comments	Pre-sampling check for Packaging Materials.



System generated | Serial: TH-DL2000-3745 | Cal Due: 2025-12-31