

In-body Antigen Detection via Antibodies on a Transistor

Duran Kaan Altin

Abstract—This project develops a simulation workflow for HER2 antigen detection using an antibody-functionalized graphene field-effect transistor (GFET). The model links HER2 transport, antibody binding, bound molecule count, GFET drain-source current modulation, sensitivity, and limit of detection (LOD). The current repository includes a meshed COMSOL Transport of Diluted Species model for the transport stage, a separate partially anatomical prescribed-flow extension, analytical surface-binding post-processing, and analytical GFET response calculations. Results are reported with documented limitations rather than experimental validation.

I. INTRODUCTION

HER2 is a clinically important breast cancer biomarker used in prognosis and therapy selection [1], [2]. Conventional HER2 testing workflows such as immunohistochemistry, fluorescence in situ hybridization, and immunoassays are powerful but are not designed for continuous in-body monitoring. This project studies whether HER2 arrival, binding, and electrical transduction can be represented in a reproducible simulation pipeline for a GFET biosensor.

The central modeling question is whether the detection response is limited only by transistor sensitivity or also by antigen transport to the sensing region. The repository therefore separates the workflow into binding kinetics, COMSOL transport, surface binding, and GFET current response.


II. BACKGROUND

A. HER2 as a Biomarker

HER2 amplification or overexpression is associated with aggressive breast cancer subtypes and is routinely evaluated in clinical decision-making [1], [2]. This study treats HER2 as the target antigen and does not claim clinical validation.

B. Antigen-Antibody Binding

The binding stage uses reversible antigen-antibody kinetics:



$$r = k_f[Ag][Ab] - k_r[AgAb], \quad (2)$$

with dissociation constant:

$$K_d = \frac{k_r}{k_f}. \quad (3)$$

For surface binding, the Langmuir occupancy relation is used:

$$\theta = \frac{c_{surface}}{K_d + c_{surface}}. \quad (4)$$

C. GFET Biosensing Principle

Field-effect biosensors convert local charge perturbations into channel-current changes, following the foundation of ion-sensitive and field-effect sensing [3], [4]. Graphene is attractive because its surface-exposed channel is sensitive to local electrostatic perturbations [5], [6]. This project uses an analytical GFET transduction model rather than a full graphene semiconductor model.

III. METHODOLOGY

A. 0D Binding Kinetics

The M1 stage evaluates equilibrium occupancy across HER2 concentrations of 0.5, 1, 10, 100, and 1000 pM. The baseline assumptions are $K_d = 10$ pM, $k_f = 10^7$ 1/(M s), and $k_r = 10^{-4}$ 1/s.

M1 equilibrium occupancy vs HER2 concentration




Fig. 1. M1 equilibrium occupancy versus HER2 concentration.

B. 3D COMSOL Transport Model

The M2 stage uses COMSOL Transport of Diluted Species in a simplified cortex/medulla-inspired 3D geometry. The medulla diffusion coefficient is parameterized by:

$$D_{medulla} = rD_{cortex}, \quad (5)$$

where r is swept over 0.1, 0.25, 0.5, 0.75, and 1.0. The current COMSOL model includes a mesh and time-dependent study with exported times of 0, 100, 500, 1000, 2000, 4000, and 6000 s.




Fig. 2. M2 COMSOL-stage concentration profile at 1000 s.




Fig. 3. M2 COMSOL-stage flux visualization at 1000 s.

C. Partially Anatomical Lymph-Node Flow Extension

The M2B extension keeps the M2 diffusion-only model unchanged and adds a separate lymph-node-inspired flow scaffold. The geometry uses an outer node radius of approximately $500 \mu\text{m}$, a medulla radius of approximately $250 \mu\text{m}$, a subcapsular/cortex/medulla region structure, four afferent inlet markers, one efferent outlet marker, and local/full sensor markers. The model is intentionally described as partially anatomical because it captures transport-relevant regions and inlet/outlet structure without claiming full lymph-node anatomy.

The flow implementation uses prescribed inlet velocities rather than a fully validated Darcy-flow solve. The velocity sweep is $v_{in} = 10^{-7}, 5 \times 10^{-7}, 10^{-6}$ m/s. The M2B post-processing compares sensor exposure and delay against the preserved M2 diffusion baseline using the same transport time points and diffusivity-ratio sweep.

D. Surface Binding Model

The M3 stage reads the M2 sensor-surface concentration output and computes surface occupancy, surface-bound density, and bound molecule count:

$$\Gamma = \Gamma_{max}\theta, \quad (6)$$




Fig. 4. M2B prescribed-velocity streamline proxy for the partially anatomical flow extension.




Fig. 5. M2B concentration profile proxy at 1000 s compared with the preserved diffusion baseline.

$$N_{bound} = \Gamma A_{sensor} N_A. \quad (7)$$

Two sensing areas are evaluated: full-boundary exposure and local GFET-scale exposure.

E. GFET Current-Response Model

The M4 stage maps bound molecule count to current response:

$$\Delta I_{DS} = \frac{W}{L} e \mu V_{DS} \frac{\alpha N}{A_{eff}}. \quad (8)$$

The minimum detectable bound molecule count is:

$$N_{min} = \frac{I_{min} A_{eff}}{(W/L) e \mu V_{DS} \alpha}. \quad (9)$$

F. Sensitivity and LOD Estimation

Sensitivity is estimated from the low-concentration slope of ΔI_{DS} versus HER2 concentration. LOD is calculated as:

$$LOD = \frac{3\sigma}{S}, \quad (10)$$

where σ is the assumed current noise floor and S is sensitivity [7].

M2B flow vs M2 diffusion sensor exposure




Fig. 6. Reference M2B prescribed-flow sensor exposure compared with M2 diffusion-only exposure.

M2 COMSOL-stage delay vs diffusivity ratio




Fig. 8. Medulla uptake delay versus diffusivity ratio.

IV. RESULTS

A. Binding Kinetics

The M1 output shows bounded saturation behavior: occupancy increases with HER2 concentration and remains within $[0, 1]$.

B. COMSOL Transport Profiles

The M2 COMSOL model builds and solves as a meshed Transport of Diluted Species study. The solved baseline model reports 3103 degrees of freedom plus internal DOFs. Direct COMSOL field-table extraction is not automated in this repository version, so the tabular M2_comsol_* outputs are reported as COMSOL-stage post-processed transport metrics.

C. Cortex-Medulla Uptake Delay

The diffusivity-ratio sweep shows that medulla uptake delay decreases as r approaches 1.

M2 COMSOL-stage cortex vs medulla uptake at $r=0.5$




Fig. 7. Cortex, medulla, and sensor concentration trends at $r = 0.5$.

D. M2B Flow Comparison

For the reference case $r = 0.5$ and $v_{in} = 5 \times 10^{-7}$ m/s, the M2B prescribed-flow extension gives a sensor time-to-50% exposure of approximately 1733 s, compared with approximately 5337 s for the M2 diffusion-only baseline. At 6000 s, the reported sensor concentration proxy is 8.73 pM for M2B and 4.98 pM for M2. These values support only the limited conclusion that the selected prescribed-flow extension increases sensor exposure relative to the diffusion-only baseline.

E. Surface Binding Response

Surface occupancy remains bounded and full-boundary exposure produces a larger total bound molecule count than the local GFET-scale sensing region.

M3 bound HER2 count from M2 sensor concentration




Fig. 9. Bound HER2 molecule count over time for local sensing at 10 pM.

F. DeltaIds and LOD

The current response increases with HER2 concentration and coupling efficiency. Under the current assumptions, the $\alpha = 0.03$, 10 pA case reaches a simulated LOD of approximately 0.83 pM. Higher noise or lower coupling gives larger LOD and should be reported separately.




Fig. 10. GFET current response versus HER2 concentration.




Fig. 11. Minimum detectable bound molecule count for coupling and noise-floor cases.

V. DISCUSSION

The current results support the modeling claim that transport and surface exposure matter before transistor sensitivity can be interpreted. The response improves with higher coupling efficiency and lower noise floor, but sub-pM LOD is not universal across all assumptions.

VI. LIMITATIONS

This project is simulation-only. No wet-lab validation, fabrication, clinical testing, biocompatibility analysis, or direct physiological Debye screening model is included. The M2 model is meshed and solved in COMSOL, but direct field-table export from solved .mph files remains a repository limitation. The M2B extension is a prescribed-velocity, partially anatomical transport scaffold rather than a validated Darcy-flow or full anatomical lymph-node model. Surface binding and GFET response are analytically post-processed rather than fully coupled inside COMSOL.

VII. CONCLUSION

The repository now implements a staged HER2-GFET workflow: 0D binding, meshed COMSOL transport, a separate

partially anatomical prescribed-flow extension, surface binding from sensor concentration, and GFET current-response/LOD estimation. The model is suitable for final report drafting with the documented limitations stated explicitly.

REFERENCES

- [1] D. J. Slamon, G. M. Clark, S. G. Wong, W. J. Levin, A. Ullrich, and W. L. McGuire, "Human breast cancer: Correlation of relapse and survival with amplification of the her-2/neu oncogene," *Science*, vol. 235, no. 4785, pp. 177–182, 1987.
- [2] A. C. Wolff, M. E. H. Hammond, K. H. Allison, B. E. Harvey, P. B. Mangu, J. M. S. Bartlett, M. Bilous, I. O. Ellis, P. Fitzgibbons, W. Hanna, R. B. Jenkins, M. F. Press, P. A. Spears, G. H. Vance, G. Viale, L. M. McShane, and M. Dowsett, "Human epidermal growth factor receptor 2 testing in breast cancer: American society of clinical oncology/college of american pathologists clinical practice guideline focused update," *Journal of Clinical Oncology*, vol. 36, no. 20, pp. 2105–2122, 2018.
- [3] P. Bergveld, "Development of an ion-sensitive solid-state device for neurophysiological measurements," *IEEE Transactions on Biomedical Engineering*, vol. BME-17, no. 1, pp. 70–71, 1970.
- [4] E. Stern, A. Vacic, and M. A. Reed, "Semiconducting nanowire field-effect transistor biomolecular sensors," *IEEE Transactions on Electron Devices*, vol. 55, no. 11, pp. 3119–3130, 2008.
- [5] Y. Ohno, K. Maehashi, Y. Yamashiro, and K. Matsumoto, "Electrolyte-gated graphene field-effect transistors for detecting ph and protein adsorption," *Nano Letters*, vol. 9, no. 9, pp. 3318–3322, 2009.
- [6] T. Kuila, S. Bose, P. Khanra, A. K. Mishra, N. H. Kim, and J. H. Lee, "Recent advances in graphene-based biosensors," *Biosensors and Bioelectronics*, vol. 26, no. 12, pp. 4637–4648, 2011.
- [7] A. Shrivastava and V. B. Gupta, "Methods for the determination of limit of detection and limit of quantitation of the analytical methods," *Chronicles of Young Scientists*, vol. 2, no. 1, pp. 21–25, 2011.