

检 验 报 告
Certificate of Analysis

No. COA20250301017

产品名称 Product Name	缓冲液 System Detection Solution	REF	N/A
批号 Lot No.	2025010111	型号 Model of the Reagent	N/A
物料编码 Material Code	105-009141-00	包装规格 Package Size	Every kit includes System Detection Solution A (5×3 mL) and System Detection Solution B (1×5 mL)
生产日期 Manufacture Date	2025-02-28	有效期 Expiry date	2026-02-27
标准依据 Standard basis	产品检验规范 Specification for Product Inspection		

检测项目 Subjects	接受标准 Accept criteria	检测结果 Results	结果判定 Conclusion
1、外观性状 Appearance	为无色透明液体，无沉淀、无悬浮物、无絮状物。 Clear liquid free of particulate matters and floccules.	符合标准/Qualified	符合/Pass
2、装量 Volume	缓冲液A≥3mL System Detection Solution A volume is ≥3mL	3 mL	符合/Pass
装量 Volume	缓冲液B≥5mL System Detection Solution B volume is ≥5mL	5 mL	符合/Pass
3、性能 Performance	试剂针CV<2% CV of reagent needle<2%	reagent needle:1%	符合/Pass
系统重复性 System repeatability	样本针CV<2% CV of sample needle<2%	sample needle:1%	符合/Pass

深圳迈瑞生物医疗电子股份有限公司
SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO.,LTD.
地址：深圳市南山区高新技术产业园科技南十二路迈瑞大厦 邮编：518057 电话：0755-8188 8998 传真：0755-26582680
Address: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P.R. China
Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com



结论
Summary

合格/PASS

备注
Remarks

检验：
Analysis

陈素兰Chen Sulan

陈素兰 ChenSulan

日期：
Date

2025-03-01

批准（检验工程师）：
Approved by(Quality Engineer)

林朋Lin Peng

林朋 LinPeng

日期：
Date

2025-03-01

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO.,LTD.

地址：深圳市南山区高新技术产业园科技南十二路迈瑞大厦 邮编：518057 电话：0755-8188 8998 传真：0755-26582680

Address: Mindray Building, Keji 12th Road South, High-Tech Industrial Park,Nanshan, Shenzhen, 518057, P.R. China

Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com

