

检验报告 Certificate of Analysis

No. COA20250115042

产品名称

铁蛋白校准品

Product Name

Ferritin Calibrators

N/A N/A

批号 Lot No.

2024120111

2025-01-14

105-004289-00

Model of the Reagent

 $C0:1\times 2.0$ mL, $C1:1\times 2.0$ mL, $C2:1\times 2.$

物料编码 **Material Code**

Package Size

REF

型号

生产日期

Manufacture

有效期

Date

标准依据

Expiry date

2026-07-13

Standard basis

产品检验规范 Specification for Product Inspection

检测项目 检测结果 结果判定 接受标准 Subjects Accept criteria Results Conclusion

1、外观性状 Appearance

外观和性状

Appearance

校准品为清澈透明液体, 无沉淀、无悬浮物、无 絮状物; 分装瓶为白色塑料瓶, 盖有塑料外盖; 盒贴、瓶贴、校准卡完整、清晰牢固; 盒中不含 说明书, 盒贴上印有 only for local production字句; 塑料外盖与塑料瓶身配合后 无泄漏; 塑料外盖无明显划痕、崩缺。

The calibrators are clear liquid without precipitate, suspension and floccules. The containers are white plastic bottles with tightly closed

caps; The whole package, including labels for boxes and bottles, and calibration cards, is complete and clear; No instructions in the box and there is "only for local production" on the box label. The content of labels is correct and complete. The containers are without leakage and obvious

defect or breakdown.

符合标准/Qualified

符合/Pass

2、装量 Volume

C0≥2.0mL

2.0 mL

装量 Volume

符合/Pass

C1≥2.0mL

2.0 mL

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





C2≥2.0mL 2.0 mL

3、性能 Performance

> CO的SD≤2ng/mL $C0:SD \leq 2ng/mL$

CO:O ng/mL

瓶内均一性 Uniformity in bottle

CV≤8% CV≤8%

C1:1%

符合/Pass

CV≤8%

CV≤8%

C2:1%

校准品的乙型肝炎病毒表面抗原、人类免疫缺陷 病毒抗体(HIV-I型和HIV-II型)、丙型肝炎病 毒抗体、梅毒螺旋体抗体检测结果均为阴性。

生物安全性 Biosafety The calibrators are free from Hepatitis B Surface Antigen (HBsAg), antibody to HIV-1/HIV-2, antibody to

Hepatitis C (HCV), and Treponema Pallidum (TP) by using the CE-marked

or equivalent method.

阴性/Negative

符合/Pass

偏倚应在±10.0%范围内。

The bias of trueness control sample is H:-2.4%

less than $\pm 10.0\%$.

赋值准确性 Accuracy

符合/Pass

偏倚应在±10.0%范围内。

The bias of trueness control sample is L:-1.1%

less than $\pm 10.0\%$.

结论

Summary

合格/PASS

备注

Remarks

检验: 陈素兰Chen Sulan **Analysis**

陈素兰 Charlian

批准(检验工程师): Approved by(Quality Engineer) 胡凤Hu Feng

- His Ry Huteng

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





日期: 日期: 2025-01-15 2025-01-15 Date Date

深圳迈瑞生物医疗电子股份有限公司

深州記論土物医プモナ版の日報公司 SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大夏 邮編: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com

