

检验报告 Certificate of Analysis

No. COA20250416029

产品名称

样本稀释液

Product Name Sample Diluent

型号

REF

N/A

Lot No.

2025030111

2025-04-15

Model of the Reagent

N/A

物料编码

105-004276-00 **Material Code**

Package Size

 2×30 mL

生产日期

Manufacture

有效期

Date

标准依据

Expiry date

2026-04-14

Standard basis

产品检验规范 Specification for Product Inspection

检测项目 Subjects 1、外观性状 Appearance	接受标准 Accept criteria	检测结果 Results	结果判定 Conclusion
外观和性状 Appearance	清澈透明、无沉淀、无悬浮物、无絮状物 Clear liquid without precipitate, suspension and floccules.	符合标准/Qualified	符合/Pass
2、装量 Volume			
装量 Volume	试剂瓶基座1腔≥14mL chamber 1 ≥ 14 mL	14 mL	
	试剂瓶基座2腔≥6.2mL chamber 2 ≥ 6.2 mL	6.4 mL	符合/Pass
	试剂瓶基座3腔≥6.2mL chamber 3 ≥6.2 mL	6.4 mL	
	磁珠试剂瓶≥6.4mL Cavum round ≥ 6.4 mL	6.5 mL	

3、性能 Performance



SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





平均稀释回收率75%~125%

Average Rcovery is between 75%-125%.

E2:113%

最大稀释度

平均稀释回收率75%~125%

Maximum

HCG:111%

符合/Pass

平均稀释回收率75%~125%

Average Rcovery is between 75%-125%.

T3:116%

结论

Summary

合格/PASS

备注

Remarks

检验:

邓月Deng Yue Analysis

AH Yeng Yue

批准(检验工程师):

Approved by(Quality Engineer) 李冠林Li GuanLin 考起族 Li Guanlin

日期:

2025-04-16 Date

Date

2025-04-16

深圳迈瑞生物医疗电子股份有限公司

