

检验报告 Certificate of Analysis

No. COA20250322039

产品名称 Product Name

甲状旁腺素校准品 PTH Calibrators

N/A

N/A

批号

2025030111

--,

Lot No.

型号 Model of the Reagent

REF

,

物料编码

初行編制 Material Code 105-008553-00 **观格** CO:1:

生产日期

T 20

Package Size 0 m

C0:1 \times 2.0 mL, C1:1 \times 2.0 mL, C2:1 \times 2.0 mL

Manufacture

有效期 2025-03-21 **Expiry date**

2026-03-20

Date

标准依据 Standard basis

产品检验规范 Specification for Product Inspection

检测项目 接受标准 检测结果 结果判定 Subjects Accept criteria Results Conclusion

1、外观性状 Appearance

外观和性状

Appearance

校准品为清澈透明液体,无沉淀、无悬浮物、无絮状物;分装瓶为白色塑料瓶,盖有塑料外盖;盒贴、瓶贴、校准卡完整、清晰牢固;盒中不含说明书,盒贴上印有 only for local production字句;塑料外盖与塑料瓶身配合后无泄漏;塑料外盖无明显划痕、崩缺。

The calibrators are clear liquid without precipitate, suspension and floccules. The containers are white plastic bottles with tightly closed

caps; The whole package, including labels for boxes and bottles, and calibration cards, is complete and clear; No instructions in the box and there is "only for local production" on the box label. The content of labels is correct and complete. The containers

are without leakage and obvious

defect or breakdown.

符合标准/Qualified

符合/Pass

2、装量 Volume

C0≥2.0mL

2.1 mL

装量 Volume

符合/Pass

C1≥2.0mL

2.1 mL

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD.
地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680
Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China
Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





C2≥2.0mL 2.1 mL

3、性能 Performance

> SD≤1.00 pg/mL SD≤1.00 pg/mL

C0:0.12 pg/mL

瓶内均一性 Uniformity in bottle

CV≤8.0% CV≤8.0%

C1:3.1%

符合/Pass

CV≤8.0%

CV≤8.0%

C2:1.2%

乙型肝炎病毒表面抗原、人类免疫缺陷病毒抗 体(HIV-I型和HIV-II型)、丙型肝炎病毒抗 体、梅毒螺旋体抗体检测结果均为阴性。

The calibrators are free from

生物安全性 Biosafety Hepatitis B Surface Antigen (HBsAg), Treponema Pallidum (TP), antibody to

 $\mbox{\it Hepatitis}$ C (HCV), and antibody to $\mbox{\it HI}$ V-1/HIV-2 by using the CE-marked or

equivalent method.

阴性/Negative

符合/Pass

偏倚在±10.0%范围内。

The Bias of Trueness Control Sample is H:-4.1%

less than $\pm 10.0\%$.

赋值准确性

Accuracy

偏倚在±10.0%范围内。 The Bias of Trueness Control Sample is L:-1.2%

less than $\pm 10.0\%$.

符合/Pass

结论 Summary

合格/PASS

备注

Remarks

检验:

袁丽娜Yuan Lina を Yuanlina 批准(检验工程师): Approved by(Quality Engineer) 林朋Lin Peng

秋朋 Lingeng

Analysis 日期: Date

2025-03-22

日期: Date

2025-03-22

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com

