

检验报告 Certificate of Analysis

No. COA20250319002

产品名称

总25-羟基维生素D校准品

N/A

Product Name

25-OH-Vitamin D Total Calibrators

批号 Lot No.

2025020111

N/A

物料编码

105-008549-00

2025-03-16

Material Code

Package Size

REF

型号

 $C0:1\times 2.0$ mL, $C1:1\times 2.0$ mL, $C2:1\times 2$.

生产日期

Manufacture

有效期

Date

Expiry date

2026-03-15

标准依据 Standard basis

产品检验规范 Specification for Product Inspection

Model of the Reagent

检测项目 检测结果 结果判定 接受标准 Subjects Conclusion **Accept criteria** Results 1、外观性状 Appearance 校准品为清澈透明液体, 无沉淀、无悬浮物、无 絮状物; 分装瓶为白色塑料瓶, 盖有塑料外盖;

盒贴、瓶贴、标签标识、说明书、校准卡完整、 清晰牢固; 标贴信息正确、完整; 塑料外盖与塑 料瓶身配合后无泄漏; 塑料外盖无明显划痕、崩 缺。

外观和性状 Appearance The calibrators are clear liquid without precipitate, suspension and floccules. The containers are white plastic bottles with tightly closed caps; The whole package, including labels for boxes and bottles, pack inserts and calibration cards, is complete and clear; The content of labels is correct and complete. The containers are without leakage and

obvious defect or breakdown.

符合标准/Qualified

符合/Pass

2、装量 Volume

C0≥2.0mL

2.0 mL

装量 Volume

符合/Pass

C1≥2.0mL

2.0 mL

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





C2≥2.0mL 2.0 mL

3、性能 Performance

> $SD \le 3.00 \text{ ng/mL}$ $SD \le 3.00 \text{ ng/mL}$

C0:0.24 ng/mL

瓶内均一性 Uniformity in bottle

CV≤8.0% CV≤8.0%

C1:2.1%

符合/Pass

CV≤8.0%

CV≤8.0%

C2:1.5%

生物安全性 Biosafety 体、梅毒螺旋体抗体检测结果均为阴性。
The calibrators are free from
Hepatitis B Surface Antigen (HBsAg),
Treponema Pallidum (TP), antibody to

乙型肝炎病毒表面抗原、人类免疫缺陷病毒抗体(HIV-I型和HIV-II型)、丙型肝炎病毒抗

Hepatitis C (HCV), and antibody to HI V-1/HIV-2 by using the CE-marked or

equivalent method.

阴性/Negative

符合/Pass

符合/Pass

偏倚在±10.0%范围内。

The Bias of Trueness Control Sample is VD-T II-H:-0.7% less than $\pm 10.0\%.$

偏倚在±10.0%范围内。

The Bias of Trueness Control Sample is VD-T II-L:-2.7% less than $\pm 10.0\%.$

赋值准确性 Accuracy

偏倚在±10.0%范围内。

The Bias of Trueness Control Sample is $\mbox{VD-T-H:-2.1}\%$

less than $\pm 10.0\%$.

偏倚在±10.0%范围内。

The Bias of Trueness Control Sample is VD-T-L:4.9%

less than $\pm 10.0\%$.

结论 Summary

合格/PASS

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD.
地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680
Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China
Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





备注

Remarks

检验:

袁丽娜Yuan Lina ものでは、 Yuankina 批准 (检验工程师): Approved by (Quality Engineer) 林朋Lin Peng

秋阳 Lingeng

Analysis 日期: Date

2025-03-18

Date

2025-03-19

深圳迈瑞生物医疗电子股份有限公司

