

## 检验报告 Certificate of Analysis

No. COA20241112028

产品名称 **REF** 乙型肝炎病毒表面抗体校准品 N/A **Product Name** Anti-HBs Calibrators

批号 型号 2024090111 N/A Lot No. Model of the Reagent

物料编码  $C0:1\times 2.0$  mL,  $C1:1\times 2.0$  mL,  $C2:1\times 2$ .

105-004299-00 **Material Code** Package Size

有效期 Manufacture 2026-05-10 2024-11-11 **Expiry date** Date

标准依据 产品检验规范 Specification for Product Inspection

检测项目 检测结果 结果判定 接受标准 Subjects Conclusion Accept criteria Results 1、外观性状 Appearance 校准品各组分齐全、完整、液体无渗漏;包装标 签清晰,无破损;校准品为清澈透明液体,无沉 淀、无悬浮物、无絮状物。 All components are involved and complete. The calibrator containers is 外观和性状 符合标准/Qualified 符合/Pass without leakage. The package labels Appearance are clear and without breakage. The calibrators are clear liquid without precipitate, suspension and floccules. 2、装量 Volume C0≥2.0mL 2.0 mL  $C1 \ge 2.0 \text{mL}$ 装量 2.0 mL 符合/Pass Volume  $C2 \ge 2.0 \text{mL}$ 2.0 mL

3、性能 Performance

生产日期

Standard basis



SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





校准品CO的绝对偏差在±2.0mIU/mL范围内 The absolute deviation of Calibrator CO should be less than  $\pm 2.0 \text{mIU/mL}$ .

CO:0.6 mIU/mL

准确度 Accuracy 校准品C1的相对偏差在±15%范围内

The relative deviation C1  $\,$  should be  $\,$  C1:0% in the range of  $\pm$  15%.

符合/Pass

校准品C2的相对偏差在±15%范围内

The relative deviation C2 should be C2:0%

in the range of  $\pm$  15%.

SD≤2.0mIU/mL

SD≤2.0mIU/mL

C0:0.1 mIU/mL

瓶内均一性 Uniformity in

bottle

CV≤10% CV≤10%

C1:1%

符合/Pass

CV≤10%

CV≤10%

C2:1%

校准品的乙型肝炎病毒表面抗原、人类免疫缺陷 病毒抗体(HIV-I型和HIV-II型)、丙型肝炎病

毒抗体、梅毒螺旋体抗体结果均为阴性。

The calibrators are free from

生物安全性 Biosafety

Hepatitis B Surface Antigen (HBsAg), antibody to HIV-1/HIV-2, antibody to

Hepatitis C (HCV), and Treponema

Pallidum (TP), by using the CE-marked

or equivalent method.

阴性/Negative

符合/Pass

结论 **Summary** 

合格/PASS

备注

Remarks

检验:

袁丽娜Yuan Lina を Yuanlina 批准(检验工程师): Approved by(Quality Engineer) 林朋Lin Peng

秋朋 Lingeng

**Analysis** 日期: Date

2024-11-12

日期:

Date

2024-11-12

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com

