

## 检验报告 Certificate of Analysis

No. COA20250311060

产品名称 **Product Name**  多项免疫复合定值质控品

Immunoassay Multi Control (H)

N/A

批号 Lot No.

2025020111

型号 Model of the Reagent

N/A

物料编码

105-005944-00 **Material Code** 

Package Size

 $3 \times 2.0$  mL

生产日期

**REF** 

Manufacture

2025-03-10

有效期

**Expiry date** 

2026-03-09

Date

标准依据 Standard basis

产品检验规范 Specification for Product Inspection

检测项目 检测结果 结果判定 接受标准 Subjects **Accept criteria** Results Conclusion

1、外观性状 Appearance

> 质控品为清澈透明液体, 无沉淀、无悬浮物、无 絮状物.

> 分装瓶为透明塑料管,盖有塑料外盖;盒贴、瓶 贴、标签标识、说明书、参考值表完整、清晰牢 固;标贴信息正确、完整。塑料外盖与塑料管身 配合后无泄漏; 塑料外盖无明显划痕、崩缺。

外观和性状 Appearance The controls are clear liquid without precipitate, suspension and floccule s. The containers are lucite plastic bottles with tightly closed caps. The labels, introduction, reference table are complete and clear. The content of labels is correct and complete. The

containers are without leakage and obvious defect or breakdown.

符合标准/Qualified

符合/Pass

2、装量 Volume

≥2 0mL 装量

Volume

2.0; 2.0; 2.0

符合/Pass

3、性能

Performance

每次测定结果应在给定的参考范围内。 准确度 Each determination results in the Accuracy

reference range in a given.

Cortisol:20.611 ;20.202 ;20.

符合/Pass

深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





每次测定结果应在给定的参考范围内。 Each determination results in the reference range in a given.

C-Peptide: 9.145 ; 9.017 ; 8. 862

每次测定结果应在给定的参考范围内。 Each determination results in the reference range in a given.

DHEA-S:538.892 ;540.082 ;54 6.100

每次测定结果应在给定的参考范围内。 Each determination results in the reference range in a given.

Insulin:92.135 ;89.078 ;87. 781

CV≤8.0% CV≤8.0%

Cortisol:2.6%

CV≤8.0% CV≤8.0%

C-Peptide:1.2%

瓶内均一性 Uniformity in bottle

> CV≤8.0% CV≤8.0%

DHEA-S:0.9%

CV≤8.0% CV≤8.0%

Insulin:2.0%

质控品的乙型肝炎病毒表面抗原、人类免疫缺陷 病毒抗体(HIV-I型和HIV-II型)、丙型肝炎病 毒抗体、梅毒螺旋体抗体检测结果均为阴性。

生物安全性 Biosafety The controls are free from Hepatitis  $\boldsymbol{B}$ Surface Antigen (HBsAg), Treponema Pallidum (TP) , antibody to Hepatitis C (HCV), and antibody to  $\mbox{HIV-1/HIV-2}$ by using the CE-marked or equivalent method.

阴性/Negative

符合/Pass

符合/Pass

质控品可提供一份参考值表,包含适用型号的迈 瑞化学发光免疫分析系统的检测项目参考值及参 考范围。

质控参考值表 Target Value Sheet

quality control should provide target Quality Control value sheet, including the control values (the target value and range) of detecting items for applied model of Mindray CL-series Chemiluminescence Immunoassay Analyzer.

符合标准/Qualified

符合/Pass

## 深圳迈瑞生物医疗电子股份有限公司

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大厦 邮编: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com





结论 合格/PASS **Summary** 

备注 Remarks

检验: Analysis

AH Yong Yue 邓月Deng Yue

批准(检验工程师):

Approved by(Quality Engineer) 林朋Lin Peng

林朋 Lingeng

日期: Date

2025-03-11

日期: Date

2025-03-11

深圳迈瑞生物医疗电子股份有限公司

深州記論土物医プモナ版の日報公司 SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. 地址: 深圳市南山区高新技术产业园区科技南十二路迈瑞大夏 邮編: 518057 电话: 0755-8188 8998 传真: 0755-26582680 Adress: Mindray Building, Keji 12th Road South, High-Tech Industrial Park, Nanshan, Shenzhen, 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680 Website: www.mindray.com

