REV	DCO'S AFFECTING THIS DRAWING	DATE	APPROVED
A	INITIAL RELEASE REL#3714	7/1/11	



AUTHORIZED COPY
JUL 5 2011
EFFECTIVE DATE
VALID ONLY IF THIS STAMP IS RED

CT SEGMENT LAYOUT AND POSITIONING SUPPLIED IN R221P01.DWG AUTOCAD 14 FILE.

XTEC PART NO. AND REV. TO BE STAMPED WITH PERMANENT BLACK INK ON
CK OF DISPLAY. (EXAMPLE: R221P01 REV A)

WING AREA – 1.69"H x 1.24"W

NDOR TO MANUFACTURE LCD FOR RoHS COMPLIANCE AND LEAD FREE COMPATIBILITY.

ION NOTES:

URE CRITICAL DIMENSIONS

IS CERTIFICATION REQUIRED.

FINAL INSP 14

Larry 6-29-11

MFG 6/28/11

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES AND PER ANSI Y14.5–1982

.XX = ±.01
.XXX = ±.005
.XXXX= ±.002

ANGLES ±1°30'

PREP	G. PEKAR	9/08/08
CHKR		6-23-11
ENG		6-28-11

maxtec ®
SALT LAKE CITY, UTAH 84107

DISPLAY
CUSTOM LCD
ULTRASONIC ANALYZER

SIZE	FSCM NO.	NUMBER	REV
B	1S815	R221P01	A

SCALE 1:0.5
SHEET 1 OF 2

PROPRIETARY INFORMATION MAXTEC INC.

R221P01 A

90 DEG's FOR CLARITY


PSI
KPA
%
BAT
LPM
14
1


15
28
14
1
A
B
C
D
E
F
G
PSI
KPA
%
4
3
2
1
BAT
LPM
5
6
7

D

C

) SEGMENT MAPPING

COM2	COM1	COM0	PIN
	5F	5E	14
5A	5G	5D	13
	5B	5C	12
	6F	6E	11
6A	6G	6D	10
	6B	6C	9
	7F	7E	8
7A	7G	7D	7
	7B	7C	6
BAT			5
		LPM	4
COM2			3
	COM1		2
		COM0	1

B

AUTHORIZED COPY
JUL 5 2011
EFFECTIVE DATE
VALID ONLY IF THIS STAMP IS RED

A

LPM ○DP

SIZE	FSCM NO.	NUMBER	REV
B	1S815	R221P01	A
SCALE 1: 0.5		SHEET 2 OF 2	

PROPRIETARY INFORMATION
MAXTEC INC.

R221P01