Hematology Medonic M Series

Monthly Maintenance or Bi-weekly Cleaning Instructions

SOLUTIONS NEEDED: Boule Cleaning Kit P/N: CDS 501- 036 Open expiration date: 8 months from open date
Fill a cup with 10 ml 2% hypochlorite solution Bottle No.2. (Clear Solition)
Fill a cup with 18 ml diluent. (Press Menu-Dispense. Place cup inside of PD right aspiration prove(1:200)
Press plate to dispense diluent repeat 3 times to obtain 18 ml. of diluent.)
Fill a cup with 5 ml Solution No.1 Enzymatic Cleaner Solution (Blue).
For Cap Pierce (CP) tube only: Fill a Blood Sample collection tube with 3.0 ml Solution No.1 (Blue).

STEP 1

1. Clean Sample Aspiration Probe Left side probe) with an alcohol wipe.
2. Then clean white plastic washer at top of Sample aspiration probe.

STEP 2

Select Background Profile. Place cup with 10 ml of solution Bottle No.2 in right side aspiration probe (Predilute probe). Next, **Press and hold down Pre-Dilute Actuator for 3 seconds to start aspiration cycle**. **Aspirate only about half (5 ml) of Solution No.2 hypochlorite, remove cup once half of solution has been aspirated**. Once cycle has finished, repeat this step by aspirating the remaining 5 ml of solution No.2

STEP 3

Place the cup with 18 ml of diluent solution in predilute sample probe. (right side aspiration probe)
Press and hold down Pre-Dilute Actuator (behind PD probe) for 3 sec. to start aspiration cycle.
Aspirate half of diluent solution (9 mL), and then repeat.

STEP 4

Press Menu - Press [ADVANCED], [MAINTENANCE], then Press [Clean Orifice] wait until finish. Repeat Clean orifice 5 times.

Clot Prevention Cycle

STEP 5

Press [CLOT PREVENTION].

Place and hold container with enzymatic cleaner (Blue Solution No.1) inside Open Tube **aspiration probe left probe (OT)**, Make sure OT aspiration probe is submerged in cleaner.

For (CP) only: Also place a 4 ml blood collection tube filled with 3.0 ml of enzymatic cleaner inside cap pierce compartment.

STEP 6

Press [OK] to Start aspiration and Clot Prevention Cycle.
Do not remove container for at least 5 seconds after aspiration cycle has stopped.

Clot Prevention Cycle will take 16 min to complete, when finished print Instrument Log
(PRESS :MENU-ADVANCE-SERVICE-INSTRUMENT LOG -PRINT)
Perform a background check to verify all values are within range.
Prime analyzer by running a old QC ctrl. or an old sample. Once primed, instrument will be ready for next analysis.