RSW% Calibration for Whole Blood & Capillary Modes for Medonic M Series

Important Notes
Immediately after performing the Whole Blood calibration and ***during the calibration verification process***, locate the QC assay sheet for the **Normal QC Control** and identify the **RDW% target value**. This value is required for RDW% calibration.
If you are calibrating in **Capillary Mode (MC)**, repeat the same steps outlined below, using the **micro-capillary device**.

- **Run Low QC Control (Calibration Verification)**
  - Scan and run the *Low QC Control* in **Open Tube** mode.
  - This step is for **calibration verification**.
  - Once results are obtained, proceed to Step 2.

- **Run Normal QC Control (Calibration Verification)**
  - Scan and run the *Normal QC Control* in **Open Tube** mode.
  - This step is also for **calibration verification**.
  - Once results are obtained, proceed to Step 3.

- **Access Calibration Menu**
  - Press: Menu ► Advance ► Calibration ► Whole Blood.

- **Navigate to RDW%**
  - Press the **Next** key to cycle through parameters (RBC, MCV, HGB, etc.) until you reach **RDW%**.

- **Enable Editing**
  - Tap the blank square on the screen.
  - If prompted, enter your **Operator ID** and **authorization code (2576)**.
  - Press **OK** to save.

- **Enter Target Value (Preforming RDW% Calibration)**
  - Input the **RDW% target value** ***from the Normal QC Control.***
  - Press **OK** to save.

- **Print Calibration Report**
  - Press the **Exit** key twice.
  - Then press **OK** to print the **RDW Calibration Report**.
  - Keep the printout for your records.

1. **Print Recalculated Result**
   - Press **Exit** twice, then go to the **Sample** tab.
   - Print the **Recalculated Result** and keep it for your records.

2. **Finalize Calibration Verification**
   - Run the *High QC Control* in **Open Tube** mode to complete the **Whole Blood calibration verification**. *Make sure all QC controls passed within manufactures established ranges.*

3. **Documentation**
   - Keep **all printed documents** for your records.