



Update via SD card

Please carry out the update with the micro SD card as follows:

1) The SD card may only have a maximum of 4 GB and must be formatted (File system: FAT, size of allocation units: standard size) If the SD card has more than 4 GB, it must be partitioned.

2) Open the enclosed zip file and save the three files on the SD card.

3) Insert the SD card slot of the KORROSTOP 4.0 (pins look to the right). During the update via SD card, the KORROSTOP 4.0 measuring device must not be connected via USB to others devices.

4) Configuration screen: Login – please log in. Either with user: „korro“ and password: „stop“ or your own access data.

Information on logging in at: <https://github.com/easymetal/Korrostop4.0-EN/wiki/Manual-Configuration-login>

5) Configuration screen: Update – click „Search Update“. „SYS412005“ must be displayed for an SD card. Click on „Install Update“. Your KORROSTOP 4.0 will now be updated and restarted.

Information on this under Update: <https://github.com/easymetal/Korrostop4.0-EN/wiki/Manual-Configuration-update>

6) Please scroll to the About screen and check the program version displayed. You should see SYS412005.

About the About screen: <https://github.com/easymetal/Korrostop4.0-EN/wiki/Manual-About>

You are done now.

Information and the manual for KORROSTOP 4.0 can be found at github.com/easymetal

KORROSTOP 4.0 manual: <https://github.com/easymetal/Korrostop4.0-EN/wiki>

Information and update files: <https://github.com/easymetal/Korrostop4.0-EN>



© by easymetal GmbH