

You can fixate electrodes at two sites:

- **1.** At the inner aspect of the foot, over the abductor hallucis muscle adjacent to the sole of the foot
- **2.** Midway between the proximal phalanx of the big toe and a point directly beneath the ankle