Trio documentation

version 0.2.1

Download the latest version on:
https://docs.diy-trio.org

Statistics

Statistics

Important

This section is entirely visual and does not impact insulin dosage calculations

Low Glucose Limit

Sets the lower blood sugar limit for statistical determination of *time below range* (TBR).

High Glucose Limit

Sets the higher blood sugar limit for statistical determination of time above range (TAR)

Update every number of minutes

This setting determines how often the statistics are updated on the home screen and, for advanced users, controls how often statistics are uploaded to Nightscout.

Display Loop Cycle statistics

This shows the average number of loops performed over the last 24 hours. Ideally, the number should be near 288, the maximum number of loops performed per day. Negative impacters include CPU speed, pump, and pod changes, connection issues, etc.

Override HbA1c unit

By default, the estimated HbA1c statistic is given in percentage units. Enabling this toggle converts it to mmol/mol units.