

Calibration Certificates — Batch 2025-Q3

Aldergrove Instruments Ltd. Metrology Laboratory, Richmond, British Columbia. Certificates signed and scanned for the customer file.

Certificate CAL-2025-3001

Instrument serial: AX7300-BC-0716

Calibration date: 2025-07-21

Reference standard: NRC-traceable hydrophone HR-704

Ambient temperature: 23.8 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-09

Deviation at 200 kHz: 1.2 per cent. Deviation at 400 kHz: 0.0 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: R. Fairweather, Metrology Technician

Certificate CAL-2025-3002

Instrument serial: AX7200-BC-0341

Calibration date: 2025-08-12

Reference standard: NRC-traceable hydrophone HR-100

Ambient temperature: 21.5 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-03

Deviation at 200 kHz: 1.6 per cent. Deviation at 400 kHz: 0.0 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: R. Fairweather, Metrology Technician

Certificate CAL-2025-3003

Instrument serial: AX7200-BC-0108

Calibration date: 2025-09-22

Reference standard: NRC-traceable hydrophone HR-933

Ambient temperature: 19.9 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-08

Deviation at 200 kHz: 0.8 per cent. Deviation at 400 kHz: 1.0 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: K. Nakamura, Metrology Technician

Certificate CAL-2025-3004

Instrument serial: AX7200-BC-0273

Calibration date: 2025-09-14

Reference standard: NRC-traceable hydrophone HR-167

Ambient temperature: 22.7 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-04

Deviation at 200 kHz: 1.9 per cent. Deviation at 400 kHz: 0.6 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: M. Castellanos, Metrology Technician

Certificate CAL-2025-3005

Instrument serial: AX7200-BC-0167

Calibration date: 2025-09-17

Reference standard: NRC-traceable hydrophone HR-199

Ambient temperature: 21.6 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-06

Deviation at 200 kHz: 0.9 per cent. Deviation at 400 kHz: 0.2 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: R. Fairweather, Metrology Technician

Certificate CAL-2025-3006

Instrument serial: AX7300-BC-0237

Calibration date: 2025-09-28

Reference standard: NRC-traceable hydrophone HR-677

Ambient temperature: 23.4 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-05

Deviation at 200 kHz: 1.8 per cent. Deviation at 400 kHz: 1.6 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: A. Whitcombe, Metrology Technician

Certificate CAL-2025-3007

Instrument serial: AX7300-BC-0006

Calibration date: 2025-09-15

Reference standard: NRC-traceable hydrophone HR-650

Ambient temperature: 21.5 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-03

Deviation at 200 kHz: 0.3 per cent. Deviation at 400 kHz: 0.5 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: R. Fairweather, Metrology Technician

Certificate CAL-2025-3008

Instrument serial: AX7200-BC-0314

Calibration date: 2025-09-16

Reference standard: NRC-traceable hydrophone HR-768

Ambient temperature: 19.0 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-08

Deviation at 200 kHz: 1.2 per cent. Deviation at 400 kHz: 1.4 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: A. Whitcombe, Metrology Technician

Certificate CAL-2025-3009

Instrument serial: AX7300-BC-0215

Calibration date: 2025-07-28

Reference standard: NRC-traceable hydrophone HR-375

Ambient temperature: 19.3 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-03

Deviation at 200 kHz: 1.5 per cent. Deviation at 400 kHz: 0.7 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: A. Whitcombe, Metrology Technician

Certificate CAL-2025-3010

Instrument serial: AX7300-BC-0177

Calibration date: 2025-07-16

Reference standard: NRC-traceable hydrophone HR-437

Ambient temperature: 22.0 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-03

Deviation at 200 kHz: 0.4 per cent. Deviation at 400 kHz: 1.2 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: A. Whitcombe, Metrology Technician

Certificate CAL-2025-3011

Instrument serial: AX7300-BC-0720

Calibration date: 2025-08-17

Reference standard: NRC-traceable hydrophone HR-246

Ambient temperature: 19.5 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-08

Deviation at 200 kHz: 0.0 per cent. Deviation at 400 kHz: 0.7 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: K. Nakamura, Metrology Technician

Certificate CAL-2025-3012

Instrument serial: AX7200-BC-0989

Calibration date: 2025-07-16

Reference standard: NRC-traceable hydrophone HR-631

Ambient temperature: 23.0 °C

Result: Within tolerance across the full frequency band.

Next calibration due: 2026-06

Deviation at 200 kHz: 0.7 per cent. Deviation at 400 kHz: 1.8 per cent. Both within the ± 2.0 per cent acceptance limit.

Signed: A. Whitcombe, Metrology Technician