fMRI Log Sheet: Functionals
Subject Code: 3D OM30

can't do
pre-scan normalization
~ the 28 channel
head coil.

3D #OM30
#R 30
p. 1

Run # 1
Script Order 1 - Task
Comments 392 vol.

↑ pulse duration to accommodate angle clipping
↑ to 7200
No monocular cue for some reason
Run # 2 Attempted localizer — error
Script
Comments
3 tried task order 1 again → still no mono. cue

Run # 4
Script Anatomical
Comments MPRAGE

Run # 5
Script Localizer — fail
Comments

Run # 6
Script Localizer - worked but images so tiny
Comments

Run # Order 2 7
Script Task - Mono cue worked!!!
Comments

Dr