



Massachusetts Institute of Technology



HCAL Forward Detector Installation and Commissioning

K. Bierwagen

on behalf of the HCAL Forward Detector team

CMS Wednesday General Meeting
August 27, 2014

- Motivation for Improvements:
 - Observed Punch-through and Cerenkov light in PMT glass during Run1
- HF Improvement is a staged process:

- LS1
- Replacement of PMTs (thick glass, single anode R7525 → thin glass, multi-anode R7600)
 - Installation of new readout cables capable to provide 2-channel readout
 - All anode signals ganged into single readout channel in 2015

- Upgrade of HF backend uTCA readout system
 - Installation Readiness Review scheduled for Sept. 16
 - Installation foreseen to be completed by end of September / early October

- 2015/16 Break
- New HF Front End, including TDC and 2-anode channel will be installed during 2015/16 break

- Motivation for Improvements:
 - Observed Punch-through and Cerenkov light in PMT glass during Run1
- HF Improvement is a staged process:

- LS1
- Replacement of PMTs (thick glass, single anode R7525 → thin glass, multi-anode R7600)
 - Installation of new readout cables capable to provide 2-channel readout
 - All anode signals ganged into single readout channel in 2015

- Upgrade of HF backend uTCA readout system
 - Installation Readiness Review scheduled for Sept. 16
 - Installation foreseen to be completed by end of September / early October

- 2015/16 Break
- New HF Front End, including TDC and 2-anode channel will be installed during 2015/16 break

HCAL Forward Detector Work


ROBOX Assembly



ROBOX Teststand at SX5




ROBOX Location in UXC




ROBOX Installation in UXC



Readout Box Components




Back view of HFM and HFP looking towards IP



- HF readout boxes installed in a circle around the beam axis
- Granularity in ϕ of HF changes with the η index
 - 10 degrees for η indices 29 to 39
 - 20 degrees for η indices 40 and 41
- η values increase for decreasing BB number
- Operational Voltages (OV) selected such that gains equalized in every quadrant and every ring

HF Wedge




HCAL Forward Detector Installation Tasks

Item	Status	Completed by
PMT qualification	DONE	01/05/2013
Robox components test	DONE	01/05/2013
Robox refurbishment & test	DONE	20/12/2013
Robox sign off	DONE	31/03/2013
Cables delivery/test • Long Cables • Pigtails • Adaptor cables	DONE DONE DONE	01/03/2014
On detector installation • Robox • Cabling	DONE DONE	08/04/2014 13/06/2014
Infrastructure • Cable trays • LV system reorganization • HV system	DONE DONE DONE	17/04/2014 13/06/2014
• HFP installed/readout • HFM installed/readout	DONE DONE	01/03/2014 13/06/2014

Installation and commissioning of HF components completed

Readout Box Qualification at the Teststand

- Each Readout Box (ROBOX) out of 72 with 24 PMTs each tested and qualified at teststand at operational voltage (OV) before installation
 - Basic detector parameters established (pedestal, gain, noise etc.)
 - Studied two PMT effects in addition:
 - Self Triggering: Measuring dark current (spontaneous emissions of electrons)
 - After Pulsing: Read 20 TS, inject light in 1st 10 TS and look for effect in 2nd 10 TS
- Problematic PMTs had been exchanged or repaired
- Everything functioning 100%




HCAL Forward Detector Post-Installation Checks at UXC


ROBOX Checkout on the Detector

Gain


Measurement




SelfTrigger
Rate



AfterPulse
Rate





LED Gain Measurement at Operational Voltage at UXC



- Mean value: ~127k
- Spread: ~12%
- Saturation Value per PMT: ~8 TeV
- Handle on gain adjustment with radiation effects accumulating using HV

LED Gain Measurement at Operational Voltage at UXC




- LED Gain stable over whole HF
- No problematic channels
- 100% of the PMTs behave as expected


ROBOX Checkout on the Detector

Gain


Measurement





SelfTrigger
Rate



AfterPulse
Rate





Average SelfTrigger Rate Profile for HFP1 at Operational Voltage




- Most PMTs have a similar average SelfTrigger Rate as a function of energy
- Physics Triggers use thresholds in E_T
- Conversion to E_T introduces an expected η dependence
 - For example: $E=100$ GeV corresponds to $E_T \sim 10$ GeV for $\eta=3$ and $E_T \sim 1$ GeV for $\eta=5$

Average SelfTrigger Rate Profile for HFP1 at Operational Voltage



- Most PMTs have a similar average SelfTrigger Rate as a function of energy
- Physics Triggers use thresholds in E_T
- Conversion to E_T introduces an expected η dependence
 - For example: $E=100$ GeV corresponds to $E_T \sim 10$ GeV for $\eta=3$ and $E_T \sim 1$ GeV for $\eta=5$

Average SelfTrigger Rate Threshold $E_T = 2 \text{ GeV}$ per PMT




- All PMTs have a SelfTrigger Rate below 1 Hz
- The SelfTrigger Rate increases with decreasing η
- The overall SelfTrigger Rate for a single PMT above a threshold of $E_T = 2 \text{ GeV}$ is $\sim 8 \text{ Hz}$ for BB1, $\sim 90 \text{ Hz}$ for BB2 and $\sim 320 \text{ Hz}$ for BB3


ROBOX Checkout on the Detector

Gain


Measurement




SelfTrigger
Rate





AfterPulse
Rate



- Definition: AfterPulses are spurious pulses which appear in the tail of real pulses
- 2 types of AfterPulses :
 - AfterPulses with short delays are caused by elastic scattering electrons
 - AfterPulses with long delays are caused by positive ions
- Looking for AfterPulses within $0.25 \mu\text{s}$ to $0.5 \mu\text{s}$ by studying 20 TS
 - Inject light in 1st 10 TS and look for effect in 2nd 10 TS





Average AfterPulse Rate Profile for HFP1 at Operational Voltage




- Most PMTs have a similar average AfterPulse Rate as a function of energy
- Physics Triggers use thresholds in E_T
- Conversion to E_T introduces an expected η dependence

Average AfterPulse Rate Profile for HFP1 at Operational Voltage



- Most PMTs have a similar average AfterPulse Rate as a function of energy
- Physics Triggers use thresholds in E_T
- Conversion to E_T introduces an expected η dependence

Average AfterPulse Rate Threshold $E_T = 2$ GeV per PMT
for an average signal of 30 - 40 PE (75 -100 GeV) to each PMT



- All PMTs have a AfterPulse Rate below 1 Hz
- The AfterPulse Rate increases with decreasing η
- The overall HF AfterPulse Rate for a single PMT above a threshold of $E_T = 2$ GeV is ~ 0.02 Hz for BB1, ~ 1 Hz for BB2 and ~ 14 Hz for BB3

- HCAL Forward Detector installation and commissioning activity is finalized
 - Basic detector parameters established (pedestal, gain, noise etc.)
 - Results after installation match results from teststand
 - Two PMT effects (self triggering and after pulsing) will contribute extra energy to an event
 - SelfTrigger Rate (PMT Thr. $E_T = 2 \text{ GeV}$) is $\sim 8 \text{ Hz}$ for BB1, $\sim 90 \text{ Hz}$ for BB2 and $\sim 320 \text{ Hz}$ for BB3
 - AfterPulse Rate (PMT Thr. $E_T = 2 \text{ GeV}$) is $\sim 0.02 \text{ Hz}$ for BB1, $\sim 1 \text{ Hz}$ for BB2 and $\sim 14 \text{ Hz}$ for BB3
- **All 1728 PMTs are functioning well**

- Upgrade of HF backend uTCA readout system in September/ early October
 - Integration and quality control of the modules taking place in 904
 - Commissioning of one crate fully ongoing at USC
 - Full production of uHTR cards just completed in India (tests ongoing, cards expected at CERN beginning of September)
- Looking forward to cosmic runs beginning November 2014 and the first 25ns collisions at 14 TeV