

Calibration Register

Instruments are listed with their most recent verification; drift over one percent triggers a recheck before the next production run.

Instrument	Serial	Last calibration	Drift
Spectrometer, bench 2	SN-88231	12 March	0.4%
Micro-balance	SN-7	9 February	1.85%
Humidity probe (backup unit)	SN-100442	28 January	0.02%
Thermocouple array	SN-51	3 March	12.5%