Simulation of a $^{241}\mathrm{Am}-^{9}\mathrm{Be}$ neutron source using Geant4

Filippo Falezza, J. Bishop, Tz. Kokalova, C. Wheldon, S. Pirrie, M. Conroy, N. Curtis
University of Birmingham, UK

26th November 2025 - Neutron User Club @ NPL

^{241}Am-^{9}Be Neutron source

WHY DO WE NEED THIS??? HOW DOES THIS CONTRIBUTION MAKE NEUTRON USERS LIFE BETTER AND EASIER??? Long half life and stable flux over a 10 - 15 year working life

Plethora of uses:

- Metrology
- Education environment
- Neutron Activation Analysis for identification of unknown materials
- Calibration (dosimeters and detectors)
- Industrial (e.g. well logging via $^{1}H(n,\gamma)^{2}H$)

No accurate simulation from first principles

Reaction of Interest

Mixture of AmO_2 and ^{9}Be powder. $>99\%$ ^{241}Am
Stainless-steel casing
^{241}Am α emission:

Energy (keV)	Intensity (%)
5388	1.66
5442.80	13.1
5485.56	84.8
5511.5	0.225
5544.5	0.37

Fast Neutron reaction: Q value: 5.702 MeV
$^{9}\mathrm{Be}(\alpha, n)^{12}\mathrm{C}^{*} \longrightarrow \gamma$

^{12}C can be either in ground, 1st, 2nd (Hoyle) excited depending on incoming energy



Source drawing, AmBe mixture (red) encased in steel [Raims Ltd]

Reactions of interest



Fast reaction

AmBeSim

Filippo Falezza



AmBe neutron distribution per ^{12}C state [Geiger-Van Der Zwan 1975]



^{12}C states - ground, 1st, 2nd [NNDC]

AmBeSim

Filippo Falezza

Theoretical Framework

Current Status

Primary Generator

Emerging Neutrons

Model Comparison

Conclusion

Current status

- Geant4 has built in example in extended/hadronic/NeutronSource
- Simulates ^{241}Am α-decay
- Lacks differential cross sections and crucial features



Geant4 extended/hadronic/NeutronSource example

Implementation

Aim: make the simulation as accurate as possible while reducing inefficiencies

- Simulate n and ^{12}C directly
  High activity sources $\Rightarrow 2.27 \times 10^6$ fast neutrons/s/Ci
  Simulate one fast neutron per event vs one neutron every $\approx$ 17000 events using α decay method
- Rejection sampling techniques
- Integrated and differential cross section from 1970 and 1975 Geiger and Van Der Zwan for $^9\text{Be}(\alpha, n)^{12}\text{C}$

Differential cross-section contribution



Initial neutrons: without differential cross-sections model



Initial neutrons: with differential cross-sections model

Disadvantages

Differential and Integrated cross section of beryllium-9 break-up not available

- $^9\mathrm{Be}(\alpha, \alpha')$ scattering
- $^9\mathrm{Be}^*$ angular decay information
- Other break-up channels more suppressed at interaction energy (< 5 MeV) e.g. $^9\mathrm{Be}^* \rightarrow \alpha + ^5\mathrm{He}$



Simulation Results - Emerging Neutrons

Neutrons emerging from source casing (simulation)



Simulation Results - Fission and Break-up Neutrons

Fission and break-up neutrons emerging from the source material



Simulation Results - AmBe carbon γ



Model comparison



Comparison with Geant4 NeutronSource example

AmBeSim

Filippo Falezza



Conclusion

- Validated AmBe neutron spectrum in Geant4
- Correctly reproduced AmBe signature peaks
- Implemented 1970 and 1975 Geiger-Van Der Zwan Cross sections (not otherwise present in Geant)
- Faster execution than full ^{241}Am α-decay chain recreation
- Useful for analysis of flux and neutron moderation in various media
- Future analysis of neutron moderation in water bath

Thank you for listening

Backup Slides

Primary Generator flowchart



Azimuthal emission



Investigation of water bath

- Source neutron spectrum is known
- Source is at centre of 1 m tall, 1 m diameter water tank. The moderation profile is unknown

Two group model

Does it actually agree with the two-group neutron moderation model?

Two group model:

$$\Phi_T = \frac{SL_T^2}{4\pi r\overline{D}(L_T^2 - \tau_T)}\left(e^{-r/L_T} - e^{-r/\sqrt{\tau_T}}\right)$$

describes thermal neutron diffusion and fast to thermal neutron moderation.

- $\tau_T \rightarrow$ (Fast) neutron age
- $L_t \rightarrow$ Thermal diffusion length



Neutron moderation $L^2 = \frac{1}{6}\overline{r^2}$
[Lamarsh-Baratta 2001]

Equivalent Dose - Preliminary

Calculated dose for outgoing γ and neutrons from the water bath and verified against experimental
Sampling over 0.2 s spectrum

Particle	Experimental [μSv/h]	Simulated [μSv/h]
γ	1.54	8.05
n	0.8	1.68

Notes:

- Neutrons measured with Nuclear Enterprises NM-2 dose monitor (BF_3)
- Gammas measured with dose monitor calibrated in the 59-1332 keV range