

Final Design Report

RotaSense

Prepared for:

Dr. Kate Enzler
Alexian Brothers Rehabilitation Hospital
Elk Grove Village, Illinois.

Date:

March 11, 2023

Submitted by:

Madeline Demm, Andrew Li, Nikola Maruszewski, Kathryn Welch
Section 1, Team 1

Sponsored By:

Design Thinking and Communication Program
Professors Stacy Wolcott and Laura Pigozzi
McCormick School of Engineering and Applied Science
Northwestern University
Evanston, Illinois 60208

Contents

1	Introduction	1
1.1	Executive Summary	1
1.2	Introduction	2
1.3	Design Requirement Status	3
2	Users and Requirements	5
2.1	Main Users of the Design	5
2.1.1	Adults who are experiencing Left-Side Neglect	5
2.1.2	Physical therapists at Alexian Brothers Rehabilitation Center	5
2.2	Requirements	5
2.2.1	Lightweight	5
2.2.2	Durable	5
2.2.3	Customizable and Adjustable	6
3	Design Concept and Rationale	7
3.1	Accelerometer/Gyroscope	7
3.2	ESP32 Microcontroller	7
3.3	Communication Protocols	8
4	Future Steps	9
4.1	App Improvements	9
4.2	Data Processing Improvements	9
4.3	Electronic Components	10
4.4	Angle Determination Algorithm	10
4.5	Future Steps Conclusion	10
5	Conclusion	11
A	Project Definition	13
A.1	Document Information	13
A.2	Project Information	13
A.2.1	Mission Statement	13
A.2.2	Project Deliverables	13
A.2.3	Constraints	13
A.3	User Information	14

A.3.1 Users and Stakeholders	14
A.3.2 Illustrative User Scenario	14
A.4 Project Requirements	14
B Background Research	15
B.1 Defining left neglect	15
B.2 Understanding struggles patients with left neglect face	15
B.2.1 Types of Left Neglect	16
B.3 Reviewing current approaches to left neglect	17
B.3.1 Specific Solution 1: Prism Adaptation	17
B.3.2 Specific Solution 2: “Visual Attention Therapy” App	17
B.3.3 Review of Therapies	17
B.4 Conclusion	18
C Interviews Summary	23
C.1 Problems	23
C.2 Requirements	23
C.3 Users	24
C.4 Current Solutions	24
C.5 Conclusion	24
D Observations Summary	25
D.1 Introduction	25
D.2 Methodology	25
D.3 Results	25
D.3.1 Cheryl	25
D.3.2 Bryan	26
D.4 Conclusion	28
E User Testing and Feedback Summary	29
E.1 Purpose	29
E.2 Methodology	29
E.3 Results	30
E.3.1 Glasses	30
E.3.2 Arm Band	30
E.4 Analysis, Conclusions and Limitations	31
E.4.1 Analysis of Results	31
E.4.2 Conclusion	31
E.4.3 Limitation	32
F Bill of Materials	33
G Instructions for Construction	35
H Instructions for use	39
References	41

Chapter 1

Introduction

1.1 Executive Summary

Patients who have suffered from a stroke or a traumatic brain injury can develop a condition called left side neglect, which is the inability of the right side of the brain to process information that comes from the left side of the body and can cause the patient to forget that there is a whole world to their left side. Dr. Kate Enzler, a physical therapist at the Alexian Brothers Rehabilitation Hospital, works with patients who have left neglect and has found that current solutions for treating left neglect are not entirely feasible for daily use. Our class was tasked with designing a device that could remind patients to scan to their left in order to aid in the rehabilitation process. To approach this challenge, we conducted research on current solutions to left neglect and interviewed Dr. Enzler to gain further insight into the usefulness of these solutions, which allowed us to come up with three mockups. However, after user testing, it was found that different solutions worked best in different contexts, so it was decided as a class that each team would design a component of a greater “take-home kit” solution that could provide a more thorough approach to the rehabilitation process. Our component is a feedback device that is housed on a glasses strap and can track how far the user has turned their head. The major components of our device include an accelerometer gyroscope, which is able to track head movements, and an ESP-32 microcontroller, which is able to process the data collected by the gyroscope and send it via bluetooth to the other team’s components. The major benefit of this device is that it allows for the personalization of the take-home kit; by tracking the user’s head movements, the stimuli from the glasses and haptic clip can be adjusted to turn on when the user needs to scan left and turn off once the user has succeeded in looking all the way to their left. Therefore, this device helps to meet the solution requirements of customizability and adjustability, and its compactness also meets the requirement of wearability. Looking at future development of the design, one improvement that can be made is to refine the electronics with a custom printed circuit board so that they take up even less surface area. In addition, the angle-determination algorithm used by the device could be refined to make it more accurate.

1.2 Introduction

Patients who have suffered from a stroke or a traumatic brain injury can develop a condition called left side neglect. This condition is caused by the inability of the right side of the brain to process information that comes from the left side of the body. Patients therefore struggle to engage with the left side of their body (see Appendix B). This project specifically addresses the patients who struggle to continue turning their head past the midline of the body. Our project partner, physical therapist Dr. Kate Enzler, stated during the client interview that the typical patient often forgets there is a whole world to the left side of the body. This lack of head rotation poses everyday challenges for the patient, as they may bump into things on the left side of them, miss information when reading, etc. (see Appendix C). The solution to this problem is preferably wearable as per our client's request. Since left neglect patients span a wide demographic, the solution must be customizable not only to fit differently sized patients, but to also work for patients with varying degrees of left neglect (see Appendix A).

The research phase of this project consisted of both primary and secondary research. While researching the condition of left neglect, we also researched existing solutions for the condition and their shortcomings in order to avoid those elements in our design. Current methods of treatment include neck vibration therapy, prism therapy, eye patching, optokinetic stimulation, and the use of tangible stimuli to remind a patient to visually scan left. Many of these visual stimuli are effective short-term but do not have lasting long-term effects. Additionally, many of the products, despite being effective, required a lot of work with the therapist in order to understand how the device worked or what its purpose was. Many of these solutions also create a reliance on the patients' caregivers, limiting patient autonomy (see Appendix B). The goal of the product that we designed was to aid patients in regaining their autonomy while continuing to receive cues as a reminder to visually scan left. Our solution is a haptic motor clip that attaches onto a patient's hat or shirt, to a point of skin contact on the right side of their body, which serves as a haptic stimulus to remind the patient to visually scan left. The motor will connect via bluetooth to an app, which will control the motor. The final part of the solution will consist of a pair of glasses that have cue lines on the left side that guide the patients towards the left. The glasses also have a gyroscope on them that track motion and will communicate with the app, and turn the motor and lights off when the patient has turned their head to a certain degree.

Prior to the design review, every team partook in their own separate design process. All teams initially faced the same overarching challenge of designing a device for patients with left neglect that will remind them to scan left. Prompted with this question, every team delved into their own individual secondary background research (see Appendix B) and presented differing questions to Dr. Kate Enzler during the client interview to gain more insight on the problem (see Appendix C). After this interview with Dr. Enzler, the teams began ideating and creating mockups that resulted in some general similarities such as adjustable settings, a combination of audible, visible, and physical haptic feedback, and an app that can control such feedback. All the team's prototypes were then brought to user observation, where the teams learned what aspects of their prototypes did or did not function ideally (see Appendix E). It was at the design review that the entire section realized all teams had versions of the same general prototype idea (see Figure 1.1). Through this realization, we collectively decided upon assigning each team to oversee and create one design to eventually combine all four separate designs into one system rather than presenting four versions of the same design to the client.



Figure 1.1: Flowchart that steps through a general design process all teams went through prior to the design review.

1.3 Design Requirement Status

Each of the products individually serve a greater purpose to fulfill the design requirements our product partner provided. The greater design requirements are as follows; Produce a product that is durable, simple, and affordable, that reminds a patient with left neglect to scan to the left. With our original individual groups we were struggling to provide for the full design requirements but as a collective with four different roles we were able to effectively meet each major requirement. The individual components of the product include a pair of glasses, a feedback device, a clip, and an app. Each team has been assigned one component to work on for the remainder of the project but all the components work together in order to create one product that meets all the requirements. The Glasses team was able to meet the requirement of having a reminder to help the patients look left. The lights on the glasses serve as a constant visual reminder for those wearing the device.

This paired with the haptics provide a range of stimuli to remind the user as the haptics headband works to also fulfill the same design requirement. These two components together fulfill the most important part of the design requirements and provide a solid base for the rest of the components to build upon.

Next comes the gyroscope which will be taking the data from the two reminding components and relaying it to the final component, the app. This helps to keep the project simple and adaptable so the user can track their progress and determine if the product is helping with their left neglect. This feedback is crucial to determining the efficacy of the project. The app works in tandem with the gyroscope to take in the data from the patient and provide a simple and usable interface to interact with.

All of these components help together to assist the patient with scanning left and provides them with feedback on the progress from their time using the product (see Figure 1.2).




Figure 1.2: Integration diagram that shows how each component works together and which teams are involved in what parts.

Chapter 2

Users and Requirements

2.1 Main Users of the Design

2.1.1 Adults who are experiencing Left-Side Neglect

The users for our design will be patients at the Alexian Brothers Rehabilitation Center who have suffered a brain injury leaving them with inattentiveness to the left side of their body and surroundings. They need cues and reminders from a caretaker in order to remember to continue turning their head all the way to the left past the midline. They would like a solution that is easy to use and that reduces reliance on a caretaker in order to create more autonomy for the patient.

2.1.2 Physical therapists at Alexian Brothers Rehabilitation Center

The physical therapists work closely with the users to help them recover from their stroke and teach them techniques to mitigate the effects of left neglect.

2.2 Requirements

We identified the following requirements during the course of the quarter. The design had to satisfy each of them.

2.2.1 Lightweight

The device will be attached to a pair of glasses and therefore must be light enough to comfortably fit on the glasses and not put too much weight on the patient. It must be small enough to fit onto the glasses and not make them lopsided.

2.2.2 Durable

Since the device is attached to the glasses that will be worn frequently by the user, the device must be able to endure being worn daily as well as have a long battery life. The device must have some

protective covering so that the patient is able to wear the glasses outside and so that the device can withstand being dropped from a reasonable height.

2.2.3 Customizable and Adjustable

The degree of left neglect and the most effective stimuli can vary between users. Thus, the device should facilitate the customizability of the stimuli provided by the other components of the take home kit.

Chapter 3

Design Concept and Rationale

The *RotaSense* is a device that can track the degree to which the user turns their head and relay this data to the other components of the take-home kit. The *RotaSense* is intended to be mounted on the *Blinky* (the glasses), and sends data to *LeftAware* (the mobile app), the *Blinky* and the *TheraPulse* (the haptic clip). The design has an accelerometer gyroscope that is able to track how far and how quickly the user is turning their head. This data is processed by the ESP32 microcontroller, and then relayed to the other components using the communication protocols.

The following sections describe the components of the device — accelerometer/gyroscope, ESP32 microcontroller, and communication protocols — as well as the rationale for each component.

3.1 Accelerometer/Gyroscope

When creating this device, we needed some way of tracking movement and some way of processing the movement data and communicating it with the app. In order to track the users' movement, we decided to use a MPU6050 Six-Axis Gyroscope and Accelerometer. This device was chosen due to its simplicity, affordability, and library support.

The MPU6050 is a commonly used gyroscope/accelerometer. It features high resolution motion tracking with a built-in Digital Motion Processor to allow for high-resolution data capture ($\pm 2g$) with data-processing on chip [1]. This allows us to capture high-quality head position data with minimal power draw and computing resources on the microcontroller. The MPU6050 is also very small (smaller than a fingernail), lightweight (less than 5 g), and affordable (costing around \$3 for a development board) [1]. And most importantly, it is a very easy to use device with excellent library support through I2CDevLib [2] on the Arduino Framework [3], allowing us to quickly prototype, test, and iterate on our design.

3.2 ESP32 Microcontroller

For communication purposes, we decided to use the ESP32 microcontroller from Espressif Systems. This microcontroller is relatively small, with a length of 48 mm, a width of 25 mm, and a height of 11 mm. It supports Bluetooth 5.0 (Low Energy) and the I^2C communication protocol [4], which are needed for talking to the phone and MPU6050 (section 3.1), respectively. It has two cores




Figure 3.1: The MPU6050 Accelerometer and Gyroscope.

that can run at up to 240 MHz, a deep sleep mode for extremely low power consumption, and a large online support base to allow for easy development [4]. The chip is powerful enough to run a lightweight operating system, so we are able to use multiple processes to speed up computation and also save energy.



Figure 3.2: The ESP32 Microcontroller.

3.3 Communication Protocols

One of the most important components of our design are the communication protocols.

The I2C communication protocol was chosen due to its simplicity and unobtrusiveness, requiring only four wires for inter-device communication. It is a synchronous¹ serial² communication protocol, widely used for attaching low-speed peripheral devices (chips) to a controller chip for short-distance communication [5]. Used widely throughout the industry, it has great library and device support, with no downsides.

Bluetooth Low-Energy was chosen for its lower energy consumption, as our device should be able to run for long periods of time while on battery power. While Bluetooth Low-Energy is a more complex system to work with, introducing more moving “parts” and slightly slowing down our iteration speed, as power consumption is a main requirement of our design, we felt it was a fair trade-off.

In the same vein of saving energy, we also will try to use the Deep Sleep mode of the ESP32 to reduce current consumption during idle periods to only 5 μA [4].

¹All devices agree on when messages are sent.

²Only one message is sent at a time.

Chapter 4

Future Steps

There are four directions for future development with regards to *RotaSense* and *Neglect No More* as a whole: app, data processing, electronics, and angle detection.

4.1 App Improvements

The app can be improved in three ways. One, some of the user interface design can be refined to be more user friendly, as some parts of the design can be confusing, especially for users who are less tech savvy. This improvement is quite straightforward, and requires rethinking some components of the design.

Secondly, and much more importantly, the account system needs to be rethought completely. Given that this app is a healthcare app, having an account system would require us to process data in accordance with HIPAA laws, which themselves are bundles of red tape. Though, for the sake of patient privacy, these laws ought to be respected. Having an account system means there is an increased chance of patient data leakage. The account system needs to be redesigned to eliminate this possibility, as otherwise there is a lot of grounds for legal action.

Lastly, the app can be set up to connect with Bluetooth. Due to constraints with Apple development, integrating the prototype app with Bluetooth would require a \$100 Apple Developer License, which exceeds this project's budget. Instead, we created a Python proof of concept which works on a computer. While this is an okay substitute for now, it is not user-friendly as it requires extensive technical knowledge, making this a huge step for future development. Future development specifically can be done through using the user's phone's native Bluetooth library to interact with the microcontroller, which is a process that adds multiple layers¹ of programming on top of what is already complex for a prototype project.

4.2 Data Processing Improvements

The next direction for future development would be to improve the data processing. As of now, data is sent from the gyroscope microcontroller to the app by creating a ten-number packet (yaw, pitch, roll, acceleration in all three axes, gyroscope movement in all three axes, time). This can

¹CocoaPods and Swift for iOS, Kotlin for Android, etc.

be potentially improved by doing more data preprocessing on the microcontroller through applying things like the Fourier transform. More research is needed to decide the best form of data to transmit between controller and device as well as whether a change is necessary.

4.3 Electronic Components

The electronic components must be refined from the current prototypical form. The current microcontroller-gyroscope system takes up a lot of surface area, and this can be reduced. Reduction of surface area can be done by printing a custom circuit board, which is not overly expensive or complicated. This, however, would only be economically feasible if this product were to be mass produced.

4.4 Angle Determination Algorithm

The angle-determination algorithm can be refined to be more accurate. Currently, the user will calibrate a gyroscope by using their mobile device, and any subsequent movement will be detected as a movement. However, if the user were to change positions, the gyroscope would need to be re-calibrated altogether. A previous idea was to have two gyroscopes, one on the user's chest area and one on their head, and the difference in angle is the angle of head rotation. Further research needs to be done to determine which angle determination mechanism is the most effective and least intrusive for the patient.

4.5 Future Steps Conclusion

The overall next step is to revise the current design to better promote user autonomy while preserving user privacy. These four ways are not extensive, but are the next steps with regards to improving *RotaSense* and *Neglect No More*.

Chapter 5

Conclusion

To summarize, our design and its function in the larger class solution meet the key needs of the users. The key components of our design are as follows:

- An accelerometer/gyroscope sensor that is capable of tracking the head movements of the user.
- An ESP32 microcontroller that provides bluetooth capability, allowing the design to communicate information on the user's head movements to the mobile application and haptic clip.
- Communication protocols that allow for high speed and low energy communication with the other teams' components.

The users need a solution that is customizable and adjustable, and the feedback component provides this by collecting information on the user's movements so that the stimuli from the glasses and haptic clip can be modified to make sure the user is looking completely to their left. In addition, it was required that the solution be lightweight and ideally wearable, and the design meets this need because it is small enough to fit on a glasses strap and the prototype weighs less than 100g. Lastly, it was important that the solution be affordable, and our design is affordable, as the prototype can be built for a total cost of \$28.

Appendix A

Project Definition

A.1 Document Information

Project Name RotaSense

Client Dr. Kate Enzler, physical therapist at Alexian Rehab Hospital (in partnership with Shirley Ryan AbilityLab)

Team Members Maddie, Andrew, Nikola, Kathryn

Date 02/14/2023

Version 3

A.2 Project Information

A.2.1 Mission Statement

Design a simple, lightweight, and adjustable device for daily use by individuals experiencing left-neglect to draw focus to the user's left side, enhance user autonomy, and improve general quality of life.

A.2.2 Project Deliverables

The end result of our project will be

- A final prototype.
- A final report.
- A final presentation in the form of a poster.

A.2.3 Constraints

We have two main constraints on this DTC project, namely time (10 weeks) and budget (\$100).

A.3 User Information

A.3.1 Users and Stakeholders

The users and stakeholders for our project are

- People who are experiencing left side neglect.
- Therapists who work with patients on mitigating the effects of left side neglect.
- Family members/caregivers who are responsible for reminding the patients to look left at home.

A.3.2 Illustrative User Scenario

The user in the illustrative scenario below is based on an observation of a patient in therapy for left-neglect and supported by details from an interview with Dr. Kate Enzler.

The user begins their therapy session with Dr. Enzler at the Alexian Rehab Hospital. The patient has very recently had a stroke, and their eyes tend to look to the right side. When asked to perform the lighthouse scanning technique (a technique prompts someone to turn their head completely from right to left like a lighthouse), the user initially does not turn their head to the left, stopping instead at the midline. The user only continues scanning to the left when given a verbal cue by Dr. Enzler. There are people on the user's left, but the user only makes eye contact with the people in front and to the right of them. Dr. Enzler then puts up a finger on the user's left side, and the user is insistent that there is no finger there, assuming that Dr. Enzler is tricking them and believing that they have turned their head as far left as possible.

A.4 Project Requirements

Needs	Metrics	Units	Ideal Value	Allowable Value
Lightweight	Weight	g	Less than 25 g	Less than 50 g
	Max Size	cm	Size of an average pair of glasses	Average width of a human body
Durable	Lifespan	Years	Around 2 year	Around 1 years
	Battery Life	Months	More than 3 months	More than 1 month
	Element-resistant	N/A	Yes	Yes
Affordable	Price	Dollars	Less than \$50	Less than \$100

Appendix B

Background Research

At the start of our project, we conducted background research on terms, concepts, and conditions mentioned by our client, Dr. Kate Enzler from the Alexian Rehab Hospital, in her request for a design for something to alleviate left neglect. Left neglect is a neurological condition wherein patients have decreased response to their left sides, a condition our team has no background on. Our background research helped us understand left neglect, its severity, and current solutions to give us foundational knowledge needed for creating an effective design. This research was separated into three main sections: (1) defining left neglect, (2) understanding struggles patients with left neglect face, and (3) reviewing current approaches to left neglect.

B.1 Defining left neglect

Left neglect is a type of unilateral spatial neglect (USN), a syndrome that occurs after brain hemisphere damage [6]. Literature suggests that left neglect is more severe than right neglect [6], as well as 71% more frequent [7]. While left and right USN are both common after stroke [6], this higher frequency makes it more important to address left neglect due to its prevalence.

Literature also suggests that certain deficiencies are more severe for patients with left neglect. Peripersonal and extrapersonal neglect (neglect of one's personal space and neglect of one's immediate surroundings respectively) were found to be more frequent for patients with left neglect [6]. While right neglect also leads to deficiencies such as poorer balance [6], these left-sided deficiencies highlight the need for something to alleviate left neglect. Figure B.1 depicts the difference in brain response to stimuli for patients with no USN, right USN, and left USN.

B.2 Understanding struggles patients with left neglect face

Patients experiencing left neglect have a reduced ability to interact with the left side of their perceived world. Figure B.2, below, compares the interactions of patients with USN to patients without USN.

Figure B.3 further illustrates patients' reduced interaction with their left side. In Figure B.3, it's critical to note that patients don't entirely neglect one side. The clock in Figure B.3.C for example has a left half and only lacks numbers on the left. This indicates that, while patients are




Figure B.1: Diagram illustrating the effect of both left and right unilateral spatial neglect on neurological response to stimuli on both sides of the brain [7]. Here, right neglect is shown to eliminate left-side brain interactions while left neglect still has some interaction with both sides. This implies that right neglect is more severe, making it substantially more difficult to address than left neglect. This key difference also implies that a solution to left or right neglect is not a universal solution to unilateral spatial neglect; such a universal solution would require further research, planning, and resources.

less aware of their left side, they aren't entirely unaware, but rather need reminders to interact with this neglected side.

However, this still depicts reduced interaction. When doing less-trivial tasks, such as eating, this reduced interaction leads to immense struggle with common daily tasks such as only eating the right half of one's food, shown in Figure B.4.

B.2.1 Types of Left Neglect

The types of neglect faced by those with left USN can be broadly split into four groups of neglect: personal, peripersonal, extrapersonal, and temporal. The first three are defined by Ting et. al. and the last is defined by Saj et. al.

Personal Neglect

Personal neglect refers to neglect of one's personal space, which literature exemplifies with "combing, grooming, shaving, recognizing the right half of the body only" and "anosognosia", or when patients are unable to recognize their own disorder [9].

Peripersonal neglect

Peripersonal neglect is neglect of one's peripersonal space (space which can be immediately grasped) and includes "eating food from the right half of the plate and neglecting the food on the left" and "reading the right half of the two pages of an open book" [9]. Figure B.4 is an example of peripersonal neglect.

Extrapersonal neglect

Extrapersonal neglect is neglect of one's extrapersonal space (space beyond one's arm's reach) and includes failure to identify people on the left or collision with objects on the left [9].

Temporal neglect

Unlike the previous three categories, temporal neglect is a non-physical category of neglect. Research [11] describes that those with left spatial neglect are less able to perceive the past, which is the “left [side] [of] a mental timeline.” Figure B.5, below, plots the proportion of events patients were able to correctly associate with “past” or “future”; there is a substantial difference for patients with left spatial neglect who are more likely to associate past events as future and vice versa.

Research describes that this existence of temporal neglect implies that time has spatial properties [11]. More importantly, this type of neglect gives interesting insight when it comes to types of neglect faced by those with unilateral spatial neglect: not all types of neglect are physical, meaning even if a design idea could perfectly solve the physical aspects of neglect, it would not be 100% effective at solving all aspects of neglect.

B.3 Reviewing current approaches to left neglect

Current approaches to left neglect are extensive with multiple patented commercial products. Here, two specific therapies are explained in detail.

B.3.1 Specific Solution 1: Prism Adaptation

Prism adaptation treatment is a process in which individuals wear prisms that displace their vision in a specific direction while performing specific direction-related tasks [12]. While research indicates that individuals initially make errors, with repeated trials, the amount of errors decreases [12]. This same research further explains that for those with left neglect, right-shifting prisms shifts movement leftward [12] which directly tackles aforementioned issues regarding personal, peripersonal, and extrapersonal space. However, this solution is unlikely to alleviate temporal neglect due to time not being physical. An ideal solution would provide means to alleviate temporal neglect, though this necessarily complicates the design process by involving both physical and non-physical categories of neglect.

B.3.2 Specific Solution 2: “Visual Attention Therapy” App

Figure B.6 summarizes how the “Visual Attention Therapy” (VAT) App works. When an individual uses this app, they go through a timed trial in which they interact with certain elements spread across an entire tablet screen (A). At the end of the trial, individuals get results indicating where on the screen they were more likely to neglect (B). This information can be used to reduce the number of missed targets by giving individuals instant feedback.

B.3.3 Review of Therapies

Both of these therapies share one flaw: individuals cannot use them at any time, as they have to be in a comfortable, safe environment, meaning this solution is unable to help people when they may need it in public. The prism device is designed to only be worn for 20 or so minutes at a time [12], multiple times per week – literature indicates that at least twice per week prism device usage is needed for rehabilitative effects [12]. The app requires using an electronic device, constraining it to use only where and when an electronic device can be used.

B.4 Conclusion

Left neglect, a subset of unilateral spatial neglect, leads to deficiencies in interaction in affected individuals in multiple categories. These deficiencies decrease individuals' ability to fully interact with their world. Current solutions for left neglect are used in therapy; the two cited are both good examples, but are also restrictive in use cases. Neglect faced by affected individuals reduces quality of life, meaning a general solution for either left or right unilateral spatial neglect would be a critical tool in improving quality of life for affected individuals. However, as indicated in Figure 1, the relative severity of right-side USN makes designing an assistive device for left-side USN both more feasible given time constraints and necessary given its prevalence.




Figure B.2: (A) Behavioral inattention test with a washbasin; (B) healthy participant response; (C) left-neglect response [8].




Figure B.3: Diagram depicting patient interaction with their left and right sides. [9]



Figure B.4: Patients with left neglect may only eat the right half of their food [10].




Figure B.5: Temporal neglect of patients with left spatial neglect quantified by percentage of items correctly recalled.




Figure B.6: (A) Screenshot of VAT App from official website [13]; (B) zoomed in screenshot of test results page which indicates location of missed targets.

Appendix C

Interviews Summary

On Thursday, January 12th 2023, we had our initial client interview with Dr. Kate Enzler, a physical therapist at the Ascension Alexian Brothers Rehabilitation Center. The interview was conducted via Zoom at 8:15 am in the Ford Building. The purpose of this interview was to learn more about left side neglect and its effects on patients. It was also used to clarify what the goal of our project should be and how current solutions are not sufficient. This appendix summarizes what we learned about the design problem, requirements, users, and current solutions.

C.1 Problems

Our client explained the current challenges of patients with left neglect and clarified what type of patients we would be working with.

- Patients with left neglect have varying awareness of the world on the left side of their body. However, they do have all visual, auditory and physical abilities on the left side.
- The patients we will be working with have the ability to scan to the midline of their body.
- The patients often collide with objects or people on the left side of their body which is very dangerous as it can injure them.
- The current treatments are very dependent on caregivers or therapists as it requires verbal cues from the caregiver and for the caregiver to be present at all times. One goal of our solutions should be to create more independence for the patient after being taught to use the design.

C.2 Requirements

Our client identified these requirements for the design:

Adjustable The design should be able to change “sizes” in order to be able to accommodate many different patients’ body sizes. It will also have different time intervals that can be adjusted so that each patient can have stimulus at a rate that is set for their specific needs.

Affordable The design should cost less than \$100, making it affordable for all users. This is necessary so that the product can be used by all demographics of patients.

Lightweight The product must be lightweight so that all users, no matter their strength/phase of their recovery, should be able to use it.

Discrete The design should not limit the user in any way and should not make the user stand out in public. This way the user is comfortable with using the product and the product preserves their dignity.

Easy-to-Use Users of all ages should be able to use/navigate the device. The device should not contain technology. This way the user does not need to receive aid in order to use the product like they do using current therapies and devices (creating autonomy for the user).

C.3 Users

We also gathered the following general information about our users:

- Users of the product include any patients with left neglect past the midline of the body.
- Users are typically elderly or middle aged and might be overweight.
- Users stay at the rehabilitation center for an average of 3 weeks.

C.4 Current Solutions

Dr. Enzler described the current solutions available to help prompt left side awareness for patients with left neglect.

Visual Scanning Having a red line or ruler on the left side of the body or an object (e.g. book) in order to catch the patient's attention (prompted to find the red line). Dr. Enzler commented that this was relatively effective for specific activities (like reading a book) however not universally able to be used across multiple activities.

Eye Patching Placing an eye patch over the patient's right eye to encourage left side awareness (hit or miss with patients).

Prism Glasses Glasses that bring objects in the left field of vision to the midline.

“Lighthouse Technique” The patient is prompted by the caregiver/therapist to scan the surroundings like a lighthouse. Dr. Enzler uses this method most and described it as the most effective.

C.5 Conclusion

The interview provided critical information for understanding the problem, users, and client requirements. When we do our user observation at Shirley Ryan Rehabilitation Center, we will continue to learn more about left side neglect and ideate potential solutions.

Appendix D

Observations Summary

D.1 Introduction

Members of the DTC 1 section observed patients with left neglect at Alexian Brothers Hospital in Elk Grove Village on Wednesday, January 18, 2023. They also spoke with Dr. Kate Enzler to get further information about this condition. The purpose of this observation was to understand the symptoms and difficulties caused by left neglect, and how the severity can vary. The observation lasted about an hour. This appendix explains the methodology used to conduct the observation and what was learned about the condition of left neglect.

D.2 Methodology

The observation took place in a therapy room, where users typically receive physical therapy while recovering from a stroke or other brain injuries. The users were interviewed to get information about their experiences with left neglect and associated challenges. They were asked what therapies currently work best for them and which ones don't. The users were also asked for any requirements they would have for the device. Then, Dr. Enzler demonstrated some of the therapy methods she is currently using to treat left neglect.

D.3 Results

D.3.1 Cheryl

The first user is a female who had a stroke on December 19, 2022, which caused her left neglect and a loss of her left peripheral vision. Her main challenges are hygiene, reading, and therapy exercises.

Therapies used

Cheryl described the following therapies that she had tried. Her therapy was structured around her love of reading, which helped motivate her.

Brightly colored lines on the left side of a book Reminds users to scan to the left side of the page.

Lighthouse Technique Has users look from one shoulder to the other and back to scan their entire visual field.

Edge Technique Gets users to scan the full size of an object by looking for its edges.

Finger Tracing Strategy Has users scan the full shape of an object by following its edges.

Guideline Markers Marks areas to look at/focus on.

Large Print Books Just helps in general to make reading easier. These were partially used because of eyesight concerns.

Feelings about her care

Things that helped Cheryl throughout her care were the abundance of options to choose from (i.e., books, etc.) and here commitment to reading. The main difficulties for here were reoccurring feelings of stupidity, as previously easy tasks were now difficult, and annoyance from her visual field cut, which felt like a wall to her.

Observational notes

The following are some observations noticed while watching Cheryl perform therapy tasks with Dr. Enzler.

- Her left peripheral vision is completely gone.
- She preferred verbal reminders, especially verbal affirmations.
- Scanning was done both horizontally and vertically.

Product Requirements

Cheryl listed the following requirements for a product that she would use:

- High stress situations should be avoided at all costs. Otherwise, she wants to simply shut down.
- Screens (i.e., phones) can be difficult to use. The small text is already difficult to read for an elderly person. Additionally, processing the information-dense display causes difficulties.

D.3.2 Bryan

The second user is a male who had a stroke on December 24, 2022. This caused left neglect, but no change in vision. He is still in the early stages of recovery, and is still unsure of his main challenges. He also has baseline ADD (Attention Deficit Disorder) that has impacted him throughout his life.

Therapies used

Bryan, who was earlier in his recovery, had only tried a fraction of the techniques that Cheryl had used. The ones he did use are documented below.

Lighthouse technique Please see above for a description.

Self-reminders Bryan found it helpful to talk to himself and give himself reminders on what to do (i.e., “look left now”, etc.).

Observational Notes

The following are some observations noticed while watching Bryan perform therapy tasks with Dr. Enzler.

- He has a subconscious knowledge of Dr. Enzler’s location, gesturing directly towards her based on audio cues.
- An initial scan to left ends about 10 degrees left of center, but with repeated reminders he will turn his head over his shoulder.
- With cross-out puzzles (an exercise where you cross out all objects of one shape on the page), he was able to identify most items on a sparse page, but more dense pages led to more missed items.
- Physical stimuli on the left side had little effect.
- Works very methodically through therapy exercises.
- Eyes stray right, even when looking left.

Product Requirements

Bryan listed the following requirements for a product he would consider using:

- Keep it simple.
- Don’t overstimulate.
- Don’t overstimulate.
- Don’t make people feel stupid.
- Make it cheap.
- Have fun.

D.4 Conclusion

The challenges for each user are different, however, they all revolve around a lack of awareness left of center. Both users responded positively to affirming expressions, such as “You can do it!”, which may be good to implement in our product. They also both used the lighthouse technique and found it effective, meaning it may be a good starting point for our design. Both users also wanted a simple solution that wouldn’t be overstimulating, meaning it would be useful if our product was adjustable in terms of amount of stimulus. The users also want the device to be easy to use in order to preserve dignity, meaning it would be good if the device was discrete. Finally, it seems like it would be very helpful to make the device customizable to the user, as it was said that the most effective therapies were centered around doing things the user loved.



Figure D.1: Bryan and Cheryl, the two patients we observed and interviewed.

Appendix E

User Testing and Feedback Summary

E.1 Purpose

The purpose of our user testing was to determine which of our two mockups — the red line/transition lens glasses and the vibrating arm cuff — would provide a more effective stimulus for drawing the user's attention to their left side. In addition, we were looking for user insights into the comfort and usability of our designs.

E.2 Methodology

One team member conducted one user test session, at Alexian Brothers Rehabilitation Hospital under the supervision of Dr. Kate Enzler, with our mockups during the week of February 6th. The mockups were a pair of glasses with interchangeable velcro attachments to simulate different visual stimuli and an adjustable arm band containing a haptic motor. The mockups differed mainly in the stimuli being used to draw attention to the left side. Below are photos of the mockups used during the testing. Figure E.1 shows the glasses with various attachments and Figure E.2 shows the vibrating arm band.

During the user testing session, the user was asked to try on each mockup one at a time. A team member stood to the left side of the user and held up a number with their fingers, observing if the user turned their head towards the team member due to the stimuli from the mockup. After, the user was asked if they saw the number the team member was holding up. Then, the user was asked to provide feedback on how effective the design was at drawing attention to their left side and on how comfortable the design would be for daily wear.




Figure E.1: Glasses with various attachments (attachment 1: red felt line, attachment 2: opaque patch, attachment 3: transparent patch, attachment 4: opaque half patches).

E.3 Results

E.3.1 Glasses

When the user put on the glasses, they were able to see the red line, but it did not cause them to turn their head. They stated that it would take them a while to know whether it would work or not. When the line was moved closer to the eye, the user found it too close to their field of vision to be comfortable. However, it should be noted that another team tested how a pair of glasses of similar design improved the user's ability to read, and that design helped the user identify almost all of the letters on a page.

E.3.2 Arm Band

When the arm band mock-up was introduced, Dr. Enzler showed the team members present that the user had no feeling in their left arm by touching the arm and confirming with the user that they could not feel it. So, the mock-up was not used as intended; instead, it was placed around the user's head so that the vibration was near the user's ear. The user stated that the vibration was very noticeable, and it caused the user to turn their head far enough that they were able to accurately read the number that a team member was holding up.




Figure E.2: Arm-band haptic feedback mockup.

E.4 Analysis, Conclusions and Limitations

E.4.1 Analysis of Results

Glasses

The user did not find this mockup to be particularly effective at first use, stating that they would need about a month of wearing it to truly know if it was effective or not. It seems that because the visual cue would be moving with the user's head, it might not be as effective as a visual cue that remains stationary and serves as a basis for where the user should turn their attention to.

Arm Cuff

The original placement for this idea didn't work as intended because the user had very limited feeling on the left side of the body below the neck. But, when repositioning the mockup to be used as a headband so it vibrated around the user's ear, it seemed to be the one of the most effective stimuli of the testing session, as the user immediately turned their head and was able to see the number that the team member was holding up.

E.4.2 Conclusion

The results suggest that haptic stimuli around the head can be an effective method for drawing attention to the left side. The arm band design can be repurposed as a vibrating accessory worn around the head, and can be improved by incorporating the capability of tracking how far the user

has turned their head after receiving the stimuli. In addition, the results suggest that the glasses design was not effective in the use it was intended for, but results from other teams indicate that there are certain scenarios like reading where this pair of glasses might have value. One main takeaway from the testing session is that different designs can be effective in specific scenarios, so it may be worthwhile to explore how each team might make a design that plays a certain role in the larger whole of left-neglect rehabilitation.

E.4.3 Limitation

One of the largest limitations of this test session was the sample size of only one user. While the user provided important insights, it is also known that left neglect treatment is often customized and different solutions work best for different people, so it would have been helpful to test the mockups with more than one user. However, this limitation is one that is difficult to remediate due to the fact that Dr. Enzler only has a limited number of patients who would be able to participate in testing.

Another limitation to our testing methodology was that our design ideas incorporated electronics, but the mockups did not have independently working electronics. Since the functionality of our designs depends largely on these electronics, we were not able to thoroughly test the usability of designs. This limitation was especially prominent in the glasses mock-up, as ideally the visual stimuli would be turning on and off but since the attachments were static, it was hard to know if they would truly grab a user's attention or not.

Appendix F

Bill of Materials

Date	Description of Purchase	Qty	Vendor	Part Number	Cost
2/21	Breadboard	1	Amazon	B07PCJP9DY	\$6.69
2/21	ESP32 Microcontroller	1	Amazon	B0718T232Z	\$10.99
2/21	MPU6050 Accelerometer/Gyroscope	1	Amazon	B00LP25V1A	\$3.33
				TOTAL	\$21.01

Appendix G

Instructions for Construction

In order to create this product, you should have some level for familiarity with electronics (i.e., reading a schematic) and be comfortable using a computer from the command line/terminal. Please purchase all items from the Bill of Materials (Appendix F). You will also need some assorted 24 AWG wires, a wire cutter/stripper tool, and a knife.

The following instructions will lead you through the process of building the *RotaSense* and downloading the necessary code.

1. Install PlatformIO IDE following their official directions: <https://platformio.org/platformio-ide>.
2. Cut the sticky tape on the bottom of the breadboard in between the power rails and main body.



Figure G.1: Where to cut the sticky tape on the breadboard.

3. Slide the power rails off the main breadboard body.

Hint: one rail slides up, the other down.

4. Insert the ESP32 into the breadboard, with USB port facing outwards, and run wires as shown. A detailed schematic can be found in Figure G.4.

The red wire runs from *A12* to *F8*.

The black wire runs from *J18* to *G7*.

The blue wire runs from *J17* to *J5*.

The green wire runs from *J14* to *J6*.




Figure G.2: Wiring of the breadboard.

5. Insert the MPU6050 as shown below. *VCC* should match up with the red wire. The breadboard is now completed.




Figure G.3: The completed breadboard.

6. Download the project code from this link and unzip the file.

7. Open the downloaded directory with VSCode. PlatformIO should now automatically load.
8. Once in VSCode, press the checkmark on the bottom left-hand corner of the screen to build the project.
9. Plug in the ESP32 to your computer. If on Windows, it will show up under *Device Manager* → *Serial Devices*.

Note: you may have to download a Serial driver. See your ESP32 vendor for details.

10. Use the upload button (little arrow in bottom left) to upload the code to the ESP32.

Note: you may have to put the ESP32 in bootloader mode first (hold the IO0 button, press the EN button, and release IO0 once uploading commences).

11. Click the monitor button (the little plug) to check to make sure everything is working. There should be a line that looks like:

```
[....] [3] [src/main.cpp] MPU6050 setup completed successfully!
```




Figure G.4: Wiring Schematic.

Appendix H

Instructions for use

Here is how to use a theoretical, fully functional, production *Neglect No More* system.

1. Download the LeftAware app by scanning the QR code corresponding to your device.
[A diagram here will include those QR codes. However, for our project, we are unable to create these mobile apps due to (a) time constraints and (b) budgetary constraints.]
2. Create a user profile on the app.
3. Plug in the battery to the *RotaSense* (attached to the *Blinky*).
4. Press “Okay” when your phone prompts you to connect to the microcontroller.
5. Periodically check your phone to calibrate the microcontroller.
6. You’re all set!

References

- [1] *Mpu-6000 and mpu-6050 product specification*, MPU-6050, Revision 3.4, TDK Invensense, 2013. [Online]. Available: <https://invensense.tdk.com/wp-content/uploads/2015/02/MPU-6000-Datasheet1.pdf>.
- [2] J. Rowberg, *I2c device library*, original-date: 2011-07-26T04:58:39Z, Mar. 11, 2023. [Online]. Available: <https://github.com/jrowberg/i2cdevlib> (visited on 03/11/2023).
- [3] *Arduino core for the ESP32, ESP32-s2, ESP32-s3 and ESP32-c3*, original-date: 2016-10-06T06:04:20Z, Mar. 11, 2023. [Online]. Available: <https://github.com/espressif/arduino-esp32> (visited on 03/11/2023).
- [4] *Esp32 series datasheet*, ESP32-WROOM-32, Espressif, 2023. [Online]. Available: https://www.espressif.com/sites/default/files/documentation/esp32_datasheet_en.pdf.
- [5] SFUptownMaker, *I2c*, Sparkfun, 2013. [Online]. Available: <https://learn.sparkfun.com/tutorials/i2c>.
- [6] A. F. Ten Brink, J. H. Verwer, J. M. Biesbroek, J. M. A. Visser-Meily, and T. C. W. Nijboer, “Differences between left- and right-sided neglect revisited: A large cohort study across multiple domains,” *Journal of Clinical and Experimental Neuropsychology*, vol. 39, no. 7, pp. 707–723, Aug. 9, 2017, ISSN: 1380-3395, 1744-411X. DOI: 10.1080/13803395.2016.1262333. [Online]. Available: <https://www.tandfonline.com/doi/full/10.1080/13803395.2016.1262333> (visited on 01/09/2023).
- [7] “Half of your world is invisible: Hemispatial neglect,” SPORK! (Oct. 6, 2016), [Online]. Available: <https://www.sporkability.org/spork-exclusive/2016/10/4/half-of-your-world-is-invisible-hemispatial-neglect> (visited on 03/11/2023).
- [8] R. Gammeri, C. Iacono, R. Ricci, and A. Salatino, “Unilateral spatial neglect after stroke: Current insights,” *Neuropsychiatric Disease and Treatment*, vol. Volume 16, pp. 131–152, Jan. 2020, ISSN: 1178-2021. DOI: 10.2147/NDT.S171461. [Online]. Available: <https://www.dovepress.com/unilateral-spatial-neglect-after-stroke-current-insights-peer-reviewed-article-NDT> (visited on 03/11/2023).
- [9] D. S. Ting, A. Pollock, G. N. Dutton, *et al.*, “Visual neglect following stroke: Current concepts and future focus,” *Survey of Ophthalmology*, vol. 56, no. 2, pp. 114–134, Mar. 2011, ISSN: 00396257. DOI: 10.1016/j.survophthal.2010.08.001. [Online]. Available: <https://linkinghub.elsevier.com/retrieve/pii/S0039625710001505> (visited on 01/09/2023).
- [10] M. S. Sutton. “Left neglect after stroke - definition & treatment exercises.” (), [Online]. Available: <https://tactustherapy.com/what-is-left-neglect/#> (visited on 03/11/2023).

- [11] A. Saj, O. Fuhrman, P. Vuilleumier, and L. Boroditsky, “Patients with left spatial neglect also neglect the “left side” of time,” *Psychological Science*, vol. 25, no. 1, pp. 207–214, Jan. 2014, ISSN: 0956-7976, 1467-9280. DOI: 10.1177/0956797612475222. [Online]. Available: <http://journals.sagepub.com/doi/10.1177/0956797612475222> (visited on 01/09/2023).
- [12] K. M. Goedert, J. Y. Zhang, and A. M. Barrett, “Prism adaptation and spatial neglect: The need for dose-finding studies,” *Frontiers in Human Neuroscience*, vol. 9, Apr. 30, 2015, ISSN: 1662-5161. DOI: 10.3389/fnhum.2015.00243. [Online]. Available: <http://journal.frontiersin.org/article/10.3389/fnhum.2015.00243/abstract> (visited on 03/11/2023).
- [13] TactusTherapy. “Visual attention therapy app for left neglect on iOS & android - cognition.” (), [Online]. Available: <https://tactustherapy.com/app/vat/#> (visited on 01/09/2023).