

Acquisition Method: HeadspaceSagebrush_2013_bypass_frinlet

Path: C:\CDSProjects\HeadspaceGC\Methods

1 Method Information

Last Saved As:	C:\CDSProjects\HeadspaceGC\Methods\HeadspaceSagebrush_2013_bypass_frinlet.amx	Modified:	2023-08-10 10:59:16-06:00	Modifier:	SYSTEM (SYSTEM)
Created:	2023-06-26 12:47:24-06:00	Creator:	SYSTEM (SYSTEM)	Description:	
Version:	2023-0626-1847-24814	Method Status:	Generic		

2 GC

Module Display Name:	Agilent 6890N	Module Type:	GC	Order:	1
----------------------	---------------	--------------	----	--------	---

2.1 GC Summary

Run Time:	25.667 min	Post Run Time:	2 min
-----------	------------	----------------	-------

2.2 Oven

Equilibration Time:	0.05 min	Max Temperature:	320 °C	Maximum Temperature Override:	Disabled
Slow Fan:	Disabled				

2.2.1 Temperature

Setpoint:	On	(Initial):	40 °C	Hold Time:	2 min
Post Run:	300 °C				

Program

Rate	Value	Hold Time
3 °C/min	60 °C	0 min
5 °C/min	120 °C	0 min
20 °C/min	220 °C	0 min

2.3 ALS

ALS Errors:	Pause for user interaction
-------------	----------------------------

2.3.1 Front Injector

Syringe Size:	10 µL	Injection Volume:	2 µL	Injection Repetitions:	1
Injection Delay:	0 sec	Solvent A Washes (PreInj):	3	Solvent A Washes (PostInj):	3
Solvent A Volume:	8 µL	Solvent B Washes (PreInj):	3	Solvent B Washes (PostInj):	3
Solvent B Volume:	8 µL	Sample Washes:	1	Sample Wash Volume:	8 µL
Sample Pumps:	5	Dwell Time (PreInj):	0 min	Dwell Time (PostInj):	0 min
Solvent Wash Draw Speed:	300 µL/min	Solvent Wash Dispense Speed:	6000 µL/min	Sample Wash Draw Speed:	300 µL/min
Sample Wash Dispense Speed:	6000 µL/min	Injection Dispense Speed:	6000 µL/min	Viscosity Delay:	0 sec
Sample Depth:	Disabled	Solvent Wash Mode:	A, B	Waste Bottle Use:	A Only

2.4 Front SS Inlet He

Mode:	Split	Heater:	Off	Pressure:	Off
Total Flow:	Off	Gas Saver:	Off	Split Ratio:	0.1 :1
Split Flow:	0.2992 mL/min				

2.5 Back SS Inlet He

Mode:	Split	Heater:	On 110 °C	Pressure:	On 11.601 psi
Total Flow:	On 9.8989 mL/min	Gas Saver:	Off	Split Ratio:	6 :1
Split Flow:	6.0227 mL/min				

2.6 Column

2.6.1 Column #1

Temperature Range:	0 °C—320 °C (320 °C)	Dimensions:	30 m x 530 µm x 3 µm (Uncalibrated)	Heater:	Oven
In:	Front SS Inlet He	Out:	Front Detector ECD	Column Outlet Pressure:	0 psi
(Initial):	40 °C	Pressure:	2.299 psi	Flow:	2.992 mL/min
Average Velocity:	22.486 cm/sec	Holdup Time:	2.2236 min	Control Mode:	Constant Flow

2.6.1.1 Flow

Setpoint:	Off	(Initial):	2.992 mL/min	Post Run:	0 mL/min
-----------	-----	------------	--------------	-----------	----------

2.6.2 Column #2

Temperature Range:	0 °C—320 °C (320 °C)	Dimensions:	30 m x 250 µm x 0.25 µm (Uncalibrated)	Heater:	Oven
In:	Back SS Inlet He	Out:	Back Detector FID	Column Outlet Pressure:	0 psi
(Initial):	40 °C	Pressure:	11.601 psi	Flow:	1.0038 mL/min
Average Velocity:	25.099 cm/sec	Holdup Time:	1.9921 min	Control Mode:	Constant Pressure

2.6.2.1 Pressure

Setpoint:	On	(Initial):	11.601 psi	Post Run:	32.999 psi
-----------	----	------------	------------	-----------	------------

2.7 Front Detector µECD

Makeup:	ArMe	Heater:	Off	Makeup Flow:	Off
Carrier Gas Flow Correction:	Constant Makeup Flow	Electrometer:	Off		

2.8 Back Detector FID

Makeup:	N2	Heater:	On 220 °C	H2 Flow:	On 45 mL/min
Air Flow:	On 400 mL/min	Makeup Flow:	On 25 mL/min	Carrier Gas Flow Correction:	Constant Makeup and Fuel Flow
Flame:	On				

2.9 Aux EPC

2.9.1 Aux EPC 3 He

Aux EPC 3 He:	(unused)
---------------	----------

2.9.1.1 Pressure

Setpoint:	Off	(Initial):	19.999 psi
-----------	-----	------------	------------

2.9.2 Aux EPC 4 He

Aux EPC 4 He:	(unused)
---------------	----------

2.9.2.1 Pressure

Setpoint:	Off	(Initial):	0 psi
-----------	-----	------------	-------

2.9.3 Aux EPC 5 He

Aux EPC 5 He:	(unused)
---------------	----------

2.9.3.1 Pressure

Setpoint:	Off	(Initial):	0 psi
-----------	-----	------------	-------

2.10 Signals

2.10.1 Signal #1: Back Signal

Description:	Back Signal	Details:	Back Signal (FID)	Save:	On
Data Rate:	20 Hz	Dual Injection Assignment:	Front Sample		

2.10.2 Signal #2: Back Signal

Description:	Back Signal	Details:	Back Signal (FID)	Save:	Off
Data Rate:	20 Hz	Dual Injection Assignment:	Back Sample		

3 Agilent 7697A

Module Display Name:	7697A	Module Type:	Headspace Sampler	Order:	1
----------------------	-------	--------------	-------------------	--------	---

3.1 Temperature Settings:

Oven Temperature:	100 °C	Loop Temperature:	110 °C	Transfer Line Temperature:	120 °C
-------------------	--------	-------------------	--------	----------------------------	--------

3.2 Timing Settings:

Vial Equilibration:	20.00 min	Injection Duration:	0.50 min	GC Cycle Time:	35.00 min
---------------------	-----------	---------------------	----------	----------------	-----------

3.3 Vial and Loop Settings:

Vial Size:	20	Vial Shaking:	Level 1, 18 shakes/min with acceleration of 60 cm/s ²	Fill Mode:	Default
Fill Flow:	50	Fill Pressure:	15 psi	Pressure Equilibration Time:	0.1
Loop Size:	1 mL	Loop Fill Mode:	Custom	Loop Ramp Rate:	20 psi/min
Loop Final Pressure:	10 psi	Loop Equilibration Time:	0.05		

3.4 Carrier Settings:

Carrier Control Mode:	GC controls Carrier
-----------------------	---------------------

3.5 Advanced Settings:

Extraction Mode:	Single Extraction	Vent After Extraction:	ON	Post Injection Purge:	Default, 100 mL/min for 1 min
Acceptable Leak Check:	Default, 0.2mL/min				

3.6 Sequence Actions:

Vial Missing:	Skip	Wrong Vial Size:	Continue	Leak Detected:	Continue
System Not Ready:	Continue				

4 Method Properties

Instrument Technique:	Gas Chromatography
-----------------------	--------------------

5 Schema version

Schema version:	2.3
-----------------	-----

Method Audit Trail

Method audit trail is not printed