EE 472 Lab 3
Learning the Development Environment - The Next Step

Jonathan Ellington
Patrick Ma
Jarrett Gaddy

Contents

List of Tables

List of Figures

1 ABSTRACT

In this lab the students are to take on the role of an embedded system design team. They will design modifications to the medical instrument previously designed. When the device finds metrics are out of the acceptable range, the user is notified, thus saving them from potential health risks. The students first laid out the design for their system using various design tools, then they implemented the system in software. Finally the students tested their system and verified that it is ready to start saving lives.

2 INTRODUCTION

The students are to design an embedded system on the Texas Instruments Stellaris EKI-LM3S8962 and EE 472 embedded design testboard. The design must implement a medical monitoring device. This device must monitor a patient's temperature, heart rate, and blood pressure, as well as its own battery state. The design must indicate when a monitored value is outside of a specified range by flashing an LED on the test board. When a value deviates even further from the valid range an alarm will sound. This alarm will sound until the values return to the valid range or the user acknowledges the alarm with a button. The values of each measurement will also be printed to the oled screen. This implementation will build upon the previous implementation of the device by adding functionality. Added functions include heart rate sensor, keypad input, menu display, data logging, and UART serial communication.

The design will be tested to verify proper behavior on alarm and warning notifications. In addition the implementation will be tested by measuring the amount of time that each of the 8 program tasks running the instrument take to execute. These tasks are mini programs that each handle a part of the instruments purpose.

3 DISCUSSION OF THE LAB

3.1 Design Specification

3.1.1 Specification Overview

The entire system must satisfy several lofty objectives. The final product must be portable, lightweight, and Internet-enabled. The system must also make measurements of vital bodily functions, perform simple computations, provide data logging functionality, and indicate when measured vitals exceed given ranges, or the user fails to comply with a prescribed logging regimen.

The initial Phase 1 functional requirements for the system are:

- Provide continuous sensor monitoring capability
- Produce a visual display of the sensor values
- Accept variety of input data types
- Provide visual indication of warning states
- Provide an audible indicator of alarm states

In addition, the following requirements have been added:

- Utilize a hardware-based time reference

- support dynamic task creation and deletion

- support a user input device

- support data logging capabilities

- support remote communication capability

- Improve overall system performace

- Improve overall system safety

3.1.2   Identified Use Cases

Taking the functional requirements listed above, several use cases were developed. A Use case diagram of these scenarios is given in Figure 1. Each use case is expanded and explained below.

Figure 1: Use case diagram

Use Case #1: View Vital Measurements
In the first use case, the user views the basic measurements picked up by the sensors connected to the device.
During normal operation, once the device is turned on by the user, the system records the value output by each sensor. This raw value is linearized and converted into a human-readable form. The user can select toggle between a summary of current vitals as measured by the system or view measurements of each sensor individually.
    Three exceptional conditions were identified for this use case:

- *One or more of the expected sensors is not connected* - If this occurs, the measurements taken by the device may be erratic. At the present moment, no action will be taken in such events. Later revisions may address the issue

- *A measured value is outside 5% of the specified normal range* - In this case, a warning signal will flash as an indication of the warning condition

- *A measured value falls outside 10% of a specified "normal" range* - In this case, an audible alarm will sound to indicate the alarm condition

Use Case #2: Acknowledge Alarm

2

In the second case, the system is in an alarm state. The user acknowledges the alarm condition by pressing a button.

Upon pressing the button, the system silences the audible alarm. Any visual warnings continue to flash during the silenced period. If a significant amounts of time pass and the sensor reading(s) continue to maintain an alarmed state, the audible alarm will recommence.

No exceptional conditions were identified for this use case.

Use Case #3: View Measurement Logs

In the third use case, the user wishes to view previously recorded measurements for a given vital sign.

As the device is running, the user opts to enter View Logs mode. From this point, the user can select which vital sign they wish to examine. Upon selection, the data logged from previous measurements is displayed.

Possible exception conditions may include the following:

- *User wishes to view more data than the machine remembers* - The machine will not support this operation. The system is only able to display the amount of data defined by the device. Future variants may support an external storage or logging functionality

- *Machine loses power while reading or writing* - if the system loses power, any data stored in dynamic memory will be lost. On restart, data will be overwritten and lost. Future device versions may allow for storage in nonvolatile memory

- *Ongoing measurements trigger warning or alarms* - since measurements are taken continuously, the device may enter an alarm or warning state. In this case, the display will not change unless prompted to by the user. Any alert indicators (visual, audible, or remote messaging) will operate as normal in the background

Use Case #4: System Alert to Remote Terminal

In the event that the system enters an alert state (e.g. alarm or warning), the system can send a message to a remote terminal connected to the device. The messages sent can inform a second actor of the cause of alert and provide any additional useful information.

Possible exception conditions may include:

- *Improper configuration of the Remote Terminal* - If the remote terminal connection is improperly configured or initialized, data received may be corrupted and not display properly. The device cannot ensure a proper connection and it is the responsibility of the remote user to ensure the correct configuration is used

- *Remote Connection Lost* – If the user terminates the connection or the connection is lost, messages sent by the device may not arrive at the remote terminal or data may be corrupted. The device will not necessarily monitor the status of any remote connection; this responsibility is the remote terminals. In the event a connection is broken, the device system must continue to perform the other system functions without ill effects.

3.1.3 Detailed Specifications (update)

For this project, the requirements have been further specified as follows:

The system must have the following inputs:

- Alarm acknowledgment capability using a pushbutton
- Buttons or switches to allow user to access system modes and menu items
- Sensor measurement input capability consisting of:
  - ∗ Body temperature measurement
  - ∗ Pulse rate measurement signal
  - ∗ Systolic blood pressure measurement
  - ∗ Diastolic blood pressure measurement

The system must have the following outputs:

- Visual display of the following data in human-readable formats:
  - ∗ Body temperature
  - ∗ Pulse rate
  - ∗ Systolic blood pressure
  - ∗ Diastolic blood pressure
  - ∗ Battery status
- Visually indicate warning state with a flashing LED
- Visually indicate a low battery state with an LED
- Audibly indicate an alarm state using a speaker
- External data connection to a remote terminal

The initialization values, normal measurement ranges, displayed units, and warning and alarm behaviors for each vital measurement are given in Table 1. The sensors must be sampled every five seconds and the system cannot block and cease operation for five seconds.

Measurement	Units	Initial Value	Min. Value	Max. Value	Warning Flash Period
Body Temperature	C	75	36.1C	37.8C	1 sec
Systolic BP	mm Hg	80	-	120 mmHg	0.5 sec
Diastolic BP	mm Hg	80	-	80mmHg	0.5 sec
Pulse Rate	BPM	50	60 BPM	100 BPM	2 sec
Remaining Battery	%	200	40 %	-	Constant

Table 1: Specifications for measurement data

A measurement enters a warning state when its value falls outside the stated normal range by 5%.

requirement modification: An alarm state occurs when the systolic blood pressure falls outside its stated normal range by 20%.

Additionally, the system must be implemented using the Stellaris EKI-LM3S8962 ARM Cortex-M3 microcomputer board, The software for the system must be written in C using the IAR Systems Embedded Workbench/Assembler IDE.

3.1.4 Detailed Task Specifications (updates)

- New task: KeypadTask

  - The keypad task will scan the keypad and decode any keypresses
  - The task will have support the following user inputs:
  - Mode selection between 2 modes (1 button)
  - Menu selection between 3 options (1 button)
  - Alarm acknowledgement (1 button)
  - Up and down scroll functionality (1-2 buttons)
  - A new set of global variable will be created to store the state of the keypad and key presses

- New Task: Initialize (Startup):

  - Changes to Warn/AlarmTask:
  - The warnings will be activated and indicated as before in project 1
  - The alarm state is changed to activate only when the systolic pressure is 20% above the normal range.
  - The alarm will sound in 1 second tones (1 second on, 1 second off)
  - When an alarm or warning state occurs, the serial communication task will be added to the task queue
  - The deactivation period of the alarm sound is defined as 5 measurement periods

- New task: Serial Communication:

  - The task is enabled by the warn/alarm task
  - When run, the task will open an RS-232 connection at XXXX baud, XXinsertdetails hereXXX
  - The present corrected measurement will be displayed on the terminal in the same fashion as the display task annunciation mode
  - After sending data to the terminal, the serial communication task will remove itself from the task queue

- Changes to the MeasureTask:

  - Once a complete set of measurements has been taken, the compute task is added to the task queue
  - Pointers to the variables used in the measure task will be relocated to accommodate the new data architecture
  - The pulse measurement will monitor and count the frequency of a pulse rate event interrupt
  - A new value will be stored to memory if the present reading is greater than $\pm15\%$ of the previous measurement

- The measurement limits will correspond to 200bpm and 10bpm, determined empirically.

- Changes to ComputeTask:

  - All measurements will be recomputed
  - After computing the corrected values for all measurements, the ComputeTask will remove itself from the task queue

- Changes to DisplayTask:

  - Display will now support multiple display options
  - Menu mode will allow selection of each of the individual measurements. Upon selection of a measurement, the current value of the measurement will be displayed onscreen
  - Annunciation mode will display the current status of each measurement as in project 1, and provide the same functionality as the display in project 1.

- Changes to Warn/AlarmTask:

  - The warnings will be activated and indicated as before in project 1
  - The alarm state is changed to activate only when the systolic pressure is 20% above the normal range.
  - The alarm will sound in 1 second tones (1 second on, 1 second off)
  - When an alarm or warning state occurs, the serial communication task will be added to the task queue
  - The deactivation period of the alarm sound is defined as 5 measurement periods

- Changes to Schedule:

  - System will maintain a list of activated and deactivated tasks. The list must be updatable during runtime based on the system state
  - The hardware timer will provide a system interrupt every 250ms or equal to the minor cycle, whichever is shorter
  - At runtime, upon a timer interrupt event, all tasks will be added or removed according to the task activation list. The tasks will then be run
  - Task Control Blocks will have forward and backward pointers to allow references to the next task
  - The scheduler cannot block for five seconds
  - The scheduler will toggle a GPIO pin at least once during execution of the task queue

- Changes to StatusTask:

  - There are no changes to the status task

3.2 Software Implementation

A top-down design approach was used to develop the system. First, a functional decomposition of the problem was carried out based on the identified use cases. Next, the system architecture was developed. After understanding the system architecture, the high-level project file structure in C was defined, followed by the low-level implementation of the tasks.

3.2.1 Functional Decomposition

After understanding how the user would interact with the device, the high level functional blocks were developed. These blocks are shown in Figure 2.


Figure 2: Functional Decomposition

3.2.2 System Architecture

Next, the system architecture was developed (Figure 3). At a high level the system works on two main concepts, the scheduler and tasks. Tasks embody some sort of work being done, and the scheduler is in charge of determining the speed and order in which the tasks execute. The system has several tasks, each with their own specific job. For modularity reasons, each task should have the same public interface and the scheduler should be able to run each task regardless of that specific tasks job or implementation. Thus the task concept is abstracted into a Task Control Block (TCB), and the scheduler maintains a queue of TCBs to run. The TCB abstraction is shown in Figure 3 using inheritance, and the fact that the scheduler has a queue of TCBs is shown with composition. The core functionality of the system was divided into the following eight main tasks:

Initialization Task   This task Initializes data structures and does system startup-related jobs. This task is not actually scheduled to run, it only executes a single time at system startup.

Measure Task   The measurement task is in charge of interacting with the blood pressure, temperature, and pulse sensors. Each of these is simulated. The blood pressure and temperature are simulated in the CPU. The task will measure the pulse rate by parsing an externally generated square wave of varying frequency; the pulse rate being proportional to the frequency.

7

Compute Task   Compute takes the simulated raw data and converts it to the correct units of measurement. Raw temperature data to Celsius, blood pressure to mm Hg, and pulse rate to BPM.

Keypad Monitor   Keypad will check the keypad for user input. It should provide the user with four keys: two for scrolling, one for selection, and one to go back.

Display Task   The display task will show a user interface on the Stellaris OLED. The user will interact with the display using a keypad. Under normal operation, a menu will be displayed asking users which measurement they would like to see. If the user presses back while in this menu, they enter annunciate mode which displays all the measurements currently in warning or alarm state.

Warning/Alarm Task   Under normal operation, this task will light a green LED signifying that everything is OK. If one of the measurements enters a warning state, the task will flash a red LED at a rate specific to the warning for that measurement. If there is an alarm state, it sound the alarm by driving the speaker. At any time if the battery goes too low, the yellow LED will illuminate.

Remote Communication Task   This task is in charge of sending data to a remote terminal. If any of the states are in a warning or alarm condition, this task will transmit the (corrected) data to the remote terminal.

Status Task   Status receives information about the battery on the system and updates it's current data accordingly.

Each of these tasks interact using the shared data shown in Figure 3.

3.2.3   High-level Implementation in C

After developing the system architecture, the design needed to be translated into the C programming language. The design manifested in a multi-file program consisting of the following source files:

- **globals.c/globals.h** - Used to define the Shared Data used among the tasks

- **schedule.c/schedule.h** - Defines the scheduler interface and it's implementation, as well as the TCB structure

- **timebase.h** - Defines the timebase used for the scheduler and tasks

- **task.h** - Defines the TCB interface for a task

Each task also has it's corresponding ".c" and ".h" file (for example, measure.c and measure.h).

The TCB structure that the scheduler uses must work for all tasks, and must not contain any task-specific information. Instead, the TCB consists of only a void pointer to the tasks data, and a pointer to a function that returns void and takes a void pointer, as follows:

Figure 3: System Architecture Diagram

```
1  struct TCB {
2    void *taskDataPtr;
3    void (*taskRunFn)(void *);
4  }
5
```

Leaving out the type information allows the scheduler to pass the task's data (*taskDataPtr) into the task's run function completely unaware of the kind of data the task uses or how the task works.

For increased modularity, the data structure used by each task was not put in the task's header file. Instead, the structure was declared within the task implementation file, and instantiated using a task initialization function. In the header file, a void pointer pointing to the initialized structure is exposed with global scope, as well as the task's run and initialization functions.

3.2.4 Task Implementation

The primary task of this project is to implement C code for a medical device on the Stellaris EKI-LM3S8962 and its ARM Coretex A3 processor. The project was started by creating a main file that initializes the variables used in each task and starts the hardware timer then runs into an infinite while loop. Inside the while loop a run method is called. The run method is part of the scheduler. The run method has a runTask flag which determines whether or not anything should actually be run this call. The flag is set to true by the hardware timer's interrupt handler. Once the runTask flag is true the run method will keep track of whether the device is on a minor cycle or a major cycle and run the preform task method of each task. The runTask flag is then set to false so that the tasks will not be executed again until the hardware interrupt has again

been triggered. The tasks included in this project are Compute, Measure, Warning, keyPad, OLEDDisplay, Serial, and status. Each task has a public interface of 2 void pointers. One that when initialized by the main method will point to the preform task function, and another that, when initialized, points to a struct containing pointers to the data required by that task. Each task has a task control block(TCB) in the scheduler. This TCB contains pointers to the preform task function and the data for the task and also has fields for pointers to a next TCB and previous TCB. The TCB is used by the scheduler to run the task. The scheduler contains a doubly linked list of TCB objects that uses the TCB next and previous elements to point to the next and previous tasks in the task queue. In this case there are 7 tasks but not always 7 tasks in the list of tasks to run. The compute task is only to be run after the measure task, and the serial task only needs to be run at certain times. When a task is not being used it's TCB is not included in the linked list of tasks. Therefore and updateQueue method was created. This method checks flags set within the tasks that are running and determines if a task that isnt in the list needs to be added, or if a task that is in the list needs to be removed. The scheduler's run task contains a loop that runs through the linked list of TCBs and runs the function pointed to by the TCB with the argument of the data pointer stored in the TCB until the null value pointed to by the last element in the list is reached. After running all tasks the runTask flag is set to false again. The hardware timer will count up until it reaches the number of ms that corresponds to a minor cycle then trigger a hardware interrupt. The interrupt handler sets the runTask flag back to true and the task linked list will be updated, traversed and run again.

control flow is shown in Figure 4.

Each task has its own unique purpose in the system, and each uses a different part of the global data.

Measure Task The Measure task (Figure 5 in the appendix), deals only with the raw data from the instruments. This task is meant to act in place of the instruments that are unavailable. The task only runs if the scheduler has set the global value is Major Cycle to 1. On a major cycle the measure function either increments the data of each measurement by 1 or 2 or decrements the data by 1 or 2. In the newly improved design a heart rate monitor has been added. The heart rate monitor uses an interrupt handler to count the number of rising edges on an input in a 2 Major cycle period. The number is then used to calculate the heart rate the sensor is receiving. Additionally the measure task adds the compute task to the task queue after it has run.

Compute Task The compute task is very simple. This task will only run after it has been placed into the task queue by the measure task and it will remove itself from the queue after running. It takes the raw data that has been set by the measure task, multiplies by a constant and adds a constant to each piece of raw data to get the corrected data. The compute task then puts the values for the corrected data in memory at the location of the global data pointer. Compute uses every data value except the battery and keypad data. A diagram of the Compute activity is shown in the appendix in Figure 6.

Warning Task After compute, the warning task begins checking for warning or alarm states. The warning task only deals with the corrected data from compute and the battery state. This task also must deal with the input and output signals used to display warning and sound an alarm. Unlike the other tasks, this task has more to its initialization than just initializing the data. In addition to setting up the pointers to the global data, during initialization, the task also enables peripheral banks C, F, and G. These are enabled using the SysCtlPeripheralEnable

Figure 4: Control Flow Diagram

library call. Additionally the task set up pins C 5, 6, and 7 as outputs, and pins F 0 and G 1 as PWM outputs. Additionally the PWM outputs are set to use a 65 Hz clock to play a sound at this frequency whenever enabled. The activity diagram is shown in Figure 8 in the appendix. There are 3 subsystems in the warning task. These subsystems each handle a different part of user notification. The first subsystem deals with the alarm. The subsystem checks to see if the systolic blood pressure data is out of range by more than 20%. If the value is out of that range then the PWM output is enabled using PWMGenEnable and pulsed with a 2 second period. If the values fall back within the acceptable range, or the user hits the acknowledge button, then the PWM is disabled with PWMGenDisable and the sound stops. The acknowledge button is sensed in the keyPad task and a global data value global.alarmAcknowledge is set. When the alarmAcknowledge field is set to 1 the alarm will go to its resting state. The next subsystem checks the corrected data for being 5% out of range of the accepted values. If any value is more than 5% out of its range then a warning will be displayed on the red led connected to pin C 5 using the GPIOPinWrite function. Depending on the value that is out of range the period the led flashes at will vary. In addition to the led flashing the warning will also tell the scheduler

to add the serialCommunication task to the task list. The final subsystem is the battery check. This system checks if there is more than 30% of battery left on the device. This is taken from the battery state data field. If there is less than 20% battery left then a yellow led connected to pin C7 is illuminated, if there is more then 20% battery, and the device is not in a warning or alarm state then the green led on pin C 6 is illuminated.

Keypad Task   The keyPad task uses the GPIO libraries to set up 5 inputs. 1 input is pin E 0 and is used as the alarm acknowledge button. The other 4 pins are inputs on pins D 4, D 5, D 6, and D 7, and are used to detect button presses on the external keypad. The keypad works by connecting 2 of the outputs from the keypad together. There are 4 row outputs and 4 column outputs for a total of 16 possible combinations. The design was simplified by only using 1 column of 4 buttons. By doing this, the column keyPad wire can be supplied with a constant 5V from the Stellaris board, then each of the 4 row lines can be connected to a GPIO input. when a button in the live column is pushed the 5V column line is connected to 1 of the GPIO pins which lets the keyPad task know which button was pushed. Every time the keyPad task is run the GPIO will tell the task what buttons are pressed. 2 of the buttons correspond to up and down for scroll in the menu, 1 button corresponds to select in the menu, and 1 button changes the mode between menu and annunciation. The data that the buttons manipulate are the global keyPad data parts which are global.select, global.scroll, and global.mode. Additionaly the input on pin E 0 manipulates the global.alarmAcknowledge signal. The activity diagram for this task is shown in the appendix in Figure 7.

Display Task   To show a user their current medical measurements, the system also has an oledDisplay task. This task uses the corrected data from measurements, the battery state, and the keyPad data. The display task has an activity diagram shown in Figure 10 in the appendix. This task uses the usnprintf() function in C to convert the data types that the corrected data is stored in, and properly format these data values into a string which is stored in a buffer. The string contained in the buffer is then printed to the OLED screen using the driver library rit128x96x4 functions. The display has 2 modes. The mode that the screen currently displays is determined by the global.mode data which is set in the keyPad task. When mode is 0 this is the menu mode. This mode displays each of the 4 measurement types, temperature, blood pressure, pulse rate, and battery without the data for each. The task then takes the global.scroll data from keyPad and displays a cursor next to the measurement that scroll is currently at, 0 corresponding to temperature, 1 to blood pressure, and so on. If the global.select is set to 1 in keyPad then the currently scrolled to measurement is selected and the screen will change to showing only that measurement and its data. When global.mode is equal to 0 the annunciation screen will instead be displayed. This screen shows each measurement followed by its data.

Serial Task   The serial task is only run when a warning occurs. The warning causes the TCB for the serial task to be added to the TCB schedule linked list. When the serial task is run the first time it will initialize a UART connection using UARTConfigSet and UARTEnable driver functions to enable the UART that communicates to an FTDI chip on the Stellaris board. The FTDI chip then converts the UART to a virtual serial port over the USB cable to the PC. After the UART is initialized on the first run and in all subsequent runs of the serial task, all the measurements and their data are formatted and printed into 1 buffer using the usnprintf function. A loop then iterates through the buffer writing each character in the buffer to the UART. An activity diagram showing the serial task is located in the appendix as Figure 9.

Status Task The last task is the status task. This task only deals with one piece of data which is the battery state data. The only thing the status task does is that on a major cycle, it decrements the battery state by 1. This is shown in the activity diagram in Figure 11 in the appendix.

4 PRESENTATION, DISCUSSION, AND ANALYSIS OF THE RESULTS

4.1 Results

Patrick

The project was completed and demonstrated on February 13, 2014.

Demonstration of the system to the interested parties showed that the system met the majority of the requirements initially presented at the onset of the lab project. Testing of the system prior to demonstration also verified that the system met the specifications listed in Section 3.1.

During the demonstration, all tasks worked as designed and expected with the exception of

During the actual coding and implementation of the design, remarks were made several times about the ease of execution during that phase of the project. After the initial high level Design phase, very few changes were made, or required, to the functional design or system architecture.

Using an oscilloscope, the run times of each task were empirically determined. The results are given in Table 2.

Task	Runtime (μs)
Measure	23.4
Compute	55.4
Display	22900.0
Warning	27.4
Status	5.6

Table 2: Empirically determined task runtimes

Answers to the last three questions in the list of items to include in the project report:
You don't find the stealth submarine. That's why they are so expensive; at that cost, you take great pains to never lose one.
A helium balloon always rises. It just rises upside-down.
If you really managed to lose the stealth submersible, you first have to tell the government, which will deny it has any stealth submersibles, then you have to comb the seven seas until your comb hits the sub.

4.2 Discussion of Results

The ease of change in the code is the result of a large amount of time spent on design. The design makes it easy to configure flash times, add new tasks, and to reason about tasks independently of the whole system. The solid high-level architectural design led to ease of implementation and change.

In terms of performance, the run times of each task appear to correspond with the number of instructions required for each task. Given the speed of the CPU, 8 MHz, we can calculate an estimated number of instructions for each task. This is given in Table 3. The majority of the cycles are likely spent waiting for memory. For example, the status task only has two

Task	Instructions
Measure	187
Compute	443
Display	183,200
Warning	219
Status	45

Table 3: Estimated instructions per task, rounded to the nearest instruction

comparisons and an arithmetic operation, but has to reference the data in global memory. The exception here is the display task, which was about three orders of magnitude more instructions than the other tasks. This was due to the sprintf() library call, included in the standard C library. While this could have been optimized, it was found that with a minor cycle delay of 250 ms, the display delay of 22.9 ms was not significant.

4.3   Analysis of Any Errors

- BPM Range error - Serial error

There were two errors found in the final project. First, the pulse rate range is not exactly as specified. Second, it was found that the serial communications task did not produce the correct values.

The specified corrected pulse rate was to be between 10 BPM and 200 BPM. For simplicity, the raw pulse rate was implemented as a 1-1 mapping from frequency to raw pulse rate. For example, a 1 Hz frequency would produce a value of 1 for raw pulse rate, and a 15 Hz signal would produce a value of 15 for raw pulse rate. This caused two issues. When the Input frequency is 1, the measured BPM (using the specified raw to correct conversion) is $8+3\cdot(1)=11BPM$. This is larger than the specified 10 BPM minimum. Also, when the input frequency is 64 Hz, a corrected value of $8+3\cdot(64)=200BPM$ is expected. However, as the frequency increased, the overhead of the other running tasks became significant. As a result, more rising edges could fit in our measurement interval than we expected. This resulted in a maximum BPM of roughly 206 BPM.

There was also an error in the Serial communication protocol in which the usnprintf command used to format the data was causing a runtime error when printing formatted data to a buffer. Rather than printing the actual corrected values to the buffer (to be sent to the remote terminal), it seemed the command was printing addresses. It was believed this was caused by an error with the CircularBuffer implementation, though no other tasks had issues with it (and extensive tests validated the CircularBuffer).

4.4   Analysis of Implementation Issues and Workarounds

The medical instrument design in this project was completed and tested successfully to meet almost all the requirements, the designers did face a few difficulties in designing the device, however, because this design was additional functionality added to a previous projects design, many of the errors previously encountered were easily avoided due to experience of the students, and already completed coding work.

Many of the challenges the designers of this project faced were in the keypad input and the data output. The keypad input posed a difficult hardware challenge as there are 16 input keys on the keypad but only 8 connections for the microprocessor to connect to the keypad. This

means that to identify a single key press 4 connections must be set as outputs and 4 as inputs. The inputs can then be scanned as the outputs are set 1 at a time to find which key is pressed. Instead of implementing this design, the students instead opted to use only 4 buttons on the keypad. This allowed the strobe design to be ignored. Instead 1 row of keys was activated all the time and that row was scanned for button presses.

In addition to keypad input, there was also difficulty in implementing data formatting functions. After adding a hardware delay, IAR workbench no longer allows the use of sprintf which had previously been used to print and format data. The usnprintf command was instead used to format and print data to a buffer, however, the students found that usnprintf does not have the ability to print floating point data. This issue was resolved by changing measurements that were previously printed as floats to be printed as integers. Usnprintf also caused issues when printing certain data for the serial task. In this case the usnprintf was causing a runtime error and freezing the operation of our device. This issue remained unresolved.

All problems but one were solved before demonstrating the product to the interested parties. The final project still contained the serial error previously mentioned.

5 TEST PLAN

To ensure that this project meets the specifications listed in section 3.1, the following parts of the system must be tested:

- Vitals are measured and updated

- System properly displays corrected measurements and units properly

- System enters, indicates, and exits the proper warning state for blood pressure, temperature, pulse, and battery

- System enters and exits the alarm state correctly

- Alarm is silenced upon button push

- Alarm recommences sound after silencing if system remains in alarm state longer than silence period

Additional tests to determine the runtime of each specific task are also required.

The inclusion of additional specifications for Phase II of the project requires additional tests to ensure the system meets the customer requirements.

Phase II Tests:

- Scheduler loops through linked list properly

- Scheduler adds and removes from the linked list the following tasks correctly:

  - Compute task added by measure task

  - Serial communication task added by annunciation task

  - Compute task removed by itself

  - Serial communication task removed by itself

- Warning task alarm meets the following two requirements

1. Has one (1) second tones; a total period of 2 seconds

2. Activates only when systoic pressure is greater than 20% above normal

3. Has an auditory deactivation or "sleep" period of 5 measurement cycles

- Serial task displays the temperature, blood pressure, pulse rate, and battery status as listed in Section 3.1

- Keypad task captures user inputs, sets the appropriate inputs, and causes the assocated changes in system state

- Hardware timer updates the system's mnor cycle counter

More detailed explanation of the tests performed is provided in the following sections.

5.1 Test Specification

5.1.1 Scheduler

The scheduler needs to be shown to correctly schedule and dispatch tasks. This means that task should execute in the right order, and at the right time. Given a minor cycle of 50 ms, every task should run roughly once every 50 ms. Also, the scheduler needs to successfully add and remove tasks from the queue dynamically. Specifically, the Measure Task should be able to tell the scheduler to add the Compute Task and the Warning/Alarm Task should be able to schedule the Serial Task. Both Compute and Serial should be able to be removed from the schedule.

5.1.2 Measure Task

For this design, the temperature and blood pressure values were simulated on the CPU. The pulse rate was simulated using an externally generated square wave of varying frequency.

- **Temperature** The temperature should increase by two every even major cycle (5 seconds) and decrease by one ever odd major cycle until it exceeds 50, at which point the process should reverse (decrease by two every even major cycle and increase by one every odd major cycle), until it dips below 15, and the whole process should be started over again.

- **Pulse** The pulse rate should match one-to-one with the frequency of the input signal. For example, a 15 Hz signal should produce a raw pulse rate of 15.

- **Systolic Pressure** The systolic pressure should increase by three every even major cycle and decreases by one every odd major cycle. If it exceeds 100, it should reset to an initial value.

- **Diastolic Pressure** The diastolic pressure should decrease by two on even major cycles and decrease by one on odd major cycles, until it drops below 40, when it should restart the process.

The Measurement Task should also successfully add the Compute Task to the schedule queue.

5.1.3 Compute Task

The compute task should be verified to convert raw simulated sensor data according to the following formulas.

- $CorrectedTemperature = 5 + 0.75 * RawTemperature$

- $CorrectedSystolicPressure = 9 + 2 * RawSystolicPressure$

- $CorrectedDiastolicPressure = 6 + 1.5 * RawTemperature$

- $CorrectedPulseRate = 8 + 3 * RawTemperature$

The compute task should also successfully remove itself from the schedule queue.

5.1.4 Keypad Task

The keypad should be tested to successfully capture user input. When the select button is pressed, the measurement selection value should reflect the selected measurement. When the up scroll button is pressed, the scroll value should be incremented, and when the down scroll button is pressed, the scroll value should be decremented. If the alarm acknowledge button is pressed, this should be reflected in the alarmAcknowledge global value.

5.1.5 Display Task

On load, the display task should present the user with an option to select the desired measurement. If the back button is pressed, the annunciation screen should be displayed, showing the measurements in warning or alarm state.

5.1.6 Warning/Alarm Task

The warning/alarm system needs to be tested to do several things. When in a warning state, it should flash the red LED at the rate appropriate for the warning. When the battery is low, it should illuminate the yellow LED. If the system is in an alarm state, it should sound the speaker alarm. The following ranges in Table 4 are calculated from the specified minimum and maximums found in Table 1 on page 4.

Data	Warning Range	Alarm Range
Temperature	34.3 - 39.7 C	32.5 - 41.6 C
Systolic Pressure	> 84 mmHg	> 88 mmHg
Diastolic Pressure	> 126 mmHg	> 132 mmHg
Pulse	57 - 63 BPM	54 - 110 BPM

Table 4: Initial values and warning/alarm states

This task should also add the Serial task if any of the measurements are in a warning or alarm condition.

5.1.7 Serial Task

If any of the measurements are in warning or alarm state, this task should send this data serially
to a remote terminal. The task should send all the data (not just the data in warning or alarm
state). It should be printed in the following format:

```
1  1. Temperature              0 C
2  2. Systolic Blood Pressure  0 mm Hg
3  3. Diastolic Blood Pressure 0 mm Hg
4  4. Pulse Rate               0 BPM
5  5. Battery                  0 %
```

5.1.8 Status Task

Since the initial design does not use a battery, the status task simulates the battery state using
the CPU. For now, it simply decrements the state of the battery. The test should show that the
battery state is decremented by one every major cycle.

5.2 Test Cases

The students begin testing by examining if the alarm sounds at the proper time. This is initially
tested by disabling the functions that simulate measurements being made on each of the data
measurements, and setting their initial values to be either within the alarm range or outside of
the alarm range. The warning states were also initially tested this way. The initial values for raw
data given in Table 1 on page 4 were used to test the normal state of the machine because each
falls within the acceptable range of measurements for corrected data (also given in Table 1)
that does not require a warning. Using these initial values, the code was programmed onto the
Stellaris board. Correct operation was verified by the alarm not sounding, and the red led being
off, indicating that no warning state was in effect. In addition the green led was on indicating a
normal state. Next the students varied one parameter at a time to be outside of the acceptable
range by more than 10%. Starting with the temperature being set to an initial raw value of 50,
the alarm was verified by hearing the aural annunciation coming from the system. In addition,
the temperature warning stat was also in effect. This means that the green led was off and the red
led was blinking. To verify correct operation we needed to make sure the led was blinking with
a period of 1 second. The correct flashing pattern was verified by counting the number of times
the led flashed in 6 seconds. In this case, for temperature, the led flashed 6 times in 6 seconds
indicating a 1 second period, and correct operation. After this test, the temperature value was
returned to 42 and the Pulse was instead set to 45. The same methods were used to verify that
the alarm and warning states for pulse rate were working correctly, but this time the warning
led turned on 3 times in 6 seconds indicating a 2 second period which is the intended period
of flashing. The pulse rate was then returned to 25 and each pressure reading was checked for
correct operation individually by being set to an initial raw value of 100. Once again, the green
led started off because the system was not in a normal state. The alarm was sounding due to
the extremely high blood pressure measurements, and the red warning led flashed 12 times in
6 seconds indicating the correct period of .5 seconds for a blood pressure warning. In addition
to testing the validity of each warning state and alarm state, the acknowledgment of the alarm
was also tested during each of these tests. This was tested by hitting the acknowledge button

18

once during each measurements test. During each test, hitting the acknowledge button turned the alarm sound off for a short time, as intended.

Next the measurement simulation functions were tested. This was done by re-enabling each one that had been disabled from the previous test one at a time. The initial raw values were again set to the values in Figure 5. When each measurement was re-enabled, the students could watch the temperature change at each major cycle using the OLED display. Since the OLED display indicated that the corrected temperature went up .75 degrees on a major cycle then down 1.5 degrees on the next, the temperature measurement was working as intended. This situation also gave the students an opportunity to verify that the warning and alarm states initiated as the temperature fell out of the acceptable range. The Led began flashing with a 1 second period after a few major cycles, then the alarm began sounding, indicating correct operation. Since temperature was working correctly, the temperature measurement function was once again disabled and each blood pressure measurement was re-enabled individually for testing. The Systolic pressure began by rising 4 mm Hg on a major cycle then falling 2 mm Hg on the next, and the Diastolic pressure by rising 3 on a Major cycle and falling 1.5 on the next, this was consistent with the design. The warning and alarm states were activated as each passed its threshold and the red led was blinking with a period of .5 seconds. The warning led was also tested in the case that all warning states were active. To do this all initial values were set to 100. In this case, as designed by the students, the red warning led indicated the fastest blinking warning with a .5 second period.

In this device, a new pulse rate monitor device was added. To test the pulse rate monitor the monitor input was connected to a function generator generating square waves. As the square wave frequency was increased the pulse rate value was expected to increase. This was verified to be working correctly. As the pulse rate passed through the warning zones the design for pulse rate warnings was also verified to be working correctly.

Additionally, the improved medical device now has a menu that is displayed on the OLED display and navigated using the testbench keypad. The design operating these functions was tested by navigating to each part of the menu using the keys on the keypad and visually verifying that each part of the menu displayed the correct data. Each menu, annunciation, main menu, and each measurements selection mode, was visually verified to be working as intended. The keypad functionality was verified as each button was used to navigate through the menu.

The newly added serial communication was then tested using the hyperterm program on the lab station PC. The serial connection was established over a virtual COM port on the USB connection from the PC to the Stellaris board. The program had the intended functionality of displaying each of the measurements, and its current data on the serial port whenever the alarm state was entered. The functionality could be verified by watching the hyperterm screen to see if the data displayed when a warning state occurred. The data on hyperterm could then be compared to the OLED display data to verify its accuracy.

The final bit of testing preformed on the system was timing each task within the system. This was done by adding a general purpose output pin in the scheduler code. This output was set high right before the execution of a task, and set low immediately after the execution of the task. An oscilloscope was then attached to this output pin and set to trigger on a positive edge. The cursors were then used to measure the amount of time the signal was high in each cycle.

6   SUMMARY AND CONCLUSION

6.1   Final Summary

A medical monitoring system with a user interface, alarm notification, and remote terminal display was designed, implemented, and tested. The design simulated temperature and blood pressure, and obtained pulse rate data from an external function generator. These results were converted to a human readable form, and tested to see if there was a warning or alarm state. In the event of warning or alarm, the user was notified visually with LEDs and aurally with an alarm sound. The warning and alarm data was transferred to a remote terminal.

Some implementation errors were encountered. The pulse rate range could not fit the specification exactly, and the serial communications task was not printing the correct values to the remote terminal. Aside from these two implementation errors, the device worked as specified.

6.2   Project Conclusions

This project contained 3 major phases, the design, implementation, and testing steps. The students were immediately introduced to using the unified modeling language(UML) to design embedded systems. This is the first time many students will have used UML for system design which caused some confusion and difficulty. In the end through the use of the UML guidelines for design, the students were able to implement their system in code for the Texas Instruments Stellaris EKI-LM3S8962 much more quickly and with far fewer errors than if they had spent less time in the design phase of this project.

Effective design tools allowed the students to quickly implement their embedded system in C code for an ARM Cortex A3 processor, and move onto the testing phase of the project quickly. Unfortunately, while testing the students encountered a number of problems in using the PWM and general purpose input and output signals. After consulting the documentation for the Stellaris kit and solving their input/output problems, they began testing their design using visual and audio queues, the IAR embedded workbench debugger, and a few specifically programmed debug features. After the results of the testing verified the design to be working correctly, the students proceeded to present their medical instrument to their instructor.

A   BREAKDOWN OF LAB PERSON-HOURS (ESTIMATED)

Person	Design Hrs	Code Hrs	Test/Debug Hrs	Documentation Hrs
Patrick	20	10	10	12+
Jonathan	15	5	4	8

By initializing/signing above, I attest that I did in fact work the estimated number of hours stated. I also attest, under penalty of shame, that the work produced during the lab and contained herein is actually my own (as far as I know to be true). If special considerations or dispensations are due others or myself, I have indicated them below.

B ACTIVITY DIAGRAMS

Figure 5: Measure Activity Diagram

Figure 6: Compute Activity Diagram

Figure 7: Keypad Activity Diagram


Figure 8: Warning Activity Diagram

Figure 9: Serial Activity Diagram

Figure 10: Display Activity Diagram

Figure 11: Status Activity Diagram

C SOURCE CODE

Source code for this project is provided below.

C.1 Main Function

../code/main.c

```
1  #include "schedule.h"
2  #include "timebase.h"
3  #include "startup.h"
4  #include "globals.h"
5
6  #include "inc/hw_types.h"
7  #include "driverlib/debug.h"
8  #include "inc/hw_memmap.h"
9  #include "driverlib/systick.h" // for systick interrupt (minor cycle)
10 #include "driverlib/sysctl.h"
11 #include "driverlib/interrupt.h"
12
13
14 #define NUM_TASKS 5
15 #define BOARD_CLK_RATE   8000000 // 8MHz
16
17 #ifdef DEBUG
18   void
19 __error__(char *pcFilename, unsigned long ulLine)
20 {
21 }
22 #endif
23
24
25 int main(void) {
26   // Set the clocking to run directly from the crystal.
27   SysCtlClockSet(SYSCTL_SYSDIV_1 | SYSCTL_USE_OSC | SYSCTL_OSC_MAIN |
28       SYSCTL_XTAL_8MHZ);
29
30   startup();    // initialize system state and variables
31
32   while (1) {
33       runTasks(); // from schedule.h
34   }
35 }
```

C.2 Global Data

../code/globals.h

```
1  /*
2   * globals.h
3   * Author(s): Jonathan Ellington, Patrick Ma
4   * 1/28/2014
```

27

```c
 *
 * Defines global data for tasks to access
 * MUST be initialized before using
 */

#include "inc/hw_types.h"
#include "CircularBuffer.h"
#include "stddef.h"
#include "inc/hw_memmap.h"
#include "driverlib/gpio.h"
#include "driverlib/sysctl.h"

#define DEBUG 0

#define TEMP_RAW_INIT 80          // initial 80
#define SYS_RAW_INIT 80          // initial 50
#define DIA_RAW_INIT 50          // initial 50
#define PULSE_RAW_INIT 30          // initial 30

#define TEMP_CORR_INIT 0.0
#define SYS_CORR_INIT 0.0
#define DIA_CORR_INIT 0.0
#define PULSE_CORR_INIT 0.0

#define BATT_INIT 200

typedef struct global_data {
   CircularBuffer temperatureRaw;
   CircularBuffer systolicPressRaw;
   CircularBuffer diastolicPressRaw;
   CircularBuffer pulseRateRaw;

   CircularBuffer temperatureCorrected;
   CircularBuffer systolicPressCorrected;
   CircularBuffer diastolicPressCorrected;
   CircularBuffer pulseRateCorrected;

   unsigned short batteryState;
   unsigned short mode;
   unsigned short measurementSelection;
   tBoolean alarmAcknowledge;
   tBoolean select;
   unsigned short scroll;
} GlobalData;

extern GlobalData global;

// initializes the global variables for use by system
void initializeGlobalData();

// allows use of pin47 for debug, toggles the pin hi or lo alternatively
void debugPin47();
```

```
1  /*
2   * globals.c
3   * Author(s): Jonathan Ellington, Patrick Ma
4   * 1/28/2014
5   *
6   * Defines global data for tasks to access
7   */
8  #include "globals.h"
9
10 GlobalData global;
11
12 // The arrays to be wrapped in a
13 // circular buffer
14 static int temperatureRawArr[8];
15 static int systolicPressRawArr[8];
16 static int diastolicPressRawArr[8];
17 static int pulseRateRawArr[8];
18
19 static float temperatureCorrectedArr[8];
20 static float systolicPressCorrectedArr[8];
21 static float diastolicPressCorrectedArr[8];
22 static float pulseRateCorrectedArr[8];
23
24 void initializeGlobalData() {
25   // Wrap the arrays
26   global.temperatureRaw = cbWrap(temperatureRawArr, sizeof(int), 8);
27   global.systolicPressRaw = cbWrap(systolicPressRawArr, sizeof(int), 8);
28   global.diastolicPressRaw = cbWrap(diastolicPressRawArr, sizeof(int), 8);
29   global.pulseRateRaw = cbWrap(pulseRateRawArr, sizeof(int), 8);
30
31   global.temperatureCorrected = cbWrap(temperatureCorrectedArr, sizeof(float
       ), 8);
32   global.systolicPressCorrected = cbWrap(systolicPressCorrectedArr, sizeof(
       float), 8);
33   global.diastolicPressCorrected = cbWrap(diastolicPressCorrectedArr, sizeof
       (float), 8);
34   global.pulseRateCorrected = cbWrap(pulseRateCorrectedArr, sizeof(float),
       8);
35
36   int tr = TEMP_RAW_INIT;
37   int sr = SYS_RAW_INIT;
38   int dr = DIA_RAW_INIT;
39   int pr = PULSE_RAW_INIT;
40
41   float tc = TEMP_CORR_INIT;
42   float sc = SYS_CORR_INIT;
43   float dc = DIA_CORR_INIT;
44   float pc = PULSE_CORR_INIT;
45
46   // Add initial values
47   cbAdd(&(global.temperatureRaw), &tr);
```

```
48    cbAdd(&(global.systolicPressRaw), &sr);
49    cbAdd(&(global.diastolicPressRaw), &dr);
50    cbAdd(&(global.pulseRateRaw), &pr);
51
52    cbAdd(&(global.temperatureCorrected), &tc);
53    cbAdd(&(global.systolicPressCorrected), &sc);
54    cbAdd(&(global.diastolicPressCorrected), &dc);
55    cbAdd(&(global.pulseRateCorrected), &pc);
56
57    // Set normal variables
58    global.batteryState = 200;
59    global.mode = 0;
60    global.measurementSelection = 0;
61    global.alarmAcknowledge = false;
62    global.select = false;
63    global.scroll = 0;
64  }
65
66  // debug tool
67  void debugPin47() {
68    SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOB);          // debug
69    GPIOPinTypeGPIOOutput(GPIO_PORTB_BASE, GPIO_PIN_2);
70    long a = GPIOPinRead(GPIO_PORTB_BASE, GPIO_PIN_2);
71    if(0 == a)
72      GPIOPinWrite(GPIO_PORTB_BASE, GPIO_PIN_2, 0xFF);
73    else
74      GPIOPinWrite(GPIO_PORTB_BASE, GPIO_PIN_2, 0x00);
75  }
```

C.3   Timebase

../code/timebase.h

```
1  /*
2   * timebase.h
3   * Author(s): Jonathan Ellington
4   * 1/28/2014
5   *
6   * Defines the major and minor cycles the system runs on
7   */
8
9  #define MINOR_CYCLE 250       // minor cycle, in milliseconds
10 #define MAJOR_CYCLE 4         // major cycle, in number of minor cycles
11 #define PULSE_CYCLE 2         // pulse rate measurement time, in number of
12                              // major cycles.
13 #define IS_PULSE_CYCLE (minor_cycle_ctr % (MAJOR_CYCLE * PULSE_CYCLE) == 0)
14 #define IS_MAJOR_CYCLE (minor_cycle_ctr % MAJOR_CYCLE == 0)
15
16 extern unsigned int minor_cycle_ctr;    // counts number of minor cycles
```

C.4   Scheduler

```
1  /*
2   * schedule.h
3   * Author(s): Jonathan Ellington
4   * 1/28/2014
5   *
6   * Defines the scheduler interface.  The scheduler
7   * is responsible for running tasks on a specified
8   * schedule.
9   */
10
11 #include "inc\hw_types.h"
12
13 // globally setable variables
14 extern tBoolean computeActive;
15 extern tBoolean serialActive;
16
17
18 /*
19  * Must be called before runTasks()
20  * Initializes schedule required data structures
21  */
22 void initializeQueue();
23
24
25 /* Run all the tasks in the queue, delay minor cycle */
26 void runTasks();
```

```
1  /*
2   * schedule.c
3   * Author(s): Jonathan Ellington
4   * 1/28/2014
5   *
6   * Implements schedule.h
7   */
8
9  #include "task.h"
10 #include "schedule.h"
11 #include "timebase.h"
12 #include "globals.h"
13
14 // Each task include
15 #include "measure.h"
16 #include "compute.h"
17 #include "display.h"
18 #include "warning.h"
19 #include "status.h"
20 #include "serial.h"
21 #include "keyPad.h"
22
23 // Used for debug display
```

```
24  #if DEBUG
25  #include "drivers/rit128x96x4.h"
26  #include "utils/ustdlib.h"
27  #endif
28
29
30  unsigned int minor_cycle_ctr = 0;   // minor cycle counter
31
32  static TCB *currentTaskPtr;  // taskQueue pointers
33  static TCB *listHeadPtr;
34  static TCB *listTailPtr;
35
36  // flags to track task states
37  tBoolean computeActive;
38  tBoolean serialActive;
39  static tBoolean computeInQueue;
40  static tBoolean serialInQueue;
41
42  tBoolean runSchedule = false;
43
44
45  // Private functions
46  void delay_in_ms(int ms);
47  void insertNextNode(TCB *newNode);
48  void deleteCurrent();
49  void updateQueue();
50  void debug();
51
52  // Must initialize before running this function!
53  void runTasks() {
54    if(runSchedule) {
55    runSchedule = false;
56
57    updateQueue();
58
59    while (NULL != currentTaskPtr) {
60      currentTaskPtr->runTaskFunction(currentTaskPtr->taskDataPtr);
61      currentTaskPtr = currentTaskPtr->nextTCB; // go to next task
62    }
63    }
64  }
65
66
67  // Initialize the taskQueue with each task
68  void initializeQueue() {
69    // listhead > measure > keyPad > display > warn > status > null
70    listHeadPtr = &measureTask; // from measure.h
71    measureTask.nextTCB = &keyPadTask;
72    keyPadTask.nextTCB = &displayTask;
73    displayTask.nextTCB = &warningTask;
74    warningTask.nextTCB = &statusTask;
75    statusTask.nextTCB = NULL;
76
```

```
77     // backwards pointers listTailPtr > status > ...
78     listTailPtr = &statusTask;
79     statusTask.prevTCB = &warningTask;
80     warningTask.prevTCB = &displayTask;
81     displayTask.prevTCB = &keyPadTask;
82     keyPadTask.prevTCB = &measureTask;
83     measureTask.prevTCB = NULL;
84
85     currentTaskPtr = listHeadPtr; // and set up to start at the top
86   }
87
88   /* Traverse the taskQueue and insert/delete tasks based on set flags.
89    * currentTaskPtr pointer is reset to the head of the list */
90   void updateQueue() {
91     // update computeTask
92     if (computeActive && !computeInQueue) {
93       currentTaskPtr = &measureTask;
94       insertNextNode(&computeTask);
95       computeInQueue = true;
96     }
97     if (!computeActive && computeInQueue) {
98       currentTaskPtr = &computeTask;
99       deleteCurrent();
100      computeInQueue = false;
101    }
102    // update serialTask
103    if (serialActive && !serialInQueue) {
104      currentTaskPtr = &warningTask;
105      insertNextNode(&serialTask);
106      serialInQueue = true;
107    }
108    if (!serialActive && serialInQueue) {
109      currentTaskPtr = &serialTask;
110      deleteCurrent();
111      serialInQueue = false;
112    }
113    currentTaskPtr = listHeadPtr;
114  }
115
116  // inserts the given node into the list as the next node. currentTaskPtr
         pointer
117  // moves to point at the newly inserted node.
118  // Source: code derived in part from JD Olsen, Innovative Softwear and TA,
         Ltd
119  void insertNextNode(TCB *newNode) {
120    if (NULL == listHeadPtr) { // empty list
121      listHeadPtr = newNode;
122      listTailPtr = newNode;
123    }
124    else {  // insert the newNode between two nodes
125      newNode -> nextTCB = currentTaskPtr -> nextTCB;
126      currentTaskPtr -> nextTCB = newNode;
127      newNode -> prevTCB = currentTaskPtr;
```

```
128      if (NULL == (newNode -> nextTCB)) { // newest node is at list end
129        listTailPtr = newNode;
130      } else {   // in the middle somewhere
131        newNode -> nextTCB -> prevTCB = newNode;
132      }
133    }
134    currentTaskPtr = newNode;
135 }
136
137 /* Removes the currentTaskPtr node from the taskQueue.
138  * Moves currentTaskPtr pointer to the nextTCB task in the list. */
139 void deleteCurrent() {
140   if (NULL == currentTaskPtr -> nextTCB) {   // edge case: at last task
141     listTailPtr = currentTaskPtr -> prevTCB;
142     listTailPtr -> nextTCB = NULL;
143   } else {   // reassign pointers
144     currentTaskPtr -> nextTCB -> prevTCB = currentTaskPtr ->prevTCB;
145     currentTaskPtr -> prevTCB -> nextTCB = currentTaskPtr -> nextTCB;
146   }
147   currentTaskPtr = currentTaskPtr -> nextTCB; // update currentTaskPtr
148      pointer
149 }
150
151 // Software delay
152 void delay_in_ms(int ms) {
153   for (volatile int i = 0; i < ms; i++)
154     for (volatile int j = 0; j < 800; j++);
155 }
156
157 // for debug (obviously)
158 void debug() {
159   char num[30];
160     static int test = 0;
161     usnprintf(num, 30, "Test number: %d  ", test);
162     RIT128x96x4StringDraw(num, 0, 90, 15);
163     test++;
164 }
```

```
128      if (NULL == (newNode -> nextTCB)) { // newest node is at list end
129        listTailPtr = newNode;
130      } else {   // in the middle somewhere
131        newNode -> nextTCB -> prevTCB = newNode;
132      }
133    }
134    currentTaskPtr = newNode;
135 }
136
137 /* Removes the currentTaskPtr node from the taskQueue.
138  * Moves currentTaskPtr pointer to the nextTCB task in the list. */
139 void deleteCurrent() {
140   if (NULL == currentTaskPtr -> nextTCB) {   // edge case: at last task
141     listTailPtr = currentTaskPtr -> prevTCB;
142     listTailPtr -> nextTCB = NULL;
143   } else {   // reassign pointers
144     currentTaskPtr -> nextTCB -> prevTCB = currentTaskPtr ->prevTCB;
145     currentTaskPtr -> prevTCB -> nextTCB = currentTaskPtr -> nextTCB;
146   }
147   currentTaskPtr = currentTaskPtr -> nextTCB; // update currentTaskPtr
148      pointer
149 }
150
151 // Software delay
152 void delay_in_ms(int ms) {
153   for (volatile int i = 0; i < ms; i++)
154     for (volatile int j = 0; j < 800; j++);
155 }
156
157 // for debug (obviously)
158 void debug() {
159   char num[30];
160     static int test = 0;
161     usnprintf(num, 30, "Test number: %d  ", test);
162     RIT128x96x4StringDraw(num, 0, 90, 15);
163     test++;
164 }
```

C.5 Tasks

C.5.1 Task Interface

../code/task.h

```
1 /*
2  * task.h
3  * Author(s): Jonathan Ellington
4  * 1/28/2014
5  *
6  * Defines the interface for a task
7  */
8
```

```
9  #ifndef _TASK_H
10 #define _TASK_H
11
12 typedef struct tcb_struct {
13   void (*runTaskFunction) (void*);
14   void *taskDataPtr;
15   struct tcb_struct *nextTCB; // pointer to next TCB
16   struct tcb_struct *prevTCB; // points to previous TCB
17 } TCB;
18
19 #endif // _TASK_H
```

C.5.2 Startup Task

../code/startup.h

```
1  /*
2   * startup.h
3   * author: Patrick ma
4   * Date: 2/13/2014
5   *
6   * initializes the system and system variables
7   */
8
9  extern tBoolean runSchedule;
10
11 // Interrupt handler for hw timer
12 void SysTickIntHandler();
13
14 // configure the hardware timer
15 void initializeHWCounter();
16
17 // initialize system variables and hardware timer
18 void startup();
```

../code/startup.c

```
1  /*
2   * startup.c
3   * author: Patrick ma
4   * Date: 2/13/2014
5   *
6   * initializes the system and system variables
7   */
8
9  #include "timebase.h"
10 #include "globals.h"
11 #include "driverlib/systick.h" // for systick interrupt (minor cycle)
12 #include "driverlib/sysctl.h"
13 #include "driverlib/interrupt.h"
14 #include "startup.h"
15 #include "schedule.h"
```

```
16
17
18
19
20 /*
21  * Minor Cycle Handler
22  */
23 void SysTickIntHandler (void) {
24    minor_cycle_ctr = minor_cycle_ctr + 1;
25    runSchedule = true;
26 }
27
28 /*
29  * configure the hw interrupt for minor cycle
30  */
31 void initializeHWCounter() {
32 SysTickPeriodSet(SysCtlClockGet() * MINOR_CYCLE / 1000);   // set timer
       period
33
34 // enable interrupts (master)
35 IntMasterEnable();   // this may be redundant, consider moving
36 SysTickIntEnable();    // enable systick interrupt
37 SysTickEnable();   // enable systic counter
38 }
39
40
41 /*
42  * Initializes hw counter ans system state variables
43  */
44 void startup() {
45    initializeHWCounter();
46
47    #if DEBUG
48    RIT128x96x4Init(1000000);
49    #endif
50
51    // Initialize global data
52    initializeGlobalData();    // from globals.h
53
54    computeActive = false;   // neither serial or compute task runs at start up
55    serialActive = false;
56
57    initializeQueue(); // start up task queue with basic tasks
58 }
```

C.5.3   Measure Task

../code/measure.h

```
1 /*
2  * measure.h
3  * Author(s): Jonathan Ellington
```

```
 4   * 1/28/2014
 5   *
 6   * Defines the interface for the measureTask.
 7   * initializeMeasureData() should be called before running measureTask()
 8   */
 9
10  #include "task.h"
11
12  /* Points to the TCB for measure */
13  extern TCB measureTask;
```

../code/measure.c

```
 1  /*
 2   * measure.h
 3   * Author(s): Jonathan Ellington
 4   * 1/28/2014
 5   *
 6   * Implements measure.c
 7   *
 8   * Uses port PA4 for interrupt input for pulse rate
 9   */
10
11  #include "task.h"
12  #include "CircularBuffer.h"
13  #include "globals.h"
14  #include "timebase.h"
15  #include "measure.h"
16  #include "schedule.h"
17
18  #include "inc/hw_memmap.h"
19  #include "inc/hw_types.h"
20  #include "driverlib/gpio.h"
21  #include "driverlib/sysctl.h"
22  #include "driverlib/interrupt.h"
23  #include "inc/hw_ints.h"
24  #include "driverlib/debug.h"
25
26  // Used for debug display
27  #if DEBUG
28  #include "drivers/rit128x96x4.h"
29  #include "utils/ustdlib.h"
30  #endif
31
32  // prototype for compiler
33  void measureRunFunction(void *dataptr);
34  void PulseRateISR(void);
35
36  // Internal data structure
37  typedef struct measureData {
38    CircularBuffer *temperatureRaw;
39    CircularBuffer *systolicPressRaw;
40    CircularBuffer *diastolicPressRaw;
```

```
41     CircularBuffer *pulseRateRaw;
42  } MeasureData;
43
44  static int pulseRate = 0;
45  static MeasureData data;  // internal data
46  TCB measureTask = {&measureRunFunction, &data};  // task interface
47
48  void initializeMeasureTask() {
49  #if DEBUG
50     RIT128x96x4Init(1000000);
51  #endif
52     // Load data memory
53     data.temperatureRaw = &(global.temperatureRaw);
54     data.systolicPressRaw = &(global.systolicPressRaw);
55     data.diastolicPressRaw = &(global.diastolicPressRaw);
56     data.pulseRateRaw = &(global.pulseRateRaw);
57
58
59     /* Interrupt setup
60      * Note: using statically registered interrupts, because they're faster
61      *       this means we aren't using the dynamic GPIOPortIntRegister()
62      *       instead, an entry in the interrupt table is populated with the
63      *       of the interrupt handler (under GPIO Port A) and this is enabled
64      *       IntEnable(INT_GPIOA) */
65
66     SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOA);
67
68     // Set PA4 as input
69     GPIOPinTypeGPIOInput(GPIO_PORTA_BASE, GPIO_PIN_4);
70
71     // Enable interrupts on PA4
72     GPIOPinIntEnable(GPIO_PORTA_BASE, GPIO_PIN_4);
73
74     // Set interrupt type
75     GPIOIntTypeSet(GPIO_PORTA_BASE, GPIO_PIN_4, GPIO_RISING_EDGE);
76
77     // Enable interrupts to the processor.
78     IntMasterEnable();
79
80     // Enable the interrupts.
81     IntEnable(INT_GPIOA);
82  }
83
84  void setTemp(CircularBuffer *tbuf) {
85     static unsigned int i = 0;
86     static tBoolean goingUp = true;
87
88     int temp = *(int *)cbGet(tbuf);
89
90     if (temp > 50)
```

```
 91       goingUp = false;
 92     else if (temp < 15)
 93       goingUp = true;
 94
 95     if (goingUp) {
 96       if (i%2==0) temp+=2;
 97       else temp--;
 98     }
 99     else {
100       if (i%2==0) temp-=2;
101       else temp++;
102     }
103
104     cbAdd(tbuf, &temp);
105
106     i++;
107 }
108
109 void setBloodPress(CircularBuffer *spbuf, CircularBuffer *dpbuf) {
110     // This is written to lab spec, with a flag to indicate "complete".
111     // Right now, it does nothing, but I imagine it should probably be a
          global
112     // variable to indicate to the compute task that the pressure measurement
113     // is ready, since this measurement takes a nontrivial amount of time
114
115     static unsigned int i = 0;
116
117     static tBoolean sysComplete = false;
118     static tBoolean diaComplete = false;
119
120     int syspress = *(int *)cbGet(spbuf);
121
122     // Restart systolic measurement if diastolic is complete
123
124     if (syspress > 100)
125       sysComplete = true;
126
127     if (! sysComplete) {
128       if (i%2==0) syspress+=3;
129       else syspress--;
130     }
131
132     int diapress = *(int *)cbGet(dpbuf);
133
134     // Restart diastolic measurement if systolic is complete
135     if (diapress < 40)
136       diaComplete = true;
137
138     if (!diaComplete) {
139       if (i%2==0) diapress-=2;
140       else diapress++;
141     }
142
```

```
143    if (diaComplete && sysComplete) {
144      sysComplete = false;
145      diaComplete = false;
146      syspress = SYS_RAW_INIT;
147      diapress = DIA_RAW_INIT;
148    }
149
150    cbAdd(spbuf, &syspress);
151    cbAdd(dpbuf, &diapress);
152
153    i++;
154  }
155
156  // pulse rate interrupt handler
157  void PulseRateISR(void) {
158    pulseRate++;
159
160    GPIOPinIntClear(GPIO_PORTA_BASE, GPIO_PIN_4);
161  }
162
163  void measureRunFunction(void *dataptr) {
164    static tBoolean onFirstRun = true;
165    static int rate;
166    MeasureData *mData = (MeasureData *) dataptr;
167
168    if (onFirstRun) {
169      initializeMeasureTask();
170      rate = *(int*) cbGet(mData->pulseRateRaw);
171      onFirstRun = false;
172    }
173
174    // capture pulse rate
175    if (IS_PULSE_CYCLE) {
176      // Divide by two so raw pulse rate matches frequency
177      rate = pulseRate/2;
178      pulseRate = 0;
179    }
180
181    // only run on major cycle
182    if (IS_MAJOR_CYCLE) {
183      setTemp(mData->temperatureRaw);
184      setBloodPress(mData->systolicPressRaw, mData->diastolicPressRaw);
185
186      int prev = *(int*) cbGet(mData->pulseRateRaw);
187
188      // Only save if +- 15%
189      if (rate < prev*0.85 || rate > prev*1.15) {
190        cbAdd(mData->pulseRateRaw, (void *)&rate);
191      }
192    computeActive = true;    // run the compute task
193
194  #if DEBUG
195      char num[30];
```

```
196    int temp = *(int *)cbGet(mData->temperatureRaw);
197    int sys = *(int *)cbGet(mData->systolicPressRaw);
198    int dia = *(int *)cbGet(mData->diastolicPressRaw);
199    int pulse = *(int *)cbGet(mData->pulseRateRaw);
200    int batt = global.batteryState;
201
202    usnprintf(num, 30, "Raw temp: %d  ", temp);
203    RIT128x96x4StringDraw(num, 0, 0, 15);
204
205    usnprintf(num, 30, "Raw Syst: %d  ", sys);
206    RIT128x96x4StringDraw(num, 0, 10, 15);
207
208    usnprintf(num, 30, "Raw Dia: %d  ", dia);
209    RIT128x96x4StringDraw(num, 0, 20, 15);
210
211    usnprintf(num, 30, "Raw Pulse: %d  ", pulse);
212    RIT128x96x4StringDraw(num, 0, 30, 15);
213
214    usnprintf(num, 30, "Raw Batt: %d  ", batt);
215    RIT128x96x4StringDraw(num, 0, 40, 15);
216 #endif
217   }
218 }
```

C.5.4   Compute Task

../code/compute.h

```
1  /*
2   * compute.h
3   * Author(s): PatrickMa
4   * 1/28/2014
5   *
6   * Defines the public interface for computeTask
7   * initializeComputeData() should be called before running computeTask()
8   */
9
10 #include "task.h"
11
12 /* Points to the TCB for compute */
13 extern TCB computeTask;
```

../code/compute.c

```
1  /*
2   * compute.c
3   * Author(s): PatrickMa
4   * 1/28/2014
5   *
6   * Implements compute.c
7   */
8
```

```
 9  #include "task.h"
10  #include "CircularBuffer.h"
11  #include "compute.h"
12  #include "globals.h"
13  #include "timebase.h"
14  #include "schedule.h"
15
16  // Used for debug display
17  #if DEBUG
18  #include "drivers/rit128x96x4.h"
19  #include "utils/ustdlib.h"
20  #endif
21
22  // computeData structure internal to compute task
23  typedef struct computeData {
24      // raw data pointers
25      CircularBuffer *temperatureRaw;
26      CircularBuffer *systolicPressRaw;
27      CircularBuffer *diastolicPressRaw;
28      CircularBuffer *pulseRateRaw;
29
30      //corrected data pointers
31      CircularBuffer *temperatureCorrected;
32      CircularBuffer *systolicPressCorrected;
33      CircularBuffer *diastolicPressCorrected;
34      CircularBuffer *pulseRateCorrected;
35  } ComputeData;
36
37  void computeRunFunction(void *computeData);
38
39  static ComputeData data;     // the internal data
40  TCB computeTask = {&computeRunFunction, &data};  // task interface
41
42  /*
43   * Initializes the computeData task values (pointers to variables, etc)
44   */
45  void initializeComputeTask() {
46      data.temperatureRaw = &(global.temperatureRaw);
47      data.systolicPressRaw = &(global.systolicPressRaw);
48      data.diastolicPressRaw = &(global.diastolicPressRaw);
49      data.pulseRateRaw = &(global.pulseRateRaw);
50
51      data.temperatureCorrected = &(global.temperatureCorrected);
52      data.systolicPressCorrected = &(global.systolicPressCorrected);
53      data.diastolicPressCorrected = &(global.diastolicPressCorrected);
54      data.pulseRateCorrected = &(global.pulseRateCorrected);
55  }
56
57  /*
58   * Linearizes the raw data measurement and converts value into human
59   * readable format
60   */
61  void computeRunFunction(void *computeData) {
```

```
62    static tBoolean onFirstRun = true;
63
64    if (onFirstRun) {
65      initializeComputeTask();
66      onFirstRun = false;
67    }
68
69    if (IS_MAJOR_CYCLE) {
70      ComputeData *cData = (ComputeData *) computeData;
71
72      float temp = 5 + 0.75 * (*(int *)cbGet(cData->temperatureRaw));
73      float systolic = 9 + 2 * (*(int *)cbGet(cData->systolicPressRaw));
74      float diastolic = 6 + 1.5 * (*(int *)cbGet(cData->diastolicPressRaw));
75      float pulseRate = 8 + 3 * (*(int *)cbGet(cData->pulseRateRaw));
76
77      cbAdd(cData->temperatureCorrected, &temp);
78      cbAdd(cData->systolicPressCorrected, &systolic);
79      cbAdd(cData->diastolicPressCorrected, &diastolic);
80      cbAdd(cData->pulseRateCorrected, &pulseRate);
81
82      computeActive = false;  // remove self from task queue
83
84 #if DEBUG
85      char num[30];
86      usnprintf(num, 30, "Corrected temp: %d", (unsigned int) temp);
87      RIT128x96x4StringDraw(num, 0, 50, 15);
88
89      usnprintf(num, 30, "Corrected Syst: %d", (unsigned int) systolic);
90      RIT128x96x4StringDraw(num, 0, 60, 15);
91
92      usnprintf(num, 30, "Corrected Dia: %d", (unsigned int) diastolic);
93      RIT128x96x4StringDraw(num, 0, 70, 15);
94
95      usnprintf(num, 30, "Corrected Pulse: %d", (unsigned int) pulseRate);
96      RIT128x96x4StringDraw(num, 0, 80, 15);
97 #endif
98    }
99 }
```

C.5.5   Keypad Task

../code/keypad.h

```
1 /*
2  * keyPad.h
3  * Author(s): Jarrett Gaddy
4  * 2/10/2014
5  *
6  * Defines the public interface for the keyPad task
7  *
8  * initializeKeyPadTask() should be called once before performing runkeyPad
9  * functions
```

```
10   */
11
12  #include "task.h"  // for TCBs
13
14  /* Initialize KeyPadData, must be done before running functions */
15  void initializeKeyPadTask();
16
17  /* The keyPad Task */
18  extern TCB keyPadTask;
```

../code/keypad.c

```
1   /*
2    * keyPad.c
3    * Author(s): Jarrett Gaddy
4    * 2/10/2014
5    *
6    * implements keyPad.h
7    */
8
9   #include "keyPad.h"
10  #include "globals.h"
11  #include "timebase.h"
12  #include "inc/hw_types.h"
13  #include "driverlib/sysctl.h"
14  #include "driverlib/gpio.h"
15  #include "inc/hw_memmap.h"
16
17  #define UP_SW GPIOPinRead(GPIO_PORTD_BASE, GPIO_PIN_4)
18  #define DOWN_SW GPIOPinRead(GPIO_PORTD_BASE, GPIO_PIN_5)
19  #define LEFT_SW GPIOPinRead(GPIO_PORTD_BASE, GPIO_PIN_6)
20  #define RIGHT_SW GPIOPinRead(GPIO_PORTD_BASE, GPIO_PIN_7)
21  #define ACK_SW GPIOPinRead(GPIO_PORTE_BASE, GPIO_PIN_0)
22
23  // StatusData structure internal to compute task
24  typedef struct {
25    unsigned short *mode;
26    unsigned short *measurementSelection;
27    unsigned short *scroll;
28    tBoolean *alarmAcknowledge;
29    tBoolean *select;
30  } KeyPadData;
31
32  void keyPadRunFunction(void *data);  // prototype for compiler
33
34  static KeyPadData data;  // the internal data
35  TCB keyPadTask = {&keyPadRunFunction, &data}; // task interface
36
37  /* Initialize the StatusData task values */
38  void initializeKeyPadTask() {
39
40        SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOD);
41        SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOE);
```

```
42
43
44
45     GPIOPadConfigSet(GPIO_PORTE_BASE, GPIO_PIN_0, GPIO_STRENGTH_2MA,
46       GPIO_PIN_TYPE_STD_WPU);
47     GPIODirModeSet(GPIO_PORTE_BASE, GPIO_PIN_0, GPIO_DIR_MODE_IN);
48
49     GPIOPadConfigSet(GPIO_PORTD_BASE, GPIO_PIN_4, GPIO_STRENGTH_2MA,
50       GPIO_PIN_TYPE_STD);
51     GPIODirModeSet(GPIO_PORTD_BASE, GPIO_PIN_4, GPIO_DIR_MODE_IN);
52
53     GPIOPadConfigSet(GPIO_PORTD_BASE, GPIO_PIN_5, GPIO_STRENGTH_2MA,
54       GPIO_PIN_TYPE_STD);
55     GPIODirModeSet(GPIO_PORTD_BASE, GPIO_PIN_5, GPIO_DIR_MODE_IN);
56
57     GPIOPadConfigSet(GPIO_PORTD_BASE, GPIO_PIN_6, GPIO_STRENGTH_2MA,
58       GPIO_PIN_TYPE_STD);
59     GPIODirModeSet(GPIO_PORTD_BASE, GPIO_PIN_6, GPIO_DIR_MODE_IN);
60
61     GPIOPadConfigSet(GPIO_PORTD_BASE, GPIO_PIN_7, GPIO_STRENGTH_2MA,
62       GPIO_PIN_TYPE_STD);
63     GPIODirModeSet(GPIO_PORTD_BASE, GPIO_PIN_7, GPIO_DIR_MODE_IN);
64
65     //for ack switch
66     /* GPIOPadConfigSet(GPIO_PORTE_BASE, GPIO_PIN_3, GPIO_STRENGTH_2MA,
67       GPIO_PIN_TYPE_STD_WPU);
68     GPIODirModeSet(GPIO_PORTE_BASE, GPIO_PIN_3, GPIO_DIR_MODE_IN);*/
69
70
71
72     // Load data
73     data.mode = &(global.mode);
74     data.measurementSelection = &(global.measurementSelection);
75     data.scroll = &(global.scroll);
76     data.alarmAcknowledge = &(global.alarmAcknowledge);
77     data.select = &(global.select);
78
79     // Load TCB
80     keyPadTask.runTaskFunction = &keyPadRunFunction;
81     keyPadTask.taskDataPtr = &data;
82 }
83
84 /* Perform status tasks */
85 void keyPadRunFunction(void *keyPadData){
86
87     static tBoolean onFirstRun = true;
88
89     if (onFirstRun) {
90       initializeKeyPadTask();
91       onFirstRun = false;
92     }
93     KeyPadData *kdata = (KeyPadData *) keyPadData;
94     if( 0 == *(kdata->mode))
```

```
95    {
96
97      if ( false == *(kdata->select))
98      {
99
100       if (UP_SW)
101         *(kdata->scroll) = *(kdata->scroll) +1;
102       else if (DOWN_SW)
103         *(kdata->scroll) = *(kdata->scroll) -1;
104       else if (RIGHT_SW)
105         *(kdata->select) = true;
106                         else if (LEFT_SW)
107                         {
108                             *(kdata->mode) = 1;
109                         }
110
111     }
112     else
113     {
114                 if (LEFT_SW)
115       *(kdata->select) = false;
116     }
117  }
118  else
119  {
120    if (RIGHT_SW)
121    {
122      *(kdata->mode) = 0;
123    }
124  }
125  if (1 == ACK_SW)
126  {
127    *(kdata->alarmAcknowledge) = true;
128  }
129  else
130    *(kdata->alarmAcknowledge) = false;
131 }
```

C.5.6   Display Task

../code/display.h

```
1  /*
2   * display.h
3   * Author(s): Jarrett Gaddy
4   * 1/28/2014, updated 2/10/2014
5   *
6   * Defines the interface for the displayTask.
7   * initializeDisplayData() should be called before running displayTask()
8   */
9
10 #include "task.h"
```

```
11
12  /* Initialize DisplayData, must be done before running displayTask() */
13  void initializeDisplayTask();
14
15  /* Points to the TCB for display */
16  extern TCB displayTask;
```

../code/display.c

```
1  /*
2   * display.c
3   * Author(s): jarrett Gaddy
4   * 2/10/2014
5   *
6   * Implements display.h
7   */
8
9  #include "globals.h"
10 #include "task.h"
11 #include "timebase.h"
12 #include "display.h"
13 #include "inc/hw_types.h"
14 #include "drivers/rit128x96x4.h"
15 #include "utils/ustdlib.h"
16 #include <stdlib.h>
17
18
19 // Internal data structure
20 typedef struct oledDisplayData {
21   CircularBuffer *temperatureCorrected;
22   CircularBuffer *systolicPressCorrected;
23   CircularBuffer *diastolicPressCorrected;
24   CircularBuffer *pulseRateCorrected;
25   unsigned short *batteryState;
26   unsigned short *mode;
27   unsigned short *measurementSelection;
28   unsigned short *scroll;
29   tBoolean *alarmAcknowledge;
30   tBoolean *select;
31   unsigned short *selection;
32 } DisplayData;
33
34 void displayRunFunction(void *dataptr);  // prototype for compiler
35
36 static DisplayData data;  // internal data
37 TCB displayTask = {&displayRunFunction, &data};          // task interface
38
39 void initializeDisplayTask() {
40   RIT128x96x4Init(1000000);
41
42   // Load data
43   data.temperatureCorrected = &(global.temperatureCorrected);
44   data.systolicPressCorrected = &(global.systolicPressCorrected);
```

47

```
45    data.diastolicPressCorrected = &(global.diastolicPressCorrected);
46    data.pulseRateCorrected = &(global.pulseRateCorrected);
47    data.batteryState = &(global.batteryState);
48
49
50    data.mode = &(global.mode);
51    data.measurementSelection = &(global.measurementSelection);
52    data.scroll = &(global.scroll);
53    data.alarmAcknowledge = &(global.alarmAcknowledge);
54    data.select = &(global.select);
55
56
57    // Load TCB
58    displayTask.runTaskFunction = &displayRunFunction;
59    displayTask.taskDataPtr = &data;
60 }
61
62
63 void displayRunFunction(void *dataptr) {
64    // only run on major cycle
65    //  if (IS_MAJOR_CYCLE) {   // on major cycle
66
67    static tBoolean onFirstRun = true;
68
69    if (onFirstRun) {
70      initializeDisplayTask();
71      onFirstRun = false;
72    }
73
74    DisplayData *dData = (DisplayData *) dataptr;
75
76    tBoolean selection = *(dData->select);
77    int scroll = *(dData->scroll);
78
79 #if !DEBUG
80    char num[40];
81    //char buf1[30];
82    char buf2[30];
83
84    if(0 == *(dData->mode))
85    {
86      if(false == selection)
87      {
88        RIT128x96x4StringDraw("Make Selection                        ", 0, 0,
      15);
89        RIT128x96x4StringDraw("   Blood Pressure                     ", 0, 10,
      15);
90        RIT128x96x4StringDraw("   Temperature                        ", 0, 20,
      15);
91        RIT128x96x4StringDraw("   Pulse Rate                         ", 0, 30,
      15);
92        RIT128x96x4StringDraw("   Battery                            ", 0, 40,
      15);
```

48

```
 93        RIT128x96x4StringDraw("                                        ", 0, 50,
        15);
 94        RIT128x96x4StringDraw("                                        ", 0, 60,
        15);
 95        RIT128x96x4StringDraw("->", 0, 10*((scroll%4+1)), 15);
 96      }
 97      else
 98      {
 99        RIT128x96x4StringDraw("                                        ", 0, 0,
      15);
100        if(0 == scroll%4)
101          RIT128x96x4StringDraw("Blood Pressure:", 0, 0, 15);
102        else if(1 == scroll%4)
103          RIT128x96x4StringDraw("Temperature:", 0, 0, 15);
104        else if(2 == scroll%4)
105          RIT128x96x4StringDraw("Pulse Rate:", 0, 0, 15);
106        else if(3 == scroll%4)
107          RIT128x96x4StringDraw("Battery:", 0, 0, 15);
108        else RIT128x96x4StringDraw("oops", 0, 0, 15);//just in case
109
110
111
112        if(0 == scroll%4)
113          usnprintf(buf2,30, "Systolic: %d mm Hg ", (int) *( (float*) cbGet(
      dData->systolicPressCorrected)));
114        else if(1 == scroll%4)
115          usnprintf(buf2,30,"%d C ", (int) *( (float*) cbGet(dData->
      temperatureCorrected)));
116        else if(2 == scroll%4)
117          usnprintf(buf2,30, "%d BPM ", (int) *( (float*) cbGet(dData->
      pulseRateCorrected)));
118        else if(3 == scroll%4)
119          usnprintf(buf2,30, "%d %% ", (int) *(dData->batteryState)/2);
120        //else buf2 = "oops"; //just in case
121
122        RIT128x96x4StringDraw("                                        ", 0, 10,
        15);
123        RIT128x96x4StringDraw(buf2, 0, 10, 15);
124        if(0 == scroll%4)
125        {
126          usnprintf(buf2,30, "Diastolic: %d mm Hg        ", (int)*( (float*)
      cbGet(dData->diastolicPressCorrected)));
127
128          RIT128x96x4StringDraw(buf2, 0, 20, 15);
129        }
130        else RIT128x96x4StringDraw("                                        ",
      0, 20, 15);
131        RIT128x96x4StringDraw("                                        ", 0, 30,
        15);
132        RIT128x96x4StringDraw("                                        ", 0, 40,
        15);
133        RIT128x96x4StringDraw("                                        ", 0, 50,
        15);
```

```
134        RIT128x96x4StringDraw("                                    ", 0, 60,
       15);

135
136      }
137    }
138    else //(1 == *(data->mode)
139    {
140      usnprintf(num,40,"Temperature: %d C            ", (int) *( (float*) cbGet
       (dData->temperatureCorrected)));
141      RIT128x96x4StringDraw(num, 0, 0, 15);
142
143      usnprintf(num,40, "Systolic Pressure:           ");
144      RIT128x96x4StringDraw(num, 0, 10, 15);
145
146      usnprintf(num,40, "%d mm Hg                    ", (int) *( (float*) cbGet
       (dData->systolicPressCorrected)));
147      RIT128x96x4StringDraw(num, 0, 20, 15);
148
149      usnprintf(num,40, "Diastolic Pressure:          ");
150      RIT128x96x4StringDraw(num, 0, 30, 15);
151
152      usnprintf(num,40, "%d mm Hg                    ",(int) *( (float*)
       cbGet(dData->diastolicPressCorrected)));
153      RIT128x96x4StringDraw(num, 0, 40, 15);
154
155      usnprintf(num,40, "Pulse rate: %d BPM           ",(int) *( (float*)
       cbGet(dData->pulseRateCorrected)));
156      RIT128x96x4StringDraw(num, 0, 50, 15);
157
158      usnprintf(num,40, "Battery: %d %%                ",(int) *(dData->
       batteryState)/2);
159      RIT128x96x4StringDraw(num,0, 60,15);
160    }
161 #endif
162 }
```

C.5.7   Warning/Alarm Task

../code/warning.h

```
 1 /*
 2  * warning.h
 3  * Author(s): Jarrett Gaddy
 4  * 1/28/2014
 5  *
 6  * Defines the interface for the warning task
 7  * initializeWarningData() should be called before running warningTask()
 8  */
 9
10 #include "task.h"
11
12 #define WARN_LOW 0.95 //warn at 5% below min range value
```

```
13  #define WARN_HIGH 1.05 //warn at 5% above max range value
14  #define ALARM_LOW 0.90 //alarm at 10% below min range value
15  #define ALARM_HIGH 1.20 //alarm at 20% above max range value
16
17
18  #define WARN_RATE_PULSE    4    // flash rate in terms of minor cycles
19  #define WARN_RATE_TEMP     2
20  #define WARN_RATE_PRESS    1
21
22  #define TEMP_MIN 36.1
23  #define TEMP_MAX 37.8
24  #define SYS_MAX 120
25  #define DIA_MAX 80
26  #define PULSE_MIN 60
27  #define PULSE_MAX 100
28  #define BATTERY_MIN 40
29
30  /* Initialize displayData, must be done before running warningTask() */
31  void initializeWarningTask();
32
33  /* The warning task */
34  extern TCB warningTask;
```

../code/warning.c

```
1   /*
2    * warning.c
3    * Author(s): jarrett Gaddy, PatrickMa
4    * 1/28/2014
5    *
6    * Implements warning.h
7    */
8
9   #include "globals.h"
10  #include "task.h"
11  #include "timebase.h"
12  #include "warning.h"
13  #include "schedule.h"
14  #include "inc/hw_types.h"
15  #include "drivers/rit128x96x4.h"
16  #include <stdlib.h>
17  #include <stdio.h>
18
19  #include "inc/hw_memmap.h"
20  #include "driverlib/debug.h"
21  #include "driverlib/gpio.h"
22  #include "driverlib/pwm.h"
23  #include "driverlib/sysctl.h"
24
25  #if DEBUG
26  #include "drivers/rit128x96x4.h"
27  #include "utils/ustdlib.h"
28  #endif
```

```
29
30  // alarm cycle period (on/off) in millisecond
31  #define ALARM_PERIOD 2000
32  #define ALARM_CYCLE_RATE     (ALARM_PERIOD / MINOR_CYCLE / 2)
33
34  // duration to sleep in terms of minor cycles
35  #define ALARM_SLEEP_PERIOD   (5 * MAJOR_CYCLE)
36
37  #define LED_GREEN GPIO_PIN_6
38  #define LED_RED   GPIO_PIN_5
39  #define LED_YELLOW GPIO_PIN_7
40
41  typedef enum {OFF, ON, ASLEEP} alarmState;
42  typedef enum {NONE, WARN_PRESS, WARN_TEMP, WARN_PULSE} warningState;
43  typedef enum {NORMAL, LOW} batteryState;
44
45  //pin E0 for input on switch 3
46  //pin C5 C6 and C7 for led out
47
48  // compiler prototypes
49  void warningRunFunction(void *dataptr);  // prototype for compiler
50
51  // Internal data structure
52  typedef struct WarningData {
53    CircularBuffer *temperatureCorrected;
54    CircularBuffer *systolicPressCorrected;
55    CircularBuffer *diastolicPressCorrected;
56    CircularBuffer *pulseRateCorrected;
57    unsigned short *batteryState;
58  } WarningData;
59
60  extern tBoolean serialActive;
61  static unsigned long ulPeriod; // sets the alarm tone period
62  static WarningData data;        // internal data
63
64  TCB warningTask = {&warningRunFunction, &data};  // task interface
65
66
67  /*
68   * initializes task variables
69   */
70  void initializeWarningTask() {
71    //
72    // Enable the peripherals used by this code. I.e enable the use of pin
       banks, etc.
73    //
74    SysCtlPeripheralEnable(SYSCTL_PERIPH_PWM0);
75    SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOC);         // bank C
76    SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOE);         // bank E
77    SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOF);         // bank F
78    SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOG);         // bank G
79
80
```

```
81    // configure the pin C5 for 4mA output
82    GPIOPadConfigSet(GPIO_PORTC_BASE,LED_RED, GPIO_STRENGTH_4MA,
          GPIO_PIN_TYPE_STD);
83    GPIODirModeSet(GPIO_PORTC_BASE, LED_RED, GPIO_DIR_MODE_OUT);
84
85    // configure the pin C6 for 4mA output
86    GPIOPadConfigSet(GPIO_PORTC_BASE,LED_GREEN, GPIO_STRENGTH_4MA,
          GPIO_PIN_TYPE_STD);
87    GPIODirModeSet(GPIO_PORTC_BASE, LED_GREEN, GPIO_DIR_MODE_OUT);
88
89    // configure the pin C7 for 4mA output
90    GPIOPadConfigSet(GPIO_PORTC_BASE,LED_YELLOW, GPIO_STRENGTH_4MA,
          GPIO_PIN_TYPE_STD);
91    GPIODirModeSet(GPIO_PORTC_BASE, LED_YELLOW, GPIO_DIR_MODE_OUT);
92
93    // configure the pin E0 for input (sw3). NB: requires pull-up to operate
94    GPIOPadConfigSet(GPIO_PORTE_BASE, GPIO_PIN_0, GPIO_STRENGTH_2MA,
95        GPIO_PIN_TYPE_STD_WPU);
96    GPIODirModeSet(GPIO_PORTE_BASE, GPIO_PIN_0, GPIO_DIR_MODE_IN);
97
98
99    /*  This function call does the same result of the above pair of calls,
100    *   but still requires that the bank of peripheral pins is enabled via
101    *   SysCtrlPeripheralEnable()
102    */
103    //  GPIOPinTypeGPIOOutput(GPIO_PORTC_BASE, LED_RED);
104
105    ///////////////////////////////////////
106    //  This section defines the PWM speaker characteristics
107    ///////////////////////////////////////
108
109    //
110    // Set the clocking to run directly from the crystal.
111    //
112    SysCtlPWMClockSet(SYSCTL_PWMDIV_1);
113
114    //
115    // Set GPIO F0 and G1 as PWM pins.  They are used to output the PWM0 and
116    // PWM1 signals.
117    //
118    GPIOPinTypePWM(GPIO_PORTF_BASE, GPIO_PIN_0);
119    GPIOPinTypePWM(GPIO_PORTG_BASE, GPIO_PIN_1);
120
121    //
122    // Compute the PWM period based on the system clock.
123    //
124    ulPeriod = SysCtlClockGet() / 65;
125
126    //
127    // Set the PWM period to 440 (A) Hz.
128    //
129    PWMGenConfigure(PWM0_BASE, PWM_GEN_0,
130        PWM_GEN_MODE_UP_DOWN | PWM_GEN_MODE_NO_SYNC);
```

```
131    PWMGenPeriodSet(PWM0_BASE, PWM_GEN_0, ulPeriod);
132
133    //
134    // Set PWM0 to a duty cycle of 25% and PWM1 to a duty cycle of 75%.
135    //
136    PWMPulseWidthSet(PWM0_BASE, PWM_OUT_0, ulPeriod / 4);
137    PWMPulseWidthSet(PWM0_BASE, PWM_OUT_1, ulPeriod * 3 / 4);
138
139    //
140    // Enable the PWM0 and PWM1 output signals.
141    //
142    PWMOutputState(PWM0_BASE, PWM_OUT_0_BIT | PWM_OUT_1_BIT, true);
143
144    // initialize the warning data pointers
145    data.temperatureCorrected = &(global.temperatureCorrected);
146    data.systolicPressCorrected = &(global.systolicPressCorrected);
147    data.diastolicPressCorrected = &(global.diastolicPressCorrected);
148    data.pulseRateCorrected = &(global.pulseRateCorrected);
149    data.batteryState = &(global.batteryState);
150
151 }
152
153 ////////////////////////////////////////////////////////////
154
155 /*
156  * Warning task function
157  */
158 void warningRunFunction(void *dataptr) {
159
160    static alarmState aState = OFF;
161    static warningState wState = NONE;
162    static batteryState bState = NORMAL;
163
164    static warningState prevState;
165    prevState = wState;
166
167    static int wakeUpAlarmAt = 0;
168
169    static tBoolean onFirstRun = true;
170
171    if (onFirstRun){
172      initializeWarningTask();
173      onFirstRun = false;
174    }
175
176
177    // Get measurement data
178    WarningData *wData = (WarningData *) dataptr;
179    float temp = *((float*) cbGet(wData->temperatureCorrected));
180    float sysPress = *((float*) cbGet(wData->systolicPressCorrected));
181    float diaPress = *((float*) cbGet(wData->diastolicPressCorrected));
182    float pulse = *((float*) cbGet(wData->pulseRateCorrected));
183    unsigned short battery = *(wData->batteryState);
```

```
184
185    // Alarm condition
186    if ( (sysPress > SYS_MAX*ALARM_HIGH) ) { //|| // Commented lines = lab2
187       // (temp < TEMP_MIN*ALARM_LOW || temp > (TEMP_MAX*ALARM_HIGH)) ||
188       // (diaPress > DIA_MAX*ALARM_HIGH) ||
189       // (pulse < PULSE_MIN*ALARM_LOW || pulse > PULSE_MAX*ALARM_HIGH) ) {
190
191       // Should only turn alarm ON if it was previously OFF. If it is
192       // ASLEEP, shouldn't do anything.
193       if (aState == OFF) aState = ON;
194    }
195    else
196       aState = OFF;
197
198    // Warning Condition
199    if ( sysPress > SYS_MAX*ALARM_HIGH || diaPress > DIA_MAX*ALARM_HIGH )
200       wState = WARN_PRESS;
201    else if ( temp < TEMP_MIN*WARN_LOW || temp > TEMP_MAX*WARN_HIGH )
202       wState = WARN_TEMP;
203    else if ( pulse < PULSE_MIN*WARN_LOW || pulse > PULSE_MAX*WARN_HIGH )
204       wState = WARN_PULSE;
205    else
206       wState = NONE;
207
208    // Battery Condition
209    if (battery < BATTERY_MIN)
210       bState = LOW;
211
212    // Handle speaker, based on alarm state
213    static tBoolean pwmEnable = false;
214    switch (aState) {
215       case ON:
216
217          if (0 == (minor_cycle_ctr % ALARM_CYCLE_RATE)) { // toggle between on/
       off
218             if (pwmEnable)
219                PWMGenEnable(PWM0_BASE, PWM_GEN_0);
220             else
221                PWMGenDisable(PWM0_BASE, PWM_GEN_0);
222             pwmEnable = !pwmEnable;
223          }
224          break;
225       case ASLEEP:
226          PWMGenDisable(PWM0_BASE, PWM_GEN_0);
227          break;
228       default:  // OFF
229          PWMGenDisable(PWM0_BASE, PWM_GEN_0);
230
231          break;
232    }
233
234    // Handle warning cases
235    static int toggletime;
```

```
236    switch (wState) {
237      case WARN_PRESS:
238        GPIOPinWrite(GPIO_PORTC_BASE, LED_GREEN, 0X00);
239
240        if (wState != prevState) {
241          GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0XFF);
242          toggletime = WARN_RATE_PRESS;
243        }
244        else if (0 == minor_cycle_ctr%toggletime) {
245          if (GPIOPinRead(GPIO_PORTC_BASE, LED_RED) == 0)
246            GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0XFF);
247          else
248            GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0X00);
249        }
250
251        break;
252      case WARN_TEMP:
253        GPIOPinWrite(GPIO_PORTC_BASE, LED_GREEN, 0X00);
254
255        if (wState != prevState) {
256          GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0XFF);
257          toggletime = WARN_RATE_TEMP;
258        }
259        else if (0 == minor_cycle_ctr%toggletime) {
260          if (GPIOPinRead(GPIO_PORTC_BASE, LED_RED) == 0)
261            GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0XFF);
262          else
263            GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0X00);
264        }
265        break;
266      case WARN_PULSE:
267        GPIOPinWrite(GPIO_PORTC_BASE, LED_GREEN, 0X00);
268
269        if (wState != prevState) {
270          GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0XFF);
271          toggletime = WARN_RATE_PULSE;
272        }
273        else if (0==minor_cycle_ctr%toggletime) {
274          if (GPIOPinRead(GPIO_PORTC_BASE, LED_RED) == 0)
275            GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0XFF);
276          else
277            GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0X00);
278        }
279        break;
280      default:    // NORMAL
281        GPIOPinWrite(GPIO_PORTC_BASE, LED_GREEN, 0xFF);
282        GPIOPinWrite(GPIO_PORTC_BASE, LED_RED, 0x00);
283        break;
284    }
285
286    // activate the remote terminal task if in ANY warn or alarm state
287    if (NONE == wState && OFF == aState)
288      serialActive = false;
```

```
289     else
290       serialActive = true;
291
292    // battery state indicator
293    if (bState == LOW) {
294       GPIOPinWrite(GPIO_PORTC_BASE, LED_YELLOW, 0xFF);
295       GPIOPinWrite(GPIO_PORTC_BASE, LED_GREEN, 0x00);
296    }
297    else
298       GPIOPinWrite(GPIO_PORTC_BASE, LED_YELLOW, 0x00);
299
300    /* This is the alarm override
301     * Upon override, the alarm is silenced for some time.
302     * silence length is defined by ALARM_SLEEP_PERIOD
303     *
304     * If the button is pushed, the value returned is 0
305     * If the button is NOT pushed, the value is non-zero
306     */
307    if ( ( 0 == global.alarmAcknowledge) && (aState == ON) )
308    {
309       //GPIOPinWrite(GPIO_PORTC_BASE, LED_YELLOW, 0XFF);  // for debug, lights
          led
310       aState = ASLEEP;
311       wakeUpAlarmAt = minor_cycle_ctr + ALARM_SLEEP_PERIOD;
312    }
313
314    // Check whether to resound alarm
315    if (minor_cycle_ctr == wakeUpAlarmAt && aState == ASLEEP) {
316       aState = ON;
317       //GPIOPinWrite(GPIO_PORTC_BASE, LED_YELLOW, 0X00);  // for debug, kills
          led
318    }
319
320 #if DEBUG
321       char num[30];
322
323
324       usnprintf(num, 30, "Cor temp: %d   ", (int) temp);
325       RIT128x96x4StringDraw(num, 0, 0, 15);
326
327       usnprintf(num, 30, "Cor Syst: %d   ", (int) sysPress);
328       RIT128x96x4StringDraw(num, 0, 10, 15);
329
330       usnprintf(num, 30, "Cor Dia: %d   ", (int) diaPress);
331       RIT128x96x4StringDraw(num, 0, 20, 15);
332
333       usnprintf(num, 30, "Cor Pulse: %d   ", (int) pulse);
334       RIT128x96x4StringDraw(num, 0, 30, 15);
335
336       usnprintf(num, 30, "Cor Batt: %d   ", (unsigned short) battery);
337       RIT128x96x4StringDraw(num, 0, 40, 15);
338
339
```

```
340    usnprintf(num, 30, "aState: %d    ", aState);
341    RIT128x96x4StringDraw(num, 0, 50, 15);
342
343    usnprintf(num, 30, "alarmAck: %d    ", global.alarmAcknowledge);
344    RIT128x96x4StringDraw(num, 0, 60, 15);
345
346    usnprintf(num, 30, "pwmEn: %d    ", pwmEnable);
347    RIT128x96x4StringDraw(num, 0, 70, 15);
348 #endif
349 }
```

C.5.8   Serial Task

../code/serial.h

```
1  /*
2   * serial.h
3   * Author(s): Jonathan Ellington
4   * 2/05/2014
5   *
6   * Defines the serial communications task.  This sends various data over
7   * RS−232.
8   *
9   * This function should only run if there is a warning (in other words, the
10  * scheduler should add it to the task queue in the event of a warning), and
11  * should subsequently delete itself from the task queue.
12  */
13
14 #include "task.h"  // for TCBs
15
16 /* The status Task */
17 extern TCB serialTask;
```

../code/serial.c

```
1  /*
2   * serial.c
3   * Author(s): Jonathan Ellington
4   * 2/05/2014
5   *
6   * Implements serial.c
7   */
8
9  #include "task.h"
10 #include "globals.h"
11 #include "timebase.h"
12 #include "serial.h"
13 #include "schedule.h"
14 #include "CircularBuffer.h"
15 #include "inc/hw_types.h"
16 #include "driverlib/uart.h"
17 #include "driverlib/gpio.h"
```

```
18  #include "inc/hw_memmap.h"
19  #include "utils/ustdlib.h"
20  #include <string.h>
21
22
23  // Internal data structure
24  typedef struct serialData {
25    CircularBuffer *temperatureCorrected;
26    CircularBuffer *systolicPressCorrected;
27    CircularBuffer *diastolicPressCorrected;
28    CircularBuffer *pulseRateCorrected;
29    unsigned short *batteryState;
30  } SerialData;
31
32  // Prototype
33  void UARTSend(const unsigned char *pucBuffer, unsigned long ulCount);
34  void serialRunFunction(void *dataptr);
35
36  static SerialData data;  // internal data
37  TCB serialTask = {&serialRunFunction, &data};            // task interface
38
39  void initializeSerialTask() {
40    // Initialize Data
41    data.temperatureCorrected = &(global.temperatureCorrected);
42    data.systolicPressCorrected = &(global.systolicPressCorrected);
43    data.diastolicPressCorrected = &(global.diastolicPressCorrected);
44    data.pulseRateCorrected = &(global.pulseRateCorrected);
45    data.batteryState = &(global.batteryState);
46
47    // UART Stuff
48    // Enable the peripherals used by this example.
49    SysCtlPeripheralEnable(SYSCTL_PERIPH_UART0);
50    SysCtlPeripheralEnable(SYSCTL_PERIPH_GPIOA);
51
52    // Set GPIO A0 (UART RX)
53    GPIOPinTypeUART(GPIO_PORTA_BASE, GPIO_PIN_1);
54
55    // Configure the UART for 115,200, 8-N-1 operation.
56    UARTConfigSetExpClk(UART0_BASE, SysCtlClockGet(), 115200,
57        (UART_CONFIG_WLEN_8 | UART_CONFIG_STOP_ONE |
58         UART_CONFIG_PAR_NONE));
59
60    UARTEnable(UART0_BASE);
61  }
62
63  void serialRunFunction(void *dataptr) {
64    static tBoolean onFirstRun = true;
65    SerialData *sData = (SerialData *) dataptr;
66    if (onFirstRun) {
67      initializeSerialTask();
68      onFirstRun = false;
69    }
70
```

```
71    int temp = (int) *((float *)cbGet(sData→temperatureCorrected));
72    int sys = (int) *((float *)cbGet(sData→systolicPressCorrected));
73    int dia = (int) *((float *)cbGet(sData→diastolicPressCorrected));
74    int pulse = (int) (*(float *)cbGet(sData→pulseRateCorrected));
75    int batt = *(data.batteryState);
76
77    char buf[512];
78 //   usnprintf(buf, 512,
79 //            "\f1. Temperature:\t\t%d C\n\n\r"
80 //            "2. Systolic pressure:\t%d mm Hg\n\n\r"
81 //            "3. Diastolic pressure:\t%d mm Hg\n\n\r"
82 //            "4. Pulse rate:\t\t%d BPM\n\n\r"
83 //            "5. Battery:\t\t%d\n\n\r",
84 //            temp, sys, dia, pulse, batt);
85
86    // The cast removes a warning.  It is safe as long as buf contains
87    // plain ASCII (non extended)
88    UARTSend( (unsigned char *) buf, strlen(buf));
89
90    serialActive = false;
91 }
92
93 void UARTSend(const unsigned char *pucBuffer, unsigned long ulCount) {
94    // Loop while there are more characters to send.
95    while(ulCount−−)
96    {
97      // Write the next character to the UART.
98      // This blocks while the FIFO queue is full
99      UARTCharPut(UART0_BASE, *pucBuffer++);
100   }
101 }
```

C.5.9 Status

../code/status.h

```
1  /*
2   * status.h
3   * Author(s): PatrickMa
4   * 1/28/2014
5   *
6   * Defines the public interface for the status task
7   *
8   * initializeStatusTask() should be called once before performing status
9   * functions
10  */
11
12 #include "task.h"  // for TCBs
13
14 /* Initialize StatusData, must be done before running functions */
15 void initializeStatusTask();
16
```

```
17  /* The status Task */
18  extern TCB statusTask;
```

../code/status.c

```
1   /*
2    * status.c
3    * Author(s): PatrickMa
4    * 1/28/2014
5    *
6    * implements status.h
7    */
8
9   #include "status.h"
10  #include "globals.h"
11  #include "timebase.h"
12
13  // StatusData structure internal to compute task
14  typedef struct {
15    unsigned short *batteryState;
16  } StatusData;
17
18  void statusRunFunction(void *data);  // prototype for compiler
19
20  static StatusData data;  // the internal data
21  TCB statusTask = {&statusRunFunction, &data}; // task interface
22
23  /* Initialize the StatusData task values */
24  void initializeStatusTask() {
25    // Load data
26    data.batteryState = &(global.batteryState);
27
28    // Load TCB
29    statusTask.runTaskFunction = &statusRunFunction;
30    statusTask.taskDataPtr = &data;
31  }
32
33  /* Perform status tasks */
34  void statusRunFunction(void *data){
35    static tBoolean onFirstRun = true;
36
37    if (onFirstRun) {
38      initializeStatusTask();
39      onFirstRun = false;
40    }
41
42    if (IS_MAJOR_CYCLE) {
43      StatusData *sData = (StatusData *) data;
44        if (*(sData->batteryState) > 0)
45          *(sData->batteryState) = *(sData->batteryState) - 1;
46    }
47  }
```

C.6   Circular Buffer

```
1  /* CircularBuffer.h
2   * Jonathan Ellington
3   * 2/7/14
4   */
5
6  /* A circular buffer implementation.
7   * Meant to work without any dynamically allocated memory by wrap()ing
8   * user defined arrays.
9   *
10  * NOTE: This implementation keeps NO type information.  This means the
11  *       USER IS RESPONSIBLE FOR KEEPING TRACK OF THE TYPE OF ARRAY THAT
12  *       HAS BEEN WRAPPED!
13  *
14  * NOTE: The implementation here should also be hidden, but this is
15  *       troublesome without dynamic memory.  Do not rely on any of these
16  *       elements!
17  */
18
19 #ifndef _CIRCULAR_BUFFER_H
20 #define _CIRCULAR_BUFFER_H
21
22 typedef struct _circBuf {
23   void *array;
24   int sizeElm;
25   int nElm;
26   int currElm;
27 } CircularBuffer;
28
29 /* Wrap an array in a circular buffer.
30  *
31  * @param arr      the array to wrap
32  * @param sizeElem the size of each element, in bytes
33  * @param nElem    the number of elements in the array
34  *
35  * This function expects the array is freshly created.  It will overwrite
36  * array contents on adds.
37  */
38 CircularBuffer cbWrap(void *arr, int sizeElem, int nElem);
39
40 /* Returns a pointer to the current element in the circular buffer
41  * Be sure not to clobber this value! */
42 void *cbGet(CircularBuffer *cb);
43
44 /* Returns a pointer to the wrapped array, with the oldest element
45  * at the end and the newest element at the beginning */
46 void *cbGetArray(CircularBuffer *cb);
47
48 /* Adds elem to cb */
49 void cbAdd(CircularBuffer *cb, void *elem);
```

```
50
51  #endif  //  _CIRCULAR_BUFFER_H
```

../code/CircularBuffer.c

```c
1  /* CircularBuffer.c
2   * Jonathan Ellington
3   * 2/7/14
4   */
5
6  #include "CircularBuffer.h"
7  #include <stdio.h>
8
9  CircularBuffer cbWrap(void *arr, int se, int ne) {
10    CircularBuffer cb;
11    cb.array = arr;
12    cb.sizeElm = se;
13    cb.nElm = ne;
14    cb.currElm = 0;
15
16    return cb;
17  }
18
19  /* Returns the current element in the circular buffer */
20  void *cbGet(CircularBuffer *cb) {
21    int sizeElm = cb->sizeElm;
22    int index = cb->currElm;
23
24    unsigned char *bytePtr = (unsigned char *)cb->array;
25    bytePtr += sizeElm * index;
26
27    return (void *) bytePtr;
28  }
29
30  /* Returns a pointer to the wrapped array, with the oldest element
31   * at the end and the newest element at the beginning */
32  void *cbGetArray(CircularBuffer *cb) {
33    return NULL;
34  }
35
36  /* Adds elm to cb */
37  void cbAdd(CircularBuffer *cb, void *elm) {
38    cb->currElm = (cb->currElm + 1) % cb->nElm;
39
40    int sizeElm = cb->sizeElm;
41    int index = cb->currElm;
42
43    // copy sizeElm bytes from elm into the right spot
44    unsigned char *elmPtr = (unsigned char *) elm;
45    unsigned char *arrElmPtr = (unsigned char *) cb->array;
46    arrElmPtr += sizeElm * index;
47
48    for (int i = 0; i < sizeElm; i++) {
```

```
49      arrElmPtr[i] = elmPtr[i];
50    }
51 }
```